

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071223\
 Data File : BN026480.D
 Acq On : 12 Jul 2023 18:35
 Operator : MA/JU
 Sample : SSTD3.218
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.2235

Quant Time: Jul 13 04:17:45 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN071223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 13 04:15:36 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.001	152	2393	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.798	136	7589	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.621	164	4168	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.368	188	7115	0.400	ng/ul	0.00
17) Chrysene-d12	21.555	240	5497	0.400	ng/ul	0.00
23) Perylene-d12	23.996	264	5921	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.343	96	9237	0.982	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.377	88	9527	1.576	ng/ul#	88
14) Pentachlorophenol	17.026	266	6349	3.659	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071223\
Data File : BN026480.D
Acq On : 12 Jul 2023 18:35
Operator : MA/JU
Sample : SST3.218
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.2235

Quant Time: Jul 13 04:17:45 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN071223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 13 04:15:36 2023
Response via : Initial Calibration

