


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071223\
Data File : BN026474.D
Acq On : 12 Jul 2023 13:43
Operator : MA/JU
Sample : DFTPP351
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP351

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN071223.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Jul 13 04:33:34 2023



AutoFind: Scans 2482, 2483, 2484; Background Corrected with Scan 2475

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	48.1	76445	PASS
68	69	0.00	2	1.9	1393	PASS
69	198	0.00	100	46.6	74090	PASS
70	69	0.00	2	0.8	595	PASS
127	198	10	80	48.4	76813	PASS
197	198	0.00	2	0.6	991	PASS
198	198	100	100	100.0	158843	PASS
199	198	5	9	6.7	10650	PASS
275	198	10	60	27.7	43941	PASS
365	198	1	100	5.7	8993	PASS
441	198	0.01	100	14.7	23363	PASS
442	198	50	100	99.2	157568	PASS
443	442	15	24	19.8	31128	PASS

m/z	Abundance
43.00	194.0
43.90	254.0
69.00	179.0
73.00	774.0
76.90	245.0
132.90	402.0
162.80	169.0
176.90	156.0
190.90	223.0
192.80	345.0
207.00	2700.0
208.00	445.0
208.90	294.0
267.00	178.0
281.00	542.0
281.90	176.0
282.90	238.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

m/z	Abundance
36.90	531.0
38.00	1565.0
39.00	6947.0
39.90	430.0
41.10	283.0
44.00	568.0
44.90	393.0
49.00	1138.0
50.00	28280.0
51.00	96368.0
52.00	4826.0
53.00	207.0
54.90	830.0
56.00	2623.0
56.90	5338.0
58.00	288.0
60.90	1453.0
62.00	1491.0
63.00	3700.0
63.90	486.0
65.00	1677.0
66.00	228.0
67.00	182.0
67.90	1699.0
68.90	89200.0
70.00	664.0
73.00	1678.0
74.00	12197.0
75.00	17368.0
76.00	6082.0
77.00	101424.0
78.00	7068.0
79.00	8712.0
79.90	6050.0
81.00	9223.0
81.90	2099.0
82.90	1471.0
83.90	236.0
85.00	1368.0
86.00	2695.0
87.00	1124.0
87.90	235.0
88.90	183.0
91.00	1903.0
92.00	2256.0
93.00	17912.0
94.00	1183.0
94.90	328.0
95.90	636.0
97.00	341.0
98.00	14140.0
99.00	8481.0
99.90	789.0
101.00	4256.0
101.90	245.0
102.90	1466.0
103.90	3125.0
104.90	3373.0
106.00	1038.0
107.00	34216.0
108.00	4876.0
109.00	1059.0
110.00	47024.0
110.90	8823.0
111.90	1201.0
112.90	365.0
114.90	161.0
116.00	1613.0
116.90	50416.0
118.00	3376.0
118.90	312.0
119.90	408.0
121.90	2629.0
123.00	3668.0
123.90	1804.0
124.90	1458.0
126.10	283.0
127.00	89824.0
128.00	7283.0
129.00	41832.0
130.00	3804.0
130.90	1031.0
132.00	358.0
133.00	567.0
134.00	1252.0
135.00	3651.0
135.90	1486.0
137.00	2323.0
137.90	477.0
138.90	252.0
139.90	619.0
140.90	6350.0

Instrument :

BNA_N

ClientSampleId :

DFTPP351

142.00	1898.0
142.90	1711.0
143.90	277.0
145.00	276.0
146.00	1146.0
146.90	3590.0
148.00	11399.0
148.90	2215.0
149.90	427.0
151.00	748.0
151.70	539.0
152.90	2019.0
153.90	1567.0
155.00	3620.0
156.00	4485.0
157.00	833.0
157.90	1225.0
159.00	917.0
160.00	1952.0
160.90	2881.0
162.00	854.0
162.80	167.0
163.90	443.0
165.00	2660.0
165.90	1487.0
167.00	17312.0
168.00	8621.0
169.00	1067.0
170.00	263.0
171.10	325.0
171.90	1355.0
172.90	1552.0
173.90	2524.0
175.00	4725.0
176.10	1283.0
176.90	2725.0
177.90	793.0
178.90	9981.0
179.90	6127.0
181.00	2796.0
182.00	553.0
182.70	194.0
183.90	705.0
185.00	4788.0
186.00	29960.0
187.00	9306.0
187.90	956.0
188.90	2150.0
189.90	594.0
190.90	1633.0
191.90	2846.0
193.00	3383.0
194.00	843.0
194.80	254.0
195.90	5385.0
197.00	1161.0
197.90	172480.0
199.00	11975.0
199.90	989.0
201.10	491.0
201.70	538.0
202.90	1849.0
204.00	9234.0
205.00	12947.0
206.00	52224.0
207.00	9766.0
208.00	3144.0
208.90	1214.0
209.90	1102.0
211.00	2446.0
212.00	363.0
213.00	228.0
214.90	756.0
216.00	1199.0
216.90	16840.0
217.90	2011.0
218.90	191.0
219.70	160.0
221.00	4607.0
221.90	1686.0
223.00	4191.0
224.00	30448.0
225.00	8025.0
226.00	1104.0
227.00	17240.0
228.00	2553.0
228.90	3005.0
230.00	430.0
230.90	815.0
232.10	153.0
232.90	250.0
234.00	1116.0
235.00	905.0
235.90	899.0
237.00	1874.0
237.90	286.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

239.00	627.0
240.00	367.0
240.90	549.0
241.90	1813.0
243.00	1588.0
244.00	22656.0
245.00	2823.0
245.90	6330.0
247.00	1471.0
247.90	256.0
248.90	1054.0
250.00	168.0
251.10	214.0
251.90	359.0
252.90	726.0
254.00	821.0
255.00	122192.0
256.00	17912.0
256.90	1685.0
257.90	9485.0
258.90	1441.0
259.80	230.0
264.20	157.0
264.90	3667.0
265.90	556.0
266.90	217.0
270.90	336.0
271.90	433.0
273.00	3516.0
274.00	10262.0
275.00	46528.0
276.00	6355.0
277.00	6186.0
277.90	1097.0
279.00	219.0
280.90	718.0
281.90	231.0
282.90	816.0
283.90	226.0
285.00	1032.0
286.00	200.0
289.90	238.0
291.90	289.0
292.90	1478.0
294.00	318.0
294.90	501.0
296.00	19600.0
296.90	3212.0
297.90	359.0
300.90	164.0
301.90	275.0
303.00	1882.0
304.00	509.0
308.00	441.0
313.90	707.0
314.90	2165.0
316.00	956.0
320.90	551.0
322.10	196.0
323.00	5135.0
323.90	700.0
325.00	177.0
326.90	1092.0
327.80	428.0
332.00	280.0
332.90	673.0
334.00	3542.0
334.90	747.0
341.00	544.0
342.10	188.0
345.80	1202.0
352.10	1283.0
352.90	990.0
354.00	1341.0
354.90	509.0
364.10	199.0
364.90	8212.0
365.90	1061.0
369.90	174.0
370.80	319.0
372.00	2198.0
373.00	650.0
382.90	657.0
389.90	307.0
390.90	189.0
401.90	930.0
402.80	1212.0
403.90	441.0
420.80	1045.0
422.00	793.0
422.90	7352.0
424.00	1740.0
424.90	268.0
441.00	22832.0
441.90	136512.0
442.90	26616.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

443.902004.0

445.00219.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

m/z	Abundance
37.10	579.0
38.00	1318.0
39.00	6502.0
40.00	442.0
41.00	183.0
44.10	575.0
45.00	350.0
48.90	733.0
50.00	26688.0
51.00	88888.0
52.00	4463.0
54.90	574.0
56.00	2512.0
57.00	5034.0
57.90	281.0
59.00	161.0
60.90	1496.0
62.00	1649.0
63.00	3432.0
63.90	432.0
65.00	1483.0
66.90	304.0
68.00	1627.0
68.90	87696.0
70.00	829.0
73.00	1784.0
74.00	12045.0
75.00	15890.0
76.00	6443.0
77.00	101704.0
78.00	6651.0
78.90	8525.0
79.90	5562.0
81.00	7920.0
82.00	1813.0
83.00	1157.0
84.90	1164.0
86.00	2587.0
87.00	933.0
87.80	301.0
89.00	216.0
90.90	1931.0
92.00	2066.0
93.00	18184.0
94.00	827.0
94.80	318.0
96.00	694.0
97.10	460.0
98.00	14990.0
99.00	8830.0
99.90	658.0
101.00	3757.0
101.90	304.0
102.90	1431.0
103.90	3113.0
104.90	3368.0
106.00	1001.0
106.90	34840.0
108.00	4958.0
108.90	1244.0
109.90	49536.0
111.00	9213.0
112.00	1275.0
113.00	415.0
114.90	270.0
116.00	2174.0
117.00	51768.0
118.00	3451.0
118.80	449.0
119.90	335.0
121.00	224.0
122.00	3074.0
122.90	3975.0
123.90	1724.0
125.00	1479.0
126.00	502.0
127.00	91144.0
128.00	6960.0
129.00	44792.0
129.90	3865.0
130.90	757.0
131.80	385.0
132.80	569.0
133.90	1508.0
134.90	3888.0
136.00	1943.0
137.00	2454.0
137.90	478.0
138.80	258.0
140.00	714.0
141.00	6816.0
142.00	2556.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

143.00	1740.0
143.80	308.0
144.90	395.0
145.90	1319.0
147.00	3792.0
147.90	11768.0
149.00	1792.0
149.90	467.0
150.90	557.0
151.90	465.0
152.90	2143.0
154.00	1381.0
155.00	4041.0
156.00	5058.0
157.00	588.0
158.00	1140.0
158.90	972.0
159.90	2114.0
160.90	2830.0
162.00	708.0
163.00	242.0
164.00	452.0
164.90	2753.0
166.00	1894.0
167.00	16672.0
168.00	8752.0
169.00	1165.0
169.80	480.0
170.80	724.0
171.90	1351.0
172.90	1704.0
174.00	2640.0
175.00	5513.0
176.00	1376.0
177.00	2153.0
177.90	906.0
178.90	9533.0
180.00	6612.0
181.00	3452.0
182.00	521.0
182.80	345.0
184.00	946.0
185.00	4686.0
186.00	34208.0
187.00	10310.0
188.00	1115.0
189.00	2563.0
189.90	455.0
190.90	1634.0
191.90	3386.0
193.00	4641.0
194.00	699.0
194.90	357.0
195.90	5200.0
197.00	1166.0
197.90	193088.0
198.90	12442.0
200.00	1249.0
201.60	719.0
203.00	1939.0
204.00	9542.0
205.00	17016.0
206.00	59208.0
207.00	10977.0
207.90	2705.0
209.00	1454.0
209.90	1284.0
211.00	2831.0
211.90	396.0
212.80	172.0
214.90	831.0
215.90	1820.0
216.90	20232.0
218.00	2374.0
219.00	217.0
220.00	242.0
221.00	6979.0
221.90	2414.0
222.90	3642.0
224.00	35440.0
225.00	9417.0
226.10	1107.0
227.00	20872.0
228.00	2731.0
229.00	3642.0
230.00	496.0
231.00	1133.0
231.80	163.0
232.90	278.0
233.90	1463.0
234.90	1277.0
235.90	1060.0
237.00	1626.0
238.00	279.0
238.90	777.0
240.00	860.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

240.90	1131.0
241.90	2234.0
243.00	1866.0
244.00	28296.0
244.90	2957.0
245.90	6631.0
246.90	1237.0
248.90	841.0
250.00	250.0
250.90	227.0
251.80	270.0
252.90	781.0
254.10	1211.0
255.00	135040.0
255.90	19120.0
257.00	1743.0
258.00	10995.0
259.00	1835.0
259.90	342.0
260.90	338.0
264.00	238.0
265.00	4445.0
266.00	610.0
267.00	205.0
270.90	364.0
271.90	420.0
272.90	4350.0
274.00	10591.0
275.00	51464.0
276.00	7245.0
276.90	6485.0
278.00	912.0
278.90	290.0
281.00	822.0
282.00	237.0
283.00	745.0
284.00	736.0
285.00	1171.0
286.00	209.0
288.80	257.0
289.90	332.0
292.00	455.0
292.90	1252.0
293.90	491.0
295.00	366.0
296.00	22432.0
297.00	3482.0
297.90	406.0
301.00	170.0
301.90	453.0
302.90	2545.0
304.00	568.0
307.90	413.0
309.00	246.0
309.90	327.0
314.00	936.0
314.90	2935.0
315.90	967.0
316.80	184.0
320.90	618.0
321.80	276.0
323.00	5228.0
323.90	833.0
324.80	168.0
327.00	1293.0
328.00	744.0
332.00	770.0
332.80	398.0
333.20	483.0
333.90	4316.0
335.00	1461.0
340.90	757.0
341.90	190.0
345.80	1304.0
346.90	272.0
351.90	1616.0
353.00	1341.0
354.00	1943.0
354.90	420.0
358.90	218.0
364.90	11866.0
365.90	1754.0
369.90	189.0
370.70	384.0
372.00	2978.0
372.90	729.0
382.90	934.0
389.80	425.0
391.10	160.0
391.80	278.0
400.90	179.0
401.90	1087.0
403.00	1343.0
403.90	666.0
414.90	178.0
420.90	1128.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

422.00	1764.0
423.00	11460.0
424.00	2142.0
424.90	308.0
441.00	25072.0
442.00	192768.0
443.00	38128.0
444.00	2862.0
444.90	170.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

m/z	Abundance
36.90	281.0
38.00	701.0
39.00	3005.0
39.90	323.0
44.00	625.0
44.90	272.0
49.10	529.0
50.00	14015.0
51.00	44080.0
52.00	2273.0
54.90	164.0
56.00	1195.0
57.00	2872.0
60.90	837.0
61.90	837.0
63.00	1790.0
64.00	264.0
64.90	815.0
67.90	854.0
68.90	45912.0
69.90	293.0
73.00	1230.0
74.00	6167.0
75.00	8519.0
76.00	2692.0
77.00	53624.0
78.00	3823.0
79.00	3773.0
80.00	3027.0
81.00	4372.0
81.90	946.0
82.90	766.0
84.90	740.0
85.90	1772.0
86.90	626.0
88.10	162.0
90.90	1043.0
91.90	1134.0
93.00	9844.0
93.90	703.0
95.00	237.0
96.00	505.0
97.00	182.0
98.00	7548.0
98.90	5092.0
99.90	478.0
100.90	2086.0
101.80	182.0
102.90	647.0
103.90	1844.0
104.90	1491.0
105.90	479.0
107.00	18936.0
108.00	2748.0
109.00	489.0
109.90	26112.0
111.00	4576.0
111.90	590.0
112.90	211.0
115.00	238.0
116.00	1244.0
117.00	27232.0
118.00	2022.0
118.80	186.0
120.00	304.0
121.00	248.0
122.00	1332.0
123.00	2285.0
123.90	933.0
124.90	915.0
126.00	200.0
126.90	49472.0
128.00	3539.0
129.00	23744.0
129.90	2558.0
130.90	486.0
133.00	411.0
133.90	735.0
135.00	2723.0
136.00	1037.0
137.00	1154.0
137.80	246.0
138.90	193.0
139.90	258.0
141.00	4038.0
142.00	1182.0
142.90	635.0
143.90	160.0
144.90	253.0
145.90	933.0
147.00	2509.0
148.00	6863.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

148.90	800.0
151.00	404.0
152.00	300.0
153.00	1498.0
154.00	668.0
154.90	2254.0
156.00	3111.0
157.10	356.0
157.90	702.0
158.90	731.0
159.90	1355.0
160.90	1759.0
162.00	411.0
162.90	348.0
163.80	258.0
164.90	2010.0
165.90	1110.0
167.00	10072.0
168.00	5001.0
168.90	640.0
169.90	250.0
170.90	309.0
171.80	649.0
173.00	988.0
173.90	1629.0
174.90	3333.0
175.90	580.0
176.90	1628.0
177.90	625.0
179.00	5948.0
180.00	4061.0
181.00	1683.0
181.90	329.0
182.80	164.0
183.90	436.0
185.00	2628.0
186.00	19736.0
187.00	5455.0
188.00	728.0
189.00	2010.0
189.80	241.0
190.90	768.0
191.90	1809.0
193.00	2806.0
194.00	383.0
195.00	254.0
195.90	3238.0
197.00	645.0
197.90	110960.0
198.90	7532.0
199.90	597.0
201.30	559.0
202.80	1328.0
204.00	5297.0
205.00	9490.0
206.00	34024.0
207.00	7041.0
208.00	1891.0
209.00	745.0
210.10	602.0
211.00	1450.0
211.80	212.0
212.90	162.0
214.90	762.0
215.90	1019.0
216.90	12632.0
217.90	1456.0
221.00	4716.0
221.90	1519.0
222.90	2401.0
224.00	20976.0
225.00	5112.0
226.00	715.0
226.90	10861.0
228.00	1452.0
228.90	1669.0
230.00	370.0
231.00	735.0
231.90	168.0
232.90	337.0
233.90	957.0
234.90	862.0
236.00	435.0
236.90	1157.0
238.90	504.0
239.90	438.0
240.90	414.0
241.90	1300.0
243.00	965.0
244.00	16055.0
245.00	1787.0
246.00	4626.0
246.90	785.0
247.90	234.0
248.90	696.0
251.70	155.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351

252.90	466.0
255.00	81632.0
256.00	10732.0
257.00	670.0
258.00	5698.0
258.90	1046.0
265.00	2727.0
265.90	489.0
266.90	191.0
271.00	303.0
272.00	273.0
273.00	2815.0
274.00	7447.0
274.90	33832.0
276.00	3607.0
277.00	4753.0
278.00	835.0
280.90	806.0
281.80	209.0
283.00	547.0
283.90	419.0
284.90	738.0
288.80	211.0
293.00	1168.0
294.00	198.0
295.00	233.0
296.00	16082.0
297.00	1952.0
301.00	314.0
301.90	300.0
302.90	1584.0
303.80	341.0
307.90	208.0
314.00	580.0
315.00	1591.0
315.90	768.0
316.90	288.0
321.00	293.0
322.10	172.0
323.00	4030.0
324.00	613.0
326.90	846.0
327.90	435.0
332.00	217.0
333.00	351.0
333.90	2477.0
334.90	706.0
338.90	207.0
340.90	713.0
341.90	212.0
345.90	1359.0
346.90	290.0
351.00	162.0
352.00	1217.0
352.90	948.0
354.00	1124.0
354.90	434.0
364.90	6900.0
365.90	997.0
370.90	417.0
371.90	1926.0
373.00	605.0
382.90	534.0
389.80	171.0
390.80	165.0
391.20	194.0
400.80	169.0
401.90	1031.0
403.00	1514.0
403.80	581.0
420.90	1479.0
421.90	1119.0
423.00	5198.0
424.00	1287.0
424.90	272.0
440.90	22184.0
442.00	143424.0
443.00	28640.0
444.00	2216.0

Instrument :
BNA_N
ClientSampleId :
DFTPP351