

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071224\
 Data File : BN032631.D
 Acq On : 11 Jul 2024 12:17
 Operator : MA/JU
 Sample : SSTD3.212
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.2242

Quant Time: Jul 12 01:33:07 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN071224.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 12 01:27:09 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.570	152	4414	0.400	ng/ul	0.00
4) Naphthalene-d8	10.332	136	14638	0.400	ng/ul	-0.03
9) Acenaphthene-d10	14.207	164	9444	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.965	188	19565	0.400	ng/ul	-0.03
17) Chrysene-d12	21.179	240	15266	0.400	ng/ul	-0.01
23) Perylene-d12	23.389	264	13217	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.106	96	15970	2.731	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.139	88	17776	2.669	ng/ul#	89
14) Pentachlorophenol	16.636	266	13693	3.989	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071224\
Data File : BN032631.D
Acq On : 11 Jul 2024 12:17
Operator : MA/JU
Sample : SSTD3.212
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.2242

Quant Time: Jul 12 01:33:07 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN071224.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 12 01:27:09 2024
Response via : Initial Calibration

