

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071421\
 Data File : BN015566.D
 Acq On : 14 Jul 2021 16:37
 Operator : CG/JU
 Sample : M2953-08
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 EB-02-070621

Quant Time: Jul 14 17:06:46 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 14 11:20:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.818	152	14687	0.400	ng	# 0.00
7) Naphthalene-d8	10.597	136	68676	0.400	ng	0.00
13) Acenaphthene-d10	14.446	164	37901	0.400	ng	0.01
19) Phenanthrene-d10	17.188	188	71938	0.400	ng	0.00
29) Chrysene-d12	21.387	240	40983	0.400	ng	# 0.00
35) Perylene-d12	23.752	264	19139	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.420	112	573	0.015	ng	0.00
5) Phenol-d6	6.988	99	1045	0.022	ng	0.00
8) Nitrobenzene-d5	8.962	82	16378	0.426	ng	0.00
11) 2-Methylnaphthalene-d10	12.190	152	40082	0.352	ng	0.00
14) 2,4,6-Tribromophenol	15.944	330	502	0.250	ng	0.01
15) 2-Fluorobiphenyl	13.063	172	54514	0.354	ng	0.00
27) Fluoranthene-d10	19.232	212	76253	0.359	ng	0.00
31) Terphenyl-d14	19.832	244	57703	0.602	ng	0.00
Target Compounds						
34) Bis(2-ethylhexyl)phtha...	21.312	149	6619	0.281	ng	# 87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071421\
Data File : BN015566.D
Acq On : 14 Jul 2021 16:37
Operator : CG/JU
Sample : M2953-08
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
EB-02-070621

Quant Time: Jul 14 17:06:46 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 14 11:20:16 2021
Response via : Initial Calibration

