

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071421\
 Data File : BN015570.D
 Acq On : 14 Jul 2021 19:04
 Operator : CG/JU
 Sample : M2953-03DL 10X
 Misc :
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-11B-070621-FDDL

Manual Integrations
 APPROVED

mohammad
 7/15/2021 8:50:29 AM

Quant Time: Jul 15 01:05:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 14 11:20:16 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.818	152	15791	0.400	ng	# 0.00
7) Naphthalene-d8	10.597	136	74290	0.400	ng	0.00
13) Acenaphthene-d10	14.447	164	40273	0.400	ng	0.01
19) Phenanthrene-d10	17.188	188	79820	0.400	ng	0.00
29) Chrysene-d12	21.387	240	50197	0.400	ng	# 0.00
35) Perylene-d12	23.745	264	24678	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.420	112	309	0.008	ng	0.00
5) Phenol-d6	6.988	99	235	0.005	ng	0.00
8) Nitrobenzene-d5	8.962	82	1633	0.039	ng	0.00
11) 2-Methylnaphthalene-d10	12.190	152	4036	0.033	ng	0.00
14) 2,4,6-Tribromophenol	15.944	330	198	0.231	ng	0.01
15) 2-Fluorobiphenyl	13.064	172	5536	0.034	ng	0.00
27) Fluoranthene-d10	19.227	212	8527	0.036	ng	0.00
31) Terphenyl-d14	19.828	244	6036	0.051	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.383	88	18765m	2.535	ng	Qvalue
24) Pentachlorophenol	16.835	266	502	0.295	ng	# 59
34) Bis(2-ethylhexyl)phtha...	21.302	149	4631	0.223	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071421\
Data File : BN015570.D
Acq On : 14 Jul 2021 19:04
Operator : CG/JU
Sample : M2953-03DL 10X
Misc :
ALS Vial : 17 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-11B-070621-FDDL

Manual Integrations
APPROVED

mohammad
7/15/2021 8:50:29 AM

Quant Time: Jul 15 01:05:17 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN063021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 14 11:20:16 2021
Response via : Initial Calibration

