




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071421\
Data File : BN015554.D
Acq On : 14 Jul 2021 08:42
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 15 05:41:55 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 15 05:41:44 2021
Response via : Initial Calibration



TIC: BN015554.D\data.ms

(70) Pentachlorophenol (C)




16.839min (0.000) 32943.67 ng

response 184593

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	67.41
264.00	61.60	61.43
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071421\
Data File : BN015554.D
Acq On : 14 Jul 2021 08:42
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 15 05:41:55 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 15 05:41:44 2021
Response via : Initial Calibration



TIC: BN015554.D\data.ms

(77) Benzidine

19.445min (0.000) 0.00 ng

response 1079305

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.22
183.00	13.20	11.36
0.00	0.00	0.00