

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071425\
 Data File : BN037489.D
 Acq On : 14 Jul 2025 13:27
 Operator : RC/JU
 Sample : Q2580-02
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 WC-URBAN-FILL-B2

Quant Time: Jul 14 14:37:28 2025
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071125.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 11 15:38:34 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.732	152	2071	0.400	ng	0.00
7) Naphthalene-d8	10.509	136	5897	0.400	ng	0.00
13) Acenaphthene-d10	14.355	164	4587	0.400	ng	-0.01
19) Phenanthrene-d10	17.099	188	9509	0.400	ng	# 0.00
29) Chrysene-d12	21.295	240	11681	0.400	ng	0.00
35) Perylene-d12	23.531	264	9121	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.319	112	1524	0.316	ng	0.00
5) Phenol-d6	6.887	99	1995	0.329	ng	0.00
8) Nitrobenzene-d5	8.864	82	1351	0.316	ng	0.00
11) 2-Methylnaphthalene-d10	12.101	152	2485	0.310	ng	0.00
14) 2,4,6-Tribromophenol	15.845	330	690	0.444	ng	0.00
15) 2-Fluorobiphenyl	12.988	172	5409	0.207	ng	0.00
27) Fluoranthene-d10	19.136	212	7712	0.321	ng	0.00
31) Terphenyl-d14	19.736	244	8955	0.367	ng	0.00
Target Compounds						
						Qvalue
9) Naphthalene	10.562	128	369633	23.329	ng	96
12) 2-Methylnaphthalene	12.177	142	69368	6.990	ng	98
16) Acenaphthylene	14.077	152	129106	6.173	ng	99
17) Acenaphthene	14.420	154	254222	18.114	ng	94
18) Fluorene	15.403	166	312082	17.698	ng	99
23) Atrazine	16.565	200	81	0.027	ng	# 1
24) Pentachlorophenol	16.751	266	1405	0.498	ng	97
25) Phenanthrene	17.148	178	3314102	114.461	ng	99
26) Anthracene	17.148	178	3313866	126.822	ng	99
28) Fluoranthene	19.173	202	3564821	112.198	ng	99
30) Pyrene	19.531	202	3282425	69.895	ng	# 92
32) Benzo(a)anthracene	21.277	228	1590188	40.199	ng	97
33) Chrysene	21.331	228	1364959	31.567	ng	97
34) Bis(2-ethylhexyl)phtha...	21.223	149	7348	0.546	ng	95
36) Indeno(1,2,3-cd)pyrene	25.811	276	726238	20.514	ng	95
37) Benzo(b)fluoranthene	23.031	252	221760	6.252	ng	97
38) Benzo(k)fluoranthene	23.031	252	234539	6.523	ng	96
39) Benzo(a)pyrene	23.344	252	777628	27.242	ng	92
40) Dibenzo(a,h)anthracene	25.814	278	163248	5.717	ng	90
41) Benzo(g,h,i)perylene	26.504	276	803235	25.680	ng	93


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071425\
Data File : BN037489.D
Acq On : 14 Jul 2025 13:27
Operator : RC/JU
Sample : Q2580-02
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
WC-URBAN-FILL-B2


Quant Time: Jul 14 14:37:28 2025
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071125.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 11 15:38:34 2025
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.732 min Scan# 644
 Delta R.T. -0.000 min
 Lab File: BN037489.D
 Acq: 14 Jul 2025 13:27

Instrument :
 BNA_N
 ClientSampleId :
 WC-URBAN-FILL-B2



Tgt Ion:152 Resp: 2071




Ion	Ratio	Lower	Upper
152	100		
150	149.7	122.5	183.7
115	65.2	49.9	74.9




#4
 2-Fluorophenol
 Concen: 0.316 ng
 RT: 5.319 min Scan# 310
 Delta R.T. -0.000 min
 Lab File: BN037489.D
 Acq: 14 Jul 2025 13:27

Tgt Ion:112 Resp: 1524

Ion	Ratio	Lower	Upper
112	100		
64	65.7	45.4	68.0
63	32.2	24.7	37.1









#5
Phenol-d6
Concen: 0.329 ng
RT: 6.887 min Scan# 51
Delta R.T. -0.007 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27




Instrument :
BNA_N
ClientSampleId :
WC-URBAN-FILL-B2


Tgt Ion: 99 Resp: 1995
Ion Ratio Lower Upper
99 100
42 31.0 19.5 29.3#
71 39.4 29.4 44.2



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.509 min Scan# 930
Delta R.T. -0.000 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27

Tgt Ion: 136 Resp: 5897
Ion Ratio Lower Upper
136 100
137 12.6 10.0 15.0
54 10.2 9.0 13.6
68 6.4 6.1 9.1









#8
Nitrobenzene-d5
Concen: 0.316 ng
RT: 8.864 min Scan# 776
Delta R.T. -0.000 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27




Instrument :
BNA_N
ClientSampleId :
WC-URBAN-FILL-B2


Tgt Ion: 82 Resp: 1351
Ion Ratio Lower Upper
82 100
128 45.8 33.1 49.7
54 64.7 53.0 79.4



#9
Naphthalene
Concen: 23.329 ng
RT: 10.562 min Scan# 935
Delta R.T. -0.000 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27

Tgt Ion: 128 Resp: 369633
Ion Ratio Lower Upper
128 100
129 11.0 10.2 15.2
127 13.1 11.8 17.6









#11
2-Methylnaphthalene-d10
Concen: 0.310 ng
RT: 12.101 min Scan# 1140
Delta R.T. -0.005 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27




Instrument :
BNA_N
ClientSampleId :
WC-URBAN-FILL-B2


Tgt Ion:152 Resp: 2485
Ion Ratio Lower Upper
152 100
151 21.3 17.0 25.6



#12
2-Methylnaphthalene
Concen: 6.990 ng
RT: 12.177 min Scan# 1155
Delta R.T. -0.005 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27

Tgt Ion:142 Resp: 69368
Ion Ratio Lower Upper
142 100
141 89.0 70.6 105.8
115 40.5 30.0 45.0









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.355 min Scan# 1446
Delta R.T. -0.011 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27




Instrument :
BNA_N
ClientSampleId :
WC-URBAN-FILL-B2


Tgt Ion:164 Resp: 4587
Ion Ratio Lower Upper
164 100
162 103.7 81.1 121.7
160 52.5 46.5 69.7



#14
2,4,6-Tribromophenol
Concen: 0.444 ng
RT: 15.845 min Scan# 1580
Delta R.T. -0.000 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27

Tgt Ion:330 Resp: 690
Ion Ratio Lower Upper
330 100
332 93.6 82.0 123.0
141 95.2 36.2 54.2#









#15
2-Fluorobiphenyl
Concen: 0.207 ng
RT: 12.988 min Scan# 1315
Delta R.T. -0.000 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27




Instrument :
BNA_N
ClientSampleId :
WC-URBAN-FILL-B2


Tgt Ion:172	Resp:	5409
Ion Ratio	Lower	Upper
172	100	
171	36.1	29.8 44.8
170	24.1	20.0 30.0



#16
Acenaphthylene
Concen: 6.173 ng
RT: 14.077 min Scan# 1419
Delta R.T. -0.000 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27

Tgt Ion:152	Resp:	129106
Ion Ratio	Lower	Upper
152	100	
151	20.0	16.4 24.6
153	13.6	10.6 15.8









#17
Acenaphthene
Concen: 18.114 ng
RT: 14.420 min Scan# 1451
Delta R.T. -0.000 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27




Instrument :
BNA_N
ClientSampleId :
WC-URBAN-FILL-B2


Tgt Ion:154 Resp: 254222
Ion Ratio Lower Upper
154 100
153 113.9 89.4 134.0
152 56.5 53.2 79.8



#18
Fluorene
Concen: 17.698 ng
RT: 15.403 min Scan# 1543
Delta R.T. -0.011 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27

Tgt Ion:166 Resp: 312082
Ion Ratio Lower Upper
166 100
165 98.3 79.4 119.2
167 14.2 10.6 15.8









#19
 Phenanthrene-d10
 Concen: 0.400 ng
 RT: 17.099 min Scan# 1681
 Delta R.T. -0.000 min
 Lab File: BN037489.D
 Acq: 14 Jul 2025 13:27




Instrument : BNA_N
 ClientSampleId : WC-URBAN-FILL-B2


Tgt Ion:	188	Resp:	9509
Ion Ratio	100	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	14.5	9.3	13.9



#23
 Atrazine
 Concen: 0.027 ng
 RT: 16.565 min Scan# 1638
 Delta R.T. 0.000 min
 Lab File: BN037489.D
 Acq: 14 Jul 2025 13:27

Tgt Ion:	200	Resp:	81
Ion Ratio	100	Lower	Upper
200	100		
173	158.3	34.3	51.5
215	83.3	41.6	62.4





#24

Pentachlorophenol

Concen: 0.498 ng

RT: 16.751 min Scan# 1

Delta R.T. 0.000 min

Lab File: BN037489.D

Acq: 14 Jul 2025 13:27

Instrument :

BNA_N

ClientSampleId :

WC-URBAN-FILL-B2





Tgt Ion:266 Resp: 1405

Ion Ratio Lower Upper

266 100

264 64.6 47.5 71.3

268 63.3 50.6 76.0



#25

Phenanthrene

Concen: 114.461 ng

RT: 17.148 min Scan# 1685

Delta R.T. 0.012 min

Lab File: BN037489.D

Acq: 14 Jul 2025 13:27




Tgt Ion:178 Resp: 3314102


Ion Ratio Lower Upper

178 100

176 18.9 15.1 22.7

179 16.1 12.2 18.2









#26
 Anthracene
 Concen: 126.822 ng
 RT: 17.148 min Scan# 1685
 Delta R.T. -0.087 min
 Lab File: BN037489.D
 Acq: 14 Jul 2025 13:27




Instrument :
 BNA_N
 ClientSampleId :
 WC-URBAN-FILL-B2


Tgt Ion:178 Resp: 3313866
 Ion Ratio Lower Upper
 178 100
 176 18.9 14.9 22.3
 179 16.1 12.3 18.5



#27
 Fluoranthene-d10
 Concen: 0.321 ng
 RT: 19.136 min Scan# 1929
 Delta R.T. 0.005 min
 Lab File: BN037489.D
 Acq: 14 Jul 2025 13:27

Tgt Ion:212 Resp: 7712
 Ion Ratio Lower Upper
 212 100
 106 18.0 11.7 17.5#
 104 12.3 6.6 10.0#









#28
Fluoranthene
Concen: 112.198 ng
RT: 19.173 min Scan# 1937
Delta R.T. 0.009 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27




Instrument :
BNA_N
ClientSampleId :
WC-URBAN-FILL-B2


Tgt Ion:202 Resp: 3564821
Ion Ratio Lower Upper
202 100
101 12.1 9.2 13.8
203 18.1 14.0 21.0



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.295 min Scan# 2260
Delta R.T. 0.009 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27

Tgt Ion:240 Resp: 11681
Ion Ratio Lower Upper
240 100
120 16.6 12.4 18.6
236 31.0 22.9 34.3









#30
Pyrene
Concen: 69.895 ng
RT: 19.531 min Scan# 2013
Delta R.T. 0.005 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27




Instrument :
BNA_N
ClientSampleId :
WC-URBAN-FILL-B2


Tgt Ion:202 Resp: 3282425
Ion Ratio Lower Upper
202 100
200 21.3 16.4 24.6
203 24.3 14.3 21.5#



#31
Terphenyl-d14
Concen: 0.367 ng
RT: 19.736 min Scan# 2058
Delta R.T. -0.005 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27

Tgt Ion:244 Resp: 8955
Ion Ratio Lower Upper
244 100
212 10.6 8.7 13.1
122 20.6 14.1 21.1









#32
Benzo(a)anthracene
Concen: 40.199 ng
RT: 21.277 min Scan# 2126
Delta R.T. 0.009 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27




Instrument :
BNA_N
ClientSampleId :
WC-URBAN-FILL-B2


Tgt Ion:228 Resp: 1590188
Ion Ratio Lower Upper
228 100
226 30.3 22.7 34.1
229 21.1 16.6 24.8



#33
Chrysene
Concen: 31.567 ng
RT: 21.331 min Scan# 2264
Delta R.T. 0.009 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27

Tgt Ion:228 Resp: 1364959
Ion Ratio Lower Upper
228 100
226 29.4 24.6 36.8
229 22.6 16.4 24.6









#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.546 ng
 RT: 21.223 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: BN037489.D
 Acq: 14 Jul 2025 13:27




Instrument :
 BNA_N
 ClientSampleId :
 WC-URBAN-FILL-B2


Tgt Ion:149 Resp: 7348
 Ion Ratio Lower Upper
 149 100
 167 26.9 19.4 29.2
 279 3.6 2.5 3.7



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.531 min Scan# 2769
 Delta R.T. 0.003 min
 Lab File: BN037489.D
 Acq: 14 Jul 2025 13:27

Tgt Ion:264 Resp: 9121
 Ion Ratio Lower Upper
 264 100
 260 29.1 21.7 32.5
 265 88.7 41.6 62.4#





#36

Indeno(1,2,3-cd)pyrene

Concen: 20.514 ng

RT: 25.811 min Scan# 31

Delta R.T. 0.012 min

Lab File: BN037489.D


Acq: 14 Jul 2025 13:27

Instrument :

BNA_N

ClientSampleId :

WC-URBAN-FILL-B2






Tgt Ion:276 Resp: 726238

Ion Ratio Lower Upper

276 100

138 24.9 23.4 35.2

277 24.3 19.8 29.8



#37

Benzo(b)fluoranthene

Concen: 6.252 ng

RT: 23.031 min Scan# 2598

Delta R.T. 0.175 min

Lab File: BN037489.D

Acq: 14 Jul 2025 13:27




Tgt Ion:252 Resp: 221760


Ion Ratio Lower Upper

252 100

253 25.8 20.1 30.1

125 14.8 13.7 20.5









#38
Benzo(k)fluoranthene
Concen: 6.523 ng
RT: 23.031 min Scan# 21
Delta R.T. 0.131 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27




Instrument :
BNA_N
ClientSampleId :
WC-URBAN-FILL-B2


Tgt Ion:252 Resp: 234539
Ion Ratio Lower Upper
252 100
253 25.8 20.4 30.6
125 14.8 14.6 21.8



#39
Benzo(a)pyrene
Concen: 27.242 ng
RT: 23.344 min Scan# 2705
Delta R.T. -0.085 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27


Tgt Ion:252 Resp: 777628
Ion Ratio Lower Upper
252 100
253 23.4 21.8 32.6
125 17.0 16.7 25.1








#40
Dibenzo(a,h)anthracene
Concen: 5.717 ng
RT: 25.814 min Scan# 31
Delta R.T. -0.006 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27

Instrument :
BNA_N
ClientSampleId :
WC-URBAN-FILL-B2



Tgt Ion:278 Resp: 163248
Ion Ratio Lower Upper
278 100
139 20.9 19.9 29.9
279 25.1 24.8 37.2



#41
Benzo(g,h,i)perylene
Concen: 25.680 ng
RT: 26.504 min Scan# 3786
Delta R.T. 0.015 min
Lab File: BN037489.D
Acq: 14 Jul 2025 13:27

Tgt Ion:276 Resp: 803235
Ion Ratio Lower Upper
276 100
277 24.1 21.7 32.5
138 25.7 24.3 36.5

