

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071521\
 Data File : BN015575.D
 Acq On : 15 Jul 2021 11:36
 Operator : CG/JU
 Sample : SSTDICCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICCC0.4

Quant Time: Jul 15 12:21:14 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN071521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 15 12:19:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.808	152	13667	0.400	ng	# 0.00
7) Naphthalene-d8	10.584	136	60406	0.400	ng	0.00
13) Acenaphthene-d10	14.434	164	30840	0.400	ng	0.00
19) Phenanthrene-d10	17.176	188	55241	0.400	ng	0.00
29) Chrysene-d12	21.376	240	50354	0.400	ng	# 0.00
35) Perylene-d12	23.724	264	45767	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.393	112	15685	0.450	ng	0.00
5) Phenol-d6	6.979	99	18580	0.424	ng	0.00
8) Nitrobenzene-d5	8.949	82	15307	0.452	ng	0.00
11) 2-Methylnaphthalene-d10	12.176	152	37153	0.371	ng	0.00
14) 2,4,6-Tribromophenol	15.918	330	3711	0.449	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	47565	0.380	ng	0.00
27) Fluoranthene-d10	19.217	212	60085	0.368	ng	0.00
31) Terphenyl-d14	19.818	244	45265	0.384	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.364	88	3079	0.481	ng	# 26
3) n-Nitrosodimethylamine	3.714	42	4239	0.353	ng	# 76
6) bis(2-Chloroethyl)ether	7.237	93	16425	0.381	ng	# 84
9) Naphthalene	10.635	128	67662	0.390	ng	97
10) Hexachlorobutadiene	10.926	225	12552	0.385	ng	# 98
12) 2-Methylnaphthalene	12.252	142	43227	0.389	ng	# 89
16) Acenaphthylene	14.142	152	51936	0.380	ng	98
17) Acenaphthene	14.485	154	36276	0.379	ng	99
18) Fluorene	15.474	166	43746	0.375	ng	99
20) 4,6-Dinitro-2-methylph...	15.576	198	121	0.114	ng	89
21) 4-Bromophenyl-phenylether	16.373	248	13106	0.374	ng	90
22) Hexachlorobenzene	16.494	284	15154	0.381	ng	# 92
23) Atrazine	16.640	200	8387	0.412	ng	94
24) Pentachlorophenol	16.835	266	1368	0.231	ng	99
25) Phenanthrene	17.212	178	69284	0.389	ng	98
26) Anthracene	17.310	178	55906	0.367	ng	99
28) Fluoranthene	19.247	202	74140	0.388	ng	# 97
30) Pyrene	19.609	202	73398	0.383	ng	99
32) Benzo(a)anthracene	21.355	228	65750	0.381	ng	99
33) Chrysene	21.408	228	71267	0.379	ng	98
34) Bis(2-ethylhexyl)phtha...	21.302	149	27353	0.482	ng	# 88
36) Indeno(1,2,3-cd)pyrene	26.134	276	58449	0.357	ng	# 86
37) Benzo(b)fluoranthene	23.012	252	68178	0.380	ng	# 95
38) Benzo(k)fluoranthene	23.057	252	71662	0.374	ng	# 94
39) Benzo(a)pyrene	23.618	252	57701	0.352	ng	# 94
40) Dibenzo(a,h)anthracene	26.154	278	45458	0.353	ng	# 91
41) Benzo(g,h,i)perylene	26.873	276	49442	0.348	ng	# 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071521\
Data File : BN015575.D
Acq On : 15 Jul 2021 11:36
Operator : CG/JU
Sample : SSTDICCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Quant Time: Jul 15 12:21:14 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN071521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 15 12:19:46 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.808 min Scan# 513
 Delta R.T. 0.000 min
 Lab File: BN015575.D
 Acq: 15 Jul 2021 11:36




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4


Tgt Ion: 152 Resp: 13667
 Ion Ratio Lower Upper
 152 100
 150 154.4 119.0 178.6
 115 59.0 62.2 93.4#



#2
 1,4-Dioxane
 Concen: 0.481 ng
 RT: 3.364 min Scan# 31
 Delta R.T. 0.000 min
 Lab File: BN015575.D
 Acq: 15 Jul 2021 11:36

Tgt Ion: 88 Resp: 3079
 Ion Ratio Lower Upper
 88 100
 43 107.4 12.9 19.3#
 58 65.5 38.1 57.1#









#3
n-Nitrosodimethylamine
Concen: 0.353 ng
RT: 3.714 min Scan# 69
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 42 Resp: 4239
Ion Ratio Lower Upper
42 100
74 151.2 149.5 224.3
44 10.6 4.2 6.2#



#4
2-Fluorophenol
Concen: 0.450 ng
RT: 5.393 min Scan# 251
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36

Tgt Ion: 112 Resp: 15685
Ion Ratio Lower Upper
112 100
64 56.1 39.0 58.6
63 27.7 20.6 30.8









#5
Phenol-d6
Concen: 0.424 ng
RT: 6.979 min Scan# 423
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion: 99 Resp: 18580
Ion Ratio Lower Upper
99 100
42 16.7 8.9 13.3#
71 30.1 35.6 53.4#



#6
bis(2-Chloroethyl)ether
Concen: 0.381 ng
RT: 7.237 min Scan# 451
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36

Tgt Ion: 93 Resp: 16425
Ion Ratio Lower Upper
93 100
63 74.8 46.6 70.0#
95 32.8 28.6 42.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 747
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	136	Resp:	60406
Ion Ratio	Lower	Upper	
136	100		
137	10.8	9.8	14.8
54	7.8	5.4	8.2
68	6.5	4.9	7.3



#8
Nitrobenzene-d5
Concen: 0.452 ng
RT: 8.949 min Scan# 618
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36

Tgt Ion:	82	Resp:	15307
Ion Ratio	Lower	Upper	
82	100		
128	37.8	31.6	47.4
54	50.4	29.0	43.4





#9

Naphthalene

Concen: 0.390 ng

RT: 10.635 min Scan# 751

Delta R.T. 0.000 min

Lab File: BN015575.D

Acq: 15 Jul 2021 11:36





Tgt Ion:	128	Resp:	67662
Ion Ratio	Lower	Upper	
128	100		
129	10.9	9.1	13.7
127	13.1	11.8	17.6

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4



#10

Hexachlorobutadiene

Concen: 0.385 ng




RT: 10.926 min Scan# 774


Delta R.T. 0.000 min

Lab File: BN015575.D

Acq: 15 Jul 2021 11:36

Tgt Ion:	225	Resp:	12552
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	64.3	50.2	75.2









#11
2-Methylnaphthalene-d10
Concen: 0.371 ng
RT: 12.176 min Scan# 970
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:152 Resp: 37153
Ion Ratio Lower Upper
152 100
151 20.4 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.389 ng
RT: 12.252 min Scan# 987
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36

Tgt Ion:142 Resp: 43227
Ion Ratio Lower Upper
142 100
141 88.6 72.3 108.5
115 32.8 40.7 61.1#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.434 min Scan# 1196
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:164 Resp: 30840
Ion Ratio Lower Upper
164 100
162 98.0 83.8 125.6
160 43.0 37.4 56.0



#14
2,4,6-Tribromophenol
Concen: 0.449 ng
RT: 15.918 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36

Tgt Ion:330 Resp: 3711
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 36.5 22.4 33.6#









#15
2-Fluorobiphenyl
Concen: 0.380 ng
RT: 13.051 min Scan# 1087
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:172 Resp: 47565
Ion Ratio Lower Upper
172 100
171 35.4 30.6 45.8
170 24.0 20.2 30.4



#16
Acenaphthylene
Concen: 0.380 ng
RT: 14.142 min Scan# 1173
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36

Tgt Ion:152 Resp: 51936
Ion Ratio Lower Upper
152 100
151 19.3 16.7 25.1
153 13.0 10.5 15.7









#17
Acenaphthene
Concen: 0.379 ng
RT: 14.485 min Scan# 1200
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:154 Resp: 36276
Ion Ratio Lower Upper
154 100
153 113.3 90.9 136.3
152 55.4 45.8 68.8



#18
Fluorene
Concen: 0.375 ng
RT: 15.474 min Scan# 1278
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36

Tgt Ion:166 Resp: 43746
Ion Ratio Lower Upper
166 100
165 98.0 79.2 118.8
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.176 min Scan# 1416
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:	188	Resp:	55241
Ion Ratio	Lower	Upper	
188	100		
94	0.0	0.0	0.0
80	9.9	9.0	13.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.114 ng
RT: 15.576 min Scan# 1286
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36

Tgt Ion:	198	Resp:	121
Ion Ratio	Lower	Upper	
198	100		
182	50.0	34.5	51.7
77	0.0	0.0	0.0









#21
4-Bromophenyl-phenylether
Concen: 0.374 ng
RT: 16.373 min Scan# 1350
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:248 Resp: 13106
Ion Ratio Lower Upper
248 100
250 96.9 86.2 129.4
141 51.9 36.2 54.2



#22
Hexachlorobenzene
Concen: 0.381 ng
RT: 16.494 min Scan# 1360
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36

Tgt Ion:284 Resp: 15154
Ion Ratio Lower Upper
284 100
142 34.2 22.2 33.2#
249 31.0 26.6 39.8









#23
Atrazine
Concen: 0.412 ng
RT: 16.640 min Scan# 1372
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:200 Resp: 8387
Ion Ratio Lower Upper
200 100
173 27.0 21.7 32.5
215 47.8 43.1 64.7



#24
Pentachlorophenol
Concen: 0.231 ng
RT: 16.835 min Scan# 1388
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36

Tgt Ion:266 Resp: 1368
Ion Ratio Lower Upper
266 100
264 62.6 51.0 76.4
268 62.2 50.6 76.0









#25
Phenanthrene
Concen: 0.389 ng
RT: 17.212 min Scan# 1419
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:178 Resp: 69284
Ion Ratio Lower Upper
178 100
176 18.7 15.6 23.4
179 15.8 12.1 18.1



#26
Anthracene
Concen: 0.367 ng
RT: 17.310 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36

Tgt Ion:178 Resp: 55906
Ion Ratio Lower Upper
178 100
176 18.2 15.1 22.7
179 15.4 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.368 ng
RT: 19.217 min Scan# 1727
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:212 Resp: 60085
Ion Ratio Lower Upper
212 100
106 14.2 7.0 10.6#
104 7.8 4.1 6.1#



#28
Fluoranthene
Concen: 0.388 ng
RT: 19.247 min Scan# 1733
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36

Tgt Ion:202 Resp: 74140
Ion Ratio Lower Upper
202 100
101 11.6 6.6 10.0#
203 17.3 13.9 20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.376 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:240 Resp: 50354
Ion Ratio Lower Upper
240 100
120 8.5 4.2 6.4#
236 26.2 21.4 32.2



#30
Pyrene
Concen: 0.383 ng
RT: 19.609 min Scan# 1806
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36

Tgt Ion:202 Resp: 73398
Ion Ratio Lower Upper
202 100
200 20.5 16.9 25.3
203 17.4 13.9 20.9









#31
Terphenyl-d14
Concen: 0.384 ng
RT: 19.818 min Scan# 1848
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:244 Resp: 45265
Ion Ratio Lower Upper
244 100
212 7.6 6.3 9.5
122 12.3 6.2 9.2#



#32
Benzo(a)anthracene
Concen: 0.381 ng
RT: 21.355 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36

Tgt Ion:228 Resp: 65750
Ion Ratio Lower Upper
228 100
226 27.9 23.0 34.6
229 19.2 15.7 23.5





#33

Chrysene

Concen: 0.379 ng

RT: 21.408 min Scan# 2018

Delta R.T. 0.000 min

Lab File: BN015575.D

Acq: 15 Jul 2021 11:36

Tgt Ion:228 Resp: 71267

Ion Ratio Lower Upper

228 100

226 30.3 25.0 37.6





229 19.2 15.9 23.9

Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.482 ng

RT: 21.302 min Scan# 2008

Delta R.T. 0.000 min

Lab File: BN015575.D

Acq: 15 Jul 2021 11:36




Tgt Ion:149 Resp: 27353


Ion Ratio Lower Upper

149 100

167 28.3 28.9 43.3#

279 3.7 4.9 7.3#









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.724 min Scan# 2497
 Delta R.T. 0.000 min
 Lab File: BN015575.D
 Acq: 15 Jul 2021 11:36




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.4


Tgt Ion:	264	Resp:	45767
Ion Ratio	Lower	Upper	
264	100		
260	25.2	20.6	31.0
265	24.2	21.4	32.2



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.357 ng
 RT: 26.134 min Scan# 3201
 Delta R.T. 0.000 min
 Lab File: BN015575.D
 Acq: 15 Jul 2021 11:36

Tgt Ion:	276	Resp:	58449
Ion Ratio	Lower	Upper	
276	100		
138	27.8	10.0	15.0#
277	24.9	20.4	30.6









#37
Benzo(b)fluoranthene
Concen: 0.380 ng
RT: 23.012 min Scan# 2289
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36




Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4


Tgt Ion:252 Resp: 68178
Ion Ratio Lower Upper
252 100
253 22.2 18.5 27.7
125 11.9 5.8 8.6#



#38
Benzo(k)fluoranthene
Concen: 0.374 ng
RT: 23.057 min Scan# 2302
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36

Tgt Ion:252 Resp: 71662
Ion Ratio Lower Upper
252 100
253 21.9 19.1 28.7
125 11.2 5.4 8.0#





#39

Benzo(a)pyrene

Concen: 0.352 ng

RT: 23.618 min Scan# 2466

Delta R.T. 0.000 min

Lab File: BN015575.D

Acq: 15 Jul 2021 11:36





Instrument :

BNA_N

ClientSampleId :

SSTDICCC0.4

Tgt Ion:252	Resp:	57701
Ion Ratio	Lower	Upper
252	100	
253	22.3	19.4 29.2
125	13.5	7.4 11.2#



#40

Dibenzo(a,h)anthracene

Concen: 0.353 ng




RT: 26.154 min Scan# 3207


Delta R.T. 0.000 min

Lab File: BN015575.D

Acq: 15 Jul 2021 11:36

Tgt Ion:278	Resp:	45458
Ion Ratio	Lower	Upper
278	100	
139	18.3	9.0 13.4#
279	23.8	21.0 31.4





#41
Benzo(g,h,i)perylene
Concen: 0.348 ng
RT: 26.873 min Scan# 3417
Delta R.T. 0.000 min
Lab File: BN015575.D
Acq: 15 Jul 2021 11:36

Instrument :
BNA_N
ClientSampleId :
SSTDICCC0.4

Tgt Ion:276 Resp: 49442
Ion Ratio Lower Upper
276 100
277 23.8 19.7 29.5
138 23.5 9.8 14.6#

