

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071521\
 Data File : BN015583.D
 Acq On : 15 Jul 2021 17:47
 Operator : CG/JU
 Sample : PB137284BSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB137284BSD

Quant Time: Jul 15 18:47:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN071521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 15 15:46:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.809	152	14390	0.400	ng	# 0.00
7) Naphthalene-d8	10.584	136	62880	0.400	ng	# 0.00
13) Acenaphthene-d10	14.434	164	32443	0.400	ng	0.00
19) Phenanthrene-d10	17.176	188	58810	0.400	ng	0.00
29) Chrysene-d12	21.376	240	53284	0.400	ng	# 0.00
35) Perylene-d12	23.724	264	51178	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.393	112	16318	0.395	ng	0.00
5) Phenol-d6	6.979	99	19455	0.397	ng	0.00
8) Nitrobenzene-d5	8.949	82	16612	0.367	ng	0.00
11) 2-Methylnaphthalene-d10	12.181	152	38829	0.375	ng	0.00
14) 2,4,6-Tribromophenol	15.918	330	4069	0.380	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	50260	0.384	ng	0.00
27) Fluoranthene-d10	19.217	212	62831	0.354	ng	0.00
31) Terphenyl-d14	19.818	244	47877	0.379	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.364	88	3279	0.466	ng	# 51
3) n-Nitrosodimethylamine	3.705	42	4580	0.386	ng	# 77
6) bis(2-Chloroethyl)ether	7.237	93	17419	0.397	ng	# 84
9) Naphthalene	10.635	128	69716	0.382	ng	97
10) Hexachlorobutadiene	10.926	225	13363	0.392	ng	# 98
12) 2-Methylnaphthalene	12.252	142	45196	0.390	ng	# 90
16) Acenaphthylene	14.142	152	54972	0.364	ng	98
17) Acenaphthene	14.497	154	38327	0.380	ng	98
18) Fluorene	15.487	166	45445	0.376	ng	99
20) 4,6-Dinitro-2-methylph...	15.563	198	376	0.617	ng	# 80
21) 4-Bromophenyl-phenylether	16.373	248	13852	0.371	ng	# 86
22) Hexachlorobenzene	16.494	284	16000	0.381	ng	# 92
23) Atrazine	16.640	200	9052	0.365	ng	93
24) Pentachlorophenol	16.835	266	2064	0.462	ng	99
25) Phenanthrene	17.225	178	71725	0.377	ng	99
26) Anthracene	17.310	178	59166	0.363	ng	99
28) Fluoranthene	19.247	202	78301	0.380	ng	# 97
30) Pyrene	19.609	202	77762	0.380	ng	99
32) Benzo(a)anthracene	21.366	228	69975	0.376	ng	96
33) Chrysene	21.408	228	76278	0.387	ng	99
34) Bis(2-ethylhexyl)phtha...	21.302	149	27815	0.349	ng	# 87
36) Indeno(1,2,3-cd)pyrene	26.141	276	68783	0.376	ng	# 85
37) Benzo(b)fluoranthene	23.012	252	75622	0.377	ng	# 95
38) Benzo(k)fluoranthene	23.060	252	78548	0.388	ng	# 94
39) Benzo(a)pyrene	23.622	252	65275	0.378	ng	# 92
40) Dibenzo(a,h)anthracene	26.158	278	54128	0.377	ng	# 90
41) Benzo(g,h,i)perylene	26.877	276	57349	0.370	ng	# 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071521\
 Data File : BN015583.D
 Acq On : 15 Jul 2021 17:47
 Operator : CG/JU
 Sample : PB137284BSD
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB137284BSD

Quant Time: Jul 15 18:47:02 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN071521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 15 15:46:51 2021
 Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.809 min Scan# 513
Delta R.T. 0.001 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:152 Resp: 14390
Ion Ratio Lower Upper
152 100
150 154.9 119.0 178.6
115 59.4 62.2 93.4#



#2
1,4-Dioxane
Concen: 0.466 ng
RT: 3.364 min Scan# 31
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Tgt Ion: 88 Resp: 3279
Ion Ratio Lower Upper
88 100
43 62.9 12.9 19.3#
58 66.9 38.1 57.1#









#3
n-Nitrosodimethylamine
Concen: 0.386 ng
RT: 3.705 min Scan# 68
Delta R.T. -0.009 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion: 42 Resp: 4580
Ion Ratio Lower Upper
42 100
74 152.7 149.5 224.3
44 7.2 4.2 6.2#



#4
2-Fluorophenol
Concen: 0.395 ng
RT: 5.393 min Scan# 251
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Tgt Ion: 112 Resp: 16318
Ion Ratio Lower Upper
112 100
64 55.8 39.0 58.6
63 27.8 20.6 30.8









#5
Phenol-d6
Concen: 0.397 ng
RT: 6.979 min Scan# 423
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion: 99 Resp: 19455
Ion Ratio Lower Upper
99 100
42 16.6 8.9 13.3#
71 30.0 35.6 53.4#



#6
bis(2-Chloroethyl)ether
Concen: 0.397 ng
RT: 7.237 min Scan# 451
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Tgt Ion: 93 Resp: 17419
Ion Ratio Lower Upper
93 100
63 74.9 46.6 70.0#
95 33.0 28.6 42.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 747
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:	136	Resp:	62880
Ion Ratio	Lower	Upper	
136	100		
137	11.0	9.8	14.8
54	9.0	5.4	8.2#
68	7.3	4.9	7.3



#8
Nitrobenzene-d5
Concen: 0.367 ng
RT: 8.949 min Scan# 618
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Tgt Ion:	82	Resp:	16612
Ion Ratio	Lower	Upper	
82	100		
128	36.4	31.6	47.4
54	51.8	29.0	43.4#









#9
Naphthalene
Concen: 0.382 ng
RT: 10.635 min Scan# 751
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:128 Resp: 69716
Ion Ratio Lower Upper
128 100
129 10.8 9.1 13.7
127 13.2 11.8 17.6



#10
Hexachlorobutadiene
Concen: 0.392 ng
RT: 10.926 min Scan# 774
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Tgt Ion:225 Resp: 13363
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.9 50.2 75.2









#11
2-Methylnaphthalene-d10
Concen: 0.375 ng
RT: 12.181 min Scan# 971
Delta R.T. 0.004 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:152 Resp: 38829
Ion Ratio Lower Upper
152 100
151 20.4 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.390 ng
RT: 12.252 min Scan# 987
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Tgt Ion:142 Resp: 45196
Ion Ratio Lower Upper
142 100
141 88.6 72.3 108.5
115 33.2 40.7 61.1#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.434 min Scan# 1196
Delta R.T. -0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:164 Resp: 32443
Ion Ratio Lower Upper
164 100
162 98.6 83.8 125.6
160 44.1 37.4 56.0



#14
2,4,6-Tribromophenol
Concen: 0.380 ng
RT: 15.918 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Tgt Ion:330 Resp: 4069
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 37.6 22.4 33.6#









#15
2-Fluorobiphenyl
Concen: 0.384 ng
RT: 13.051 min Scan# 1087
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:	172	Resp:	50260
Ion Ratio	Lower	Upper	
172	100		
171	35.5	30.6	45.8
170	24.1	20.2	30.4



#16
Acenaphthylene
Concen: 0.364 ng
RT: 14.142 min Scan# 1173
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Tgt Ion:	152	Resp:	54972
Ion Ratio	Lower	Upper	
152	100		
151	19.4	16.7	25.1
153	13.1	10.5	15.7









#17
Acenaphthene
Concen: 0.380 ng
RT: 14.497 min Scan# 1201
Delta R.T. 0.013 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:154 Resp: 38327
Ion Ratio Lower Upper
154 100
153 112.0 90.9 136.3
152 55.1 45.8 68.8



#18
Fluorene
Concen: 0.376 ng
RT: 15.487 min Scan# 1279
Delta R.T. 0.013 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Tgt Ion:166 Resp: 45445
Ion Ratio Lower Upper
166 100
165 97.7 79.2 118.8
167 13.4 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.176 min Scan# 1416
Delta R.T. -0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:188 Resp: 58810
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.0 9.0 13.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.617 ng
RT: 15.563 min Scan# 1285
Delta R.T. -0.013 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Tgt Ion:198 Resp: 376
Ion Ratio Lower Upper
198 100
182 30.1 34.5 51.7#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.371 ng
RT: 16.373 min Scan# 1350
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:248 Resp: 13852
Ion Ratio Lower Upper
248 100
250 96.4 86.2 129.4
141 59.2 36.2 54.2#



#22
Hexachlorobenzene
Concen: 0.381 ng
RT: 16.494 min Scan# 1360
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Tgt Ion:284 Resp: 16000
Ion Ratio Lower Upper
284 100
142 34.9 22.2 33.2#
249 31.2 26.6 39.8





#23

Atrazine

Concen: 0.365 ng

RT: 16.640 min Scan# 1372

Delta R.T. 0.000 min

Lab File: BN015583.D

Acq: 15 Jul 2021 17:47





Tgt Ion:	200	Resp:	9052
Ion Ratio	Lower	Upper	
200	100		
173	28.3	21.7	32.5
215	47.4	43.1	64.7

Instrument :

BNA_N

ClientSampleId :

PB137284BSD



#24

Pentachlorophenol

Concen: 0.462 ng




RT: 16.835 min Scan# 1388


Delta R.T. 0.000 min

Lab File: BN015583.D

Acq: 15 Jul 2021 17:47

Tgt Ion:	266	Resp:	2064
Ion Ratio	Lower	Upper	
266	100		
264	62.0	51.0	76.4
268	63.8	50.6	76.0









#25
Phenanthrene
Concen: 0.377 ng
RT: 17.225 min Scan# 1420
Delta R.T. 0.012 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:178 Resp: 71725
Ion Ratio Lower Upper
178 100
176 18.8 15.6 23.4
179 15.7 12.1 18.1



#26
Anthracene
Concen: 0.363 ng
RT: 17.310 min Scan# 1427
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Tgt Ion:178 Resp: 59166
Ion Ratio Lower Upper
178 100
176 18.2 15.1 22.7
179 15.2 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.354 ng
RT: 19.217 min Scan# 1727
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:212 Resp: 62831
Ion Ratio Lower Upper
212 100
106 14.4 7.0 10.6#
104 7.8 4.1 6.1#



#28
Fluoranthene
Concen: 0.380 ng
RT: 19.247 min Scan# 1733
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Tgt Ion:202 Resp: 78301
Ion Ratio Lower Upper
202 100
101 11.9 6.6 10.0#
203 17.3 13.9 20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.376 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:240 Resp: 53284
Ion Ratio Lower Upper
240 100
120 10.8 4.2 6.4#
236 26.6 21.4 32.2



#30
Pyrene
Concen: 0.380 ng
RT: 19.609 min Scan# 1806
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Tgt Ion:202 Resp: 77762
Ion Ratio Lower Upper
202 100
200 20.5 16.9 25.3
203 17.4 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.379 ng
 RT: 19.818 min Scan# 1848
 Delta R.T. 0.000 min
 Lab File: BN015583.D
 Acq: 15 Jul 2021 17:47




Instrument :
 BNA_N
 ClientSampleId :
 PB137284BSD


Tgt Ion:244 Resp: 47877
 Ion Ratio Lower Upper
 244 100
 212 7.8 6.3 9.5
 122 13.5 6.2 9.2#



#32
 Benzo(a)anthracene
 Concen: 0.376 ng
 RT: 21.366 min Scan# 2014
 Delta R.T. 0.011 min
 Lab File: BN015583.D
 Acq: 15 Jul 2021 17:47

Tgt Ion:228 Resp: 69975
 Ion Ratio Lower Upper
 228 100
 226 26.1 23.0 34.6
 229 20.2 15.7 23.5





#33

Chrysene

Concen: 0.387 ng

RT: 21.408 min Scan# 2018

Delta R.T. 0.000 min

Lab File: BN015583.D

Acq: 15 Jul 2021 17:47

Tgt Ion:228 Resp: 76278

Ion Ratio Lower Upper

228 100

226 30.7 25.0 37.6





229 19.1 15.9 23.9

Instrument :

BNA_N

ClientSampleId :

PB137284BSD



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.349 ng

RT: 21.302 min Scan# 2008

Delta R.T. -0.000 min

Lab File: BN015583.D

Acq: 15 Jul 2021 17:47




Tgt Ion:149 Resp: 27815


Ion Ratio Lower Upper

149 100

167 28.1 28.9 43.3#


279 3.6 4.9 7.3#








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.724 min Scan# 2497
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Instrument :
BNA_N
ClientSampleId :
PB137284BSD



Tgt Ion:264 Resp: 51178




Ion	Ratio	Lower	Upper
264	100		
260	25.0	20.6	31.0
265	23.5	21.4	32.2




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.376 ng
RT: 26.141 min Scan# 3203
Delta R.T. 0.007 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Tgt Ion:276 Resp: 68783

Ion	Ratio	Lower	Upper
276	100		
138	29.5	10.0	15.0#
277	25.0	20.4	30.6









#37
Benzo(b)fluoranthene
Concen: 0.377 ng
RT: 23.012 min Scan# 2289
Delta R.T. 0.000 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47




Instrument :
BNA_N
ClientSampleId :
PB137284BSD


Tgt Ion:252 Resp: 75622
Ion Ratio Lower Upper
252 100
253 22.1 18.5 27.7
125 12.2 5.8 8.6#



#38
Benzo(k)fluoranthene
Concen: 0.388 ng
RT: 23.060 min Scan# 2303
Delta R.T. 0.003 min
Lab File: BN015583.D
Acq: 15 Jul 2021 17:47

Tgt Ion:252 Resp: 78548
Ion Ratio Lower Upper
252 100
253 22.1 19.1 28.7
125 11.2 5.4 8.0#





#39

Benzo(a)pyrene

Concen: 0.378 ng





RT: 23.622 min Scan# 2467

Delta R.T. 0.003 min

Lab File: BN015583.D

Acq: 15 Jul 2021 17:47

Tgt Ion:252	Resp:	65275
Ion Ratio	Lower	Upper
252	100	
253	22.6	19.4 29.2
125	16.2	7.4 11.2#



#40

Dibenzo(a,h)anthracene

Concen: 0.377 ng




RT: 26.158 min Scan# 3208


Delta R.T. 0.003 min

Lab File: BN015583.D

Acq: 15 Jul 2021 17:47

Tgt Ion:278	Resp:	54128
Ion Ratio	Lower	Upper
278	100	
139	19.4	9.0 13.4#
279	23.9	21.0 31.4





#41
 Benzo(g,h,i)perylene
 Concen: 0.370 ng
 RT: 26.877 min Scan# 3418
 Delta R.T. 0.003 min
 Lab File: BN015583.D
 Acq: 15 Jul 2021 17:47

Instrument :
 BNA_N
ClientSampleId :
 PB137284BSD

Tgt Ion:	276	Resp:	57349
Ion Ratio	Lower	Upper	
276	100		
277	23.8	19.7	29.5
138	23.8	9.8	14.6#

