

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071521\
 Data File : BN015586.D
 Acq On : 15 Jul 2021 19:39
 Operator : CG/JU
 Sample : M2799-04
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 MW-8

Quant Time: Jul 16 01:16:13 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN071521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 15 15:46:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.808	152	16636	0.400	ng	# 0.00
7) Naphthalene-d8	10.584	136	75663	0.400	ng	# 0.00
13) Acenaphthene-d10	14.434	164	40475	0.400	ng	0.00
19) Phenanthrene-d10	17.176	188	74295	0.400	ng	0.00
29) Chrysene-d12	21.376	240	67151	0.400	ng	# 0.00
35) Perylene-d12	23.724	264	65059	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.429	112	8140	0.171	ng	0.04
5) Phenol-d6	6.988	99	5570	0.098	ng	0.00
8) Nitrobenzene-d5	8.949	82	22635	0.416	ng	0.00
11) 2-Methylnaphthalene-d10	12.181	152	43158	0.347	ng	0.00
14) 2,4,6-Tribromophenol	15.918	330	6261	0.437	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	66677	0.408	ng	0.00
27) Fluoranthene-d10	19.217	212	83065	0.370	ng	0.00
31) Terphenyl-d14	19.817	244	76083	0.478	ng	0.00
Target Compounds						
9) Naphthalene	10.635	128	366017	1.666	ng	Qvalue 98
12) 2-Methylnaphthalene	12.252	142	30089	0.216	ng	# 90
16) Acenaphthylene	14.142	152	24027	0.128	ng	98
24) Pentachlorophenol	16.823	266	393	0.309	ng	# 47
34) Bis(2-ethylhexyl)phtha...	21.291	149	23831	0.237	ng	# 88

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071521\
Data File : BN015586.D
Acq On : 15 Jul 2021 19:39
Operator : CG/JU
Sample : M2799-04
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
MW-8

Quant Time: Jul 16 01:16:13 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN071521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 15 15:46:51 2021
Response via : Initial Calibration

