


2,4,6-Tribromophenol

Response Ratio




$$\text{Response} = 2.049\text{e-}001 * \text{Amt} - 6.911\text{e-}002$$

Coef of Det (r^2) = 0.996154 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071521.M

Calibration Table Last Updated: Thu Jul 15 15:46:51 2021


4,6-Dinitro-2-methylphenol



$R = 4.576e-003 A^2 - 3.685e-003 A + 1.193e-003$
 Coef of Det (r^2) = 0.999969 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071521.M
 Calibration Table Last Updated: Thu Jul 15 15:46:51 2021

Atrazine

Response Ratio



$$\text{Response} = 2.284\text{e-}001 * \text{Amt} - 5.454\text{e-}002$$


Coef of Det (r^2) = 0.999171 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071521.M

Calibration Table Last Updated: Thu Jul 15 15:46:51 2021

Pentachlorophenol

Response Ratio



$R = 4.028e-003 A^2 + 6.988e-002 A - 5.104e-002$
 Coef of Det (r^2) = 0.998283 Curve Fit: Quadratic
 Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071521.M
 Calibration Table Last Updated: Thu Jul 15 15:46:51 2021