

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071521\
 Data File : BN015580.D
 Acq On : 15 Jul 2021 15:54
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN071521

Quant Time: Jul 15 16:39:25 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN071521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 15 15:46:51 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	101	0.00
2	1,4-Dioxane	0.196	0.212	-8.2	110	0.00
3	n-Nitrosodimethylamine	0.330	0.325	1.5	105	-0.02
4 S	2-Fluorophenol	1.147	1.172	-2.2	103	0.00
5 S	Phenol-d6	1.364	1.399	-2.6	103	0.00
6	bis(2-Chloroethyl)ether	1.220	1.215	0.4	102	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	101	0.00
8 S	Nitrobenzene-d5	0.288	0.270	6.2	107	0.00
9	Naphthalene	1.161	1.108	4.6	100	0.00
10	Hexachlorobutadiene	0.217	0.210	3.2	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.658	0.621	5.6	102	0.00
12	2-Methylnaphthalene	0.736	0.717	2.6	101	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	102	0.00
14 S	2,4,6-Tribromophenol	0.145	0.131	9.7	111	0.00
15 S	2-Fluorobiphenyl	1.614	1.537	4.8	102	0.00
16	Acenaphthylene	1.861	1.722	7.5	104	0.00
17	Acenaphthene	1.243	1.190	4.3	103	0.01
18	Fluorene	1.492	1.430	4.2	103	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
20	4,6-Dinitro-2-methylphenol	0.016	0.006	62.5#	299#	-0.01
21	4-Bromophenyl-phenylether	0.254	0.243	4.3	105	0.00
22	Hexachlorobenzene	0.286	0.273	4.5	102	0.00
23	Atrazine	0.183	0.165	9.8	112	0.00
24	Pentachlorophenol	0.061	0.035	42.6#	143	0.00
25	Phenanthrene	1.293	1.243	3.9	102	0.01
26	Anthracene	1.110	1.040	6.3	106	0.00
27 SURR	Fluoranthene-d10	1.209	1.113	7.9	105	0.00
28	Fluoranthene	1.400	1.355	3.2	104	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
30	Pyrene	1.538	1.477	4.0	104	0.00
31 S	Terphenyl-d14	0.949	0.901	5.1	103	0.00
32	Benzo(a)anthracene	1.397	1.346	3.7	106	0.01
33	Chrysene	1.480	1.424	3.8	104	0.01
34	Bis(2-ethylhexyl)phthalate	0.599	0.696	-16.2	132	0.00
35 I	Perylene-d12	1.000	1.000	0.0	102	0.01
36	Indeno(1,2,3-cd)pyrene	1.428	1.312	8.1	105	0.01
37	Benzo(b)fluoranthene	1.569	1.503	4.2	103	0.00
38	Benzo(k)fluoranthene	1.584	1.534	3.2	100	0.01
39 C	Benzo(a)pyrene	1.351	1.281	5.2	104	0.01
40	Dibenzo(a,h)anthracene	1.122	1.040	7.3	107	0.01
41	Benzo(g,h,i)perylene	1.211	1.095	9.6	104	0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071521\
Data File : BN015580.D
Acq On : 15 Jul 2021 15:54
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN071521

Quant Time: Jul 15 16:39:25 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN071521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 15 15:46:51 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range				
SPCC's out = 0 CCC's out = 0				