

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071521\
 Data File : BN015573.D
 Acq On : 15 Jul 2021 10:23
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

mohammad
 7/16/2021 8:53:11 AM

Quant Time: Jul 15 12:20:32 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN071521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 15 12:19:46 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.808	152	13836	0.400	ng	# 0.00
7) Naphthalene-d8	10.584	136	60787	0.400	ng	0.00
13) Acenaphthene-d10	14.421	164	30402	0.400	ng	-0.01
19) Phenanthrene-d10	17.176	188	55008	0.400	ng	# 0.00
29) Chrysene-d12	21.376	240	44820	0.400	ng	0.00
35) Perylene-d12	23.714	264	42562	0.400	ng	-0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.392	112	4017	0.114	ng	0.00
5) Phenol-d6	6.969	99	4678	0.105	ng	0.00
8) Nitrobenzene-d5	8.949	82	3896	0.114	ng	0.00
11) 2-Methylnaphthalene-d10	12.176	152	10492	0.104	ng	0.00
14) 2,4,6-Tribromophenol	15.918	330	819	0.101	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	12891	0.104	ng	0.00
27) Fluoranthene-d10	19.212	212	16425	0.101	ng	0.00
31) Terphenyl-d14	19.812	244	11190	0.107	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.364	88	648m	0.100	ng	
3) n-Nitrosodimethylamine	3.714	42	1038	0.085	ng	# 79
6) bis(2-Chloroethyl)ether	7.237	93	4398	0.101	ng	# 85
9) Naphthalene	10.635	128	18909	0.108	ng	98
10) Hexachlorobutadiene	10.926	225	3384	0.103	ng	# 97
12) 2-Methylnaphthalene	12.252	142	11297	0.101	ng	# 90
16) Acenaphthylene	14.142	152	12506	0.093	ng	98
17) Acenaphthene	14.484	154	9521	0.101	ng	99
18) Fluorene	15.474	166	10924	0.095	ng	99
20) 4,6-Dinitro-2-methylph...	15.576	198	24m	0.023	ng	
21) 4-Bromophenyl-phenylether	16.372	248	3371	0.097	ng	94
22) Hexachlorobenzene	16.482	284	4099	0.103	ng	# 92
23) Atrazine	16.640	200	1947	0.096	ng	97
25) Phenanthrene	17.212	178	17948	0.101	ng	98
26) Anthracene	17.310	178	13521	0.089	ng	99
28) Fluoranthene	19.242	202	18055	0.095	ng	# 97
30) Pyrene	19.604	202	18114	0.106	ng	99
32) Benzo(a)anthracene	21.355	228	14184	0.092	ng	98
33) Chrysene	21.408	228	17686	0.106	ng	98
34) Bis(2-ethylhexyl)phtha...	21.291	149	5653	0.112	ng	# 88
36) Indeno(1,2,3-cd)pyrene	26.120	276	12375	0.081	ng	# 82
37) Benzo(b)fluoranthene	23.005	252	16314	0.098	ng	# 92
38) Benzo(k)fluoranthene	23.050	252	17099	0.096	ng	# 96
39) Benzo(a)pyrene	23.614	252	13470	0.088	ng	# 81
40) Dibenzo(a,h)anthracene	26.144	278	9365	0.078	ng	# 92
41) Benzo(g,h,i)perylene	26.859	276	10368	0.079	ng	# 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071521\
Data File : BN015573.D
Acq On : 15 Jul 2021 10:23
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Manual Integrations
APPROVED

mohammad
7/16/2021 8:53:11 AM


Quant Time: Jul 15 12:20:32 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN071521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 15 12:19:46 2021
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.808 min Scan# 513
Delta R.T. 0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:152 Resp: 13836
Ion Ratio Lower Upper
152 100
150 155.0 119.0 178.6
115 59.4 62.2 93.4#

Manual Integrations
APPROVED



mohammad
7/16/2021 8:53:11 AM




#2
1,4-Dioxane
Concen: 0.100 ng m
RT: 3.364 min Scan# 31
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23



Tgt Ion: 88 Resp: 648
Ion Ratio Lower Upper
88 100
43 523.9 12.9 19.3#
58 201.5 38.1 57.1#









#3
n-Nitrosodimethylamine
Concen: 0.085 ng
RT: 3.714 min Scan# 69
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23

Tgt Ion: 42 Resp: 1038
Ion Ratio Lower Upper
42 100
74 155.7 149.5 224.3
44 11.6 4.2 6.2#

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
7/16/2021 8:53:11 AM



#4
2-Fluorophenol
Concen: 0.114 ng
RT: 5.392 min Scan# 251
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23

Tgt Ion: 112 Resp: 4017
Ion Ratio Lower Upper
112 100
64 55.2 39.0 58.6
63 27.7 20.6 30.8









#5
Phenol-d6
Concen: 0.105 ng
RT: 6.969 min Scan# 422
Delta R.T. -0.009 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23

Tgt Ion: 99 Resp: 4678
Ion Ratio Lower Upper
99 100
42 16.8 8.9 13.3#
71 30.4 35.6 53.4#

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Manual Integrations
APPROVED


mohammad
7/16/2021 8:53:11 AM



#6
bis(2-Chloroethyl)ether
Concen: 0.101 ng
RT: 7.237 min Scan# 451
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23


Tgt Ion: 93 Resp: 4398
Ion Ratio Lower Upper
93 100
63 74.5 46.6 70.0#
95 32.7 28.6 42.8





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 747
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	136	Resp:	60787
Ion	Ratio	Lower	Upper
136	100		
137	10.9	9.8	14.8
54	7.2	5.4	8.2
68	6.1	4.9	7.3

Manual Integrations
APPROVED



mohammad
7/16/2021 8:53:11 AM




#8
Nitrobenzene-d5
Concen: 0.114 ng
RT: 8.949 min Scan# 618
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23



Tgt Ion:	82	Resp:	3896
Ion	Ratio	Lower	Upper
82	100		
128	39.5	31.6	47.4
54	50.9	29.0	43.4









#9
Naphthalene
Concen: 0.108 ng
RT: 10.635 min Scan# 751
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23

Tgt Ion:	128	Resp:	18909
Ion Ratio	Lower	Upper	
128	100		
129	11.3	9.1	13.7
127	13.2	11.8	17.6

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
7/16/2021 8:53:11 AM



#10
Hexachlorobutadiene
Concen: 0.103 ng
RT: 10.926 min Scan# 774
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23

Tgt Ion:	225	Resp:	3384
Ion Ratio	Lower	Upper	
225	100		
223	0.0	0.0	0.0
227	65.0	50.2	75.2









#11
2-Methylnaphthalene-d10
Concen: 0.104 ng
RT: 12.176 min Scan# 970
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23

Tgt Ion:152 Resp: 10492
Ion Ratio Lower Upper
152 100
151 20.4 16.7 25.1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Manual Integrations
APPROVED


mohammad
7/16/2021 8:53:11 AM



#12
2-Methylnaphthalene
Concen: 0.101 ng
RT: 12.252 min Scan# 987
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23


Tgt Ion:142 Resp: 11297
Ion Ratio Lower Upper
142 100
141 88.6 72.3 108.5
115 33.2 40.7 61.1#





#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.421 min Scan# 1195
Delta R.T. -0.013 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:164 Resp: 30402
Ion Ratio Lower Upper
164 100
162 104.1 83.8 125.6
160 48.6 37.4 56.0

Manual Integrations
APPROVED



mohammad
7/16/2021 8:53:11 AM




#14
2,4,6-Tribromophenol
Concen: 0.101 ng
RT: 15.918 min Scan# 1313
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23




Tgt Ion:330 Resp: 819
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 38.8 22.4 33.6#









#15
2-Fluorobiphenyl
Concen: 0.104 ng
RT: 13.051 min Scan# 1087
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23





Tgt Ion:	172	Resp:	12891
Ion Ratio	Lower	Upper	
172	100		
171	35.2	30.6	45.8
170	23.6	20.2	30.4



#16
Acenaphthylene
Concen: 0.093 ng
RT: 14.142 min Scan# 1173
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23




Tgt Ion:	152	Resp:	12506
Ion Ratio	Lower	Upper	
152	100		
151	19.4	16.7	25.1
153	13.3	10.5	15.7



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Manual Integrations
APPROVED

mohammad
7/16/2021 8:53:11 AM



#17
Acenaphthene
Concen: 0.101 ng
RT: 14.484 min Scan# 1200
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:154 Resp: 9521
Ion Ratio Lower Upper
154 100
153 113.5 90.9 136.3
152 56.1 45.8 68.8

Manual Integrations
APPROVED



mohammad
7/16/2021 8:53:11 AM




#18
Fluorene
Concen: 0.095 ng
RT: 15.474 min Scan# 1278
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23




Tgt Ion:166 Resp: 10924
Ion Ratio Lower Upper
166 100
165 97.9 79.2 118.8
167 13.5 10.7 16.1





#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.176 min Scan# 1416
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	8.2	9.0	13.4#

Manual Integrations
APPROVED



mohammad
7/16/2021 8:53:11 AM




#20
4,6-Dinitro-2-methylphenol
Concen: 0.023 ng m
RT: 15.576 min Scan# 1286
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23




Tgt Ion	Ratio	Lower	Upper
198	100		
182	96.2	34.5	51.7#
77	0.0	0.0	0.0





#21
4-Bromophenyl-phenylether
Concen: 0.097 ng
RT: 16.372 min Scan# 1350
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	248	Resp:	3371
Ion Ratio	Lower	Upper	
248	100		
250	100.2	86.2	129.4
141	43.2	36.2	54.2

Manual Integrations
APPROVED



mohammad
7/16/2021 8:53:11 AM




#22
Hexachlorobenzene
Concen: 0.103 ng
RT: 16.482 min Scan# 1359
Delta R.T. -0.012 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23




Tgt Ion:	284	Resp:	4099
Ion Ratio	Lower	Upper	
284	100		
142	34.6	22.2	33.2#
249	31.2	26.6	39.8









#23
Atrazine
Concen: 0.096 ng
RT: 16.640 min Scan# 1372
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23





Tgt Ion: 200 Resp: 1947
Ion Ratio Lower Upper
200 100
173 27.5 21.7 32.5
215 50.9 43.1 64.7



#25
Phenanthrene
Concen: 0.101 ng
RT: 17.212 min Scan# 1419
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23




Tgt Ion: 178 Resp: 17948
Ion Ratio Lower Upper
178 100
176 18.8 15.6 23.4
179 16.6 12.1 18.1




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Manual Integrations
APPROVED




mohammad
7/16/2021 8:53:11 AM




#26
 Anthracene
 Concen: 0.089 ng
 RT: 17.310 min Scan# 1427
 Delta R.T. -0.000 min
 Lab File: BN015573.D
 Acq: 15 Jul 2021 10:23





Tgt Ion:	178	Resp:	13521
Ion Ratio	Lower	Upper	
178	100		
176	18.1	15.1	22.7
179	15.3	12.2	18.2



#27
 Fluoranthene-d10
 Concen: 0.101 ng
 RT: 19.212 min Scan# 1726
 Delta R.T. -0.005 min
 Lab File: BN015573.D
 Acq: 15 Jul 2021 10:23




Tgt Ion:	212	Resp:	16425
Ion Ratio	Lower	Upper	
212	100		
106	14.3	7.0	10.6#
104	7.8	4.1	6.1#



Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Manual Integrations
 APPROVED

mohammad
 7/16/2021 8:53:11 AM



#28
Fluoranthene
Concen: 0.095 ng
RT: 19.242 min Scan# 1732
Delta R.T. -0.005 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23



Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:202 Resp: 18055
Ion Ratio Lower Upper
202 100
101 11.7 6.6 10.0#
203 17.3 13.9 20.9

Manual Integrations
APPROVED



mohammad
7/16/2021 8:53:11 AM




#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.376 min Scan# 2015
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23



Tgt Ion:240 Resp: 44820
Ion Ratio Lower Upper
240 100
120 5.8 4.2 6.4
236 25.4 21.4 32.2









#30
Pyrene
Concen: 0.106 ng
RT: 19.604 min Scan# 1805
Delta R.T. -0.005 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23

Tgt Ion:	202	Resp:	18114
Ion Ratio	Lower	Upper	
202	100		
200	20.6	16.9	25.3
203	17.3	13.9	20.9

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
7/16/2021 8:53:11 AM



#31
Terphenyl-d14
Concen: 0.107 ng
RT: 19.812 min Scan# 1847
Delta R.T. -0.005 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23

Tgt Ion:	244	Resp:	11190
Ion Ratio	Lower	Upper	
244	100		
212	8.1	6.3	9.5
122	13.5	6.2	9.2#









#32
Benzo(a)anthracene
Concen: 0.092 ng
RT: 21.355 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23

Tgt Ion	Ratio	Lower	Upper
228	100		
226	26.9	23.0	34.6
229	19.8	15.7	23.5

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Manual Integrations
APPROVED


mohammad
7/16/2021 8:53:11 AM



#33
Chrysene
Concen: 0.106 ng
RT: 21.408 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23

Tgt Ion	Ratio	Lower	Upper
228	100		
226	29.6	25.0	37.6
229	19.8	15.9	23.9






#34
 Bis(2-ethylhexyl)phthalate
 Concen: 0.112 ng
 RT: 21.291 min Scan# 2007
 Delta R.T. -0.011 min
 Lab File: BN015573.D
 Acq: 15 Jul 2021 10:23




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

mohammad
 7/16/2021 8:53:11 AM







Tgt Ion	Ratio	Lower	Upper
149	100		
167	28.9	28.9	43.3#
279	3.8	4.9	7.3#



#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.714 min Scan# 2494
 Delta R.T. -0.010 min
 Lab File: BN015573.D
 Acq: 15 Jul 2021 10:23

Tgt Ion	Ratio	Lower	Upper
264	100		
260	25.1	20.6	31.0
265	24.3	21.4	32.2









#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.081 ng
 RT: 26.120 min Scan# 3197
 Delta R.T. -0.014 min
 Lab File: BN015573.D
 Acq: 15 Jul 2021 10:23

Tgt Ion: 276 Resp: 12375
 Ion Ratio Lower Upper
 276 100
 138 32.6 10.0 15.0#
 277 24.9 20.4 30.6

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1




Manual Integrations
 APPROVED


mohammad
 7/16/2021 8:53:11 AM



#37
 Benzo(b)fluoranthene
 Concen: 0.098 ng
 RT: 23.005 min Scan# 2287
 Delta R.T. -0.007 min
 Lab File: BN015573.D
 Acq: 15 Jul 2021 10:23

Tgt Ion: 252 Resp: 16314
 Ion Ratio Lower Upper
 252 100
 253 24.1 18.5 27.7
 125 16.1 5.8 8.6#









#38
Benzo(k)fluoranthene
Concen: 0.096 ng
RT: 23.050 min Scan# 2300
Delta R.T. -0.007 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23

Tgt Ion:252 Resp: 17099
Ion Ratio Lower Upper
252 100
253 23.4 19.1 28.7
125 12.4 5.4 8.0#

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
7/16/2021 8:53:11 AM




#39
Benzo(a)pyrene
Concen: 0.088 ng
RT: 23.614 min Scan# 2465
Delta R.T. -0.004 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23

Tgt Ion:252 Resp: 13470
Ion Ratio Lower Upper
252 100
253 26.7 19.4 29.2
125 29.9 7.4 11.2#









#40
Dibenzo(a,h)anthracene
Concen: 0.078 ng
RT: 26.144 min Scan# 3204
Delta R.T. -0.010 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23





Tgt Ion:278 Resp: 9365
Ion Ratio Lower Upper
278 100
139 21.3 9.0 13.4#
279 26.3 21.0 31.4



#41
Benzo(g,h,i)perylene
Concen: 0.079 ng
RT: 26.859 min Scan# 3413
Delta R.T. -0.014 min
Lab File: BN015573.D
Acq: 15 Jul 2021 10:23



Tgt Ion:276 Resp: 10368
Ion Ratio Lower Upper
276 100
277 24.8 19.7 29.5
138 25.0 9.8 14.6#



Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Manual Integrations
APPROVED

mohammad
7/16/2021 8:53:11 AM