

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071521\
 Data File : BN015580.D
 Acq On : 15 Jul 2021 15:54
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN071521

Quant Time: Jul 15 16:39:25 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN071521.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Jul 15 15:46:51 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.799	152	13736	0.400	ng	# 0.00
7) Naphthalene-d8	10.584	136	60799	0.400	ng	0.00
13) Acenaphthene-d10	14.434	164	31425	0.400	ng	0.00
19) Phenanthrene-d10	17.176	188	56715	0.400	ng	0.00
29) Chrysene-d12	21.376	240	51912	0.400	ng	# 0.00
35) Perylene-d12	23.734	264	46876	0.400	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.392	112	16103	0.409	ng	0.00
5) Phenol-d6	6.969	99	19210	0.410	ng	0.00
8) Nitrobenzene-d5	8.949	82	16408	0.375	ng	0.00
11) 2-Methylnaphthalene-d10	12.176	152	37728	0.377	ng	0.00
14) 2,4,6-Tribromophenol	15.918	330	4128	0.391	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	48285	0.381	ng	0.00
27) Fluoranthene-d10	19.217	212	63151	0.368	ng	0.00
31) Terphenyl-d14	19.822	244	46793	0.380	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.355	88	2917	0.434	ng	# 33
3) n-Nitrosodimethylamine	3.696	42	4462	0.394	ng	79
6) bis(2-Chloroethyl)ether	7.227	93	16687	0.398	ng	# 85
9) Naphthalene	10.635	128	67343	0.382	ng	97
10) Hexachlorobutadiene	10.926	225	12798	0.388	ng	# 98
12) 2-Methylnaphthalene	12.252	142	43580	0.389	ng	# 89
16) Acenaphthylene	14.142	152	54115	0.370	ng	98
17) Acenaphthene	14.497	154	37389	0.383	ng	98
18) Fluorene	15.474	166	44942	0.383	ng	99
20) 4,6-Dinitro-2-methylph...	15.563	198	362	0.616	ng	# 82
21) 4-Bromophenyl-phenylether	16.372	248	13792	0.383	ng	# 87
22) Hexachlorobenzene	16.494	284	15460	0.382	ng	# 92
23) Atrazine	16.640	200	9383	0.385	ng	92
24) Pentachlorophenol	16.835	266	1960	0.460	ng	99
25) Phenanthrene	17.224	178	70478	0.385	ng	99
26) Anthracene	17.310	178	58988	0.375	ng	99
28) Fluoranthene	19.247	202	76858	0.387	ng	# 97
30) Pyrene	19.609	202	76679	0.384	ng	99
32) Benzo(a)anthracene	21.365	228	69882	0.385	ng	98
33) Chrysene	21.419	228	73912	0.385	ng	97
34) Bis(2-ethylhexyl)phtha...	21.302	149	36148	0.465	ng	# 87
36) Indeno(1,2,3-cd)pyrene	26.144	276	61510	0.367	ng	# 86
37) Benzo(b)fluoranthene	23.019	252	70432	0.383	ng	# 96
38) Benzo(k)fluoranthene	23.067	252	71906	0.387	ng	# 94
39) Benzo(a)pyrene	23.628	252	60036	0.379	ng	# 93
40) Dibenzo(a,h)anthracene	26.168	278	48733	0.371	ng	# 91
41) Benzo(g,h,i)perylene	26.883	276	51347	0.362	ng	# 89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071521\
Data File : BN015580.D
Acq On : 15 Jul 2021 15:54
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN071521

Quant Time: Jul 15 16:39:25 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN071521.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 15 15:46:51 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.799 min Scan# 512
 Delta R.T. -0.009 min
 Lab File: BN015580.D
 Acq: 15 Jul 2021 15:54




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN071521


Tgt Ion: 152 Resp: 13736
 Ion Ratio Lower Upper
 152 100
 150 156.4 119.0 178.6
 115 61.1 62.2 93.4#



#2
 1,4-Dioxane
 Concen: 0.434 ng
 RT: 3.355 min Scan# 30
 Delta R.T. -0.009 min
 Lab File: BN015580.D
 Acq: 15 Jul 2021 15:54

Tgt Ion: 88 Resp: 2917
 Ion Ratio Lower Upper
 88 100
 43 73.6 12.9 19.3#
 58 76.7 38.1 57.1#









#3
n-Nitrosodimethylamine
Concen: 0.394 ng
RT: 3.696 min Scan# 67
Delta R.T. -0.019 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54




Instrument :
BNA_N
ClientSampleId :
ICVBN071521


Tgt Ion: 42 Resp: 4462
Ion Ratio Lower Upper
42 100
74 155.0 149.5 224.3
44 5.9 4.2 6.2



#4
2-Fluorophenol
Concen: 0.409 ng
RT: 5.392 min Scan# 251
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54

Tgt Ion: 112 Resp: 16103
Ion Ratio Lower Upper
112 100
64 55.8 39.0 58.6
63 27.7 20.6 30.8









#5
Phenol-d6
Concen: 0.410 ng
RT: 6.969 min Scan# 422
Delta R.T. -0.009 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54




Instrument :
BNA_N
ClientSampled :
ICVBN071521


Tgt Ion: 99 Resp: 19210
Ion Ratio Lower Upper
99 100
42 16.8 8.9 13.3#
71 30.2 35.6 53.4#



#6
bis(2-Chloroethyl)ether
Concen: 0.398 ng
RT: 7.227 min Scan# 450
Delta R.T. -0.009 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54

Tgt Ion: 93 Resp: 16687
Ion Ratio Lower Upper
93 100
63 74.4 46.6 70.0#
95 33.3 28.6 42.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.584 min Scan# 747
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54




Instrument :
BNA_N
ClientSampleId :
ICVBN071521


Tgt Ion:	136	Resp:	60799
Ion Ratio	Lower	Upper	
136	100		
137	11.0	9.8	14.8
54	7.3	5.4	8.2
68	6.2	4.9	7.3



#8
Nitrobenzene-d5
Concen: 0.375 ng
RT: 8.949 min Scan# 618
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54

Tgt Ion:	82	Resp:	16408
Ion Ratio	Lower	Upper	
82	100		
128	42.1	31.6	47.4
54	47.0	29.0	43.4









#9
Naphthalene
Concen: 0.382 ng
RT: 10.635 min Scan# 751
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54




Instrument :
BNA_N
ClientSampleId :
ICVBN071521


Tgt Ion:	128	Resp:	67343
Ion	Ratio	Lower	Upper
128	100		
129	11.0	9.1	13.7
127	13.0	11.8	17.6



#10
Hexachlorobutadiene
Concen: 0.388 ng
RT: 10.926 min Scan# 774
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54

Tgt Ion:	225	Resp:	12798
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.1	50.2	75.2









#11
2-Methylnaphthalene-d10
Concen: 0.377 ng
RT: 12.176 min Scan# 970
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54




Instrument :
BNA_N
ClientSampleId :
ICVBN071521


Tgt Ion:152 Resp: 37728
Ion Ratio Lower Upper
152 100
151 20.4 16.7 25.1



#12
2-Methylnaphthalene
Concen: 0.389 ng
RT: 12.252 min Scan# 987
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54

Tgt Ion:142 Resp: 43580
Ion Ratio Lower Upper
142 100
141 87.9 72.3 108.5
115 32.6 40.7 61.1#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.434 min Scan# 1196
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54




Instrument :
BNA_N
ClientSampled :
ICVBN071521


Tgt Ion:	164	Resp:	31425
Ion Ratio	Lower	Upper	
164	100		
162	99.6	83.8	125.6
160	44.0	37.4	56.0



#14
2,4,6-Tribromophenol
Concen: 0.391 ng
RT: 15.918 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54

Tgt Ion:	330	Resp:	4128
Ion Ratio	Lower	Upper	
330	100		
332	0.0	0.0	0.0
141	36.1	22.4	33.6#









#15
2-Fluorobiphenyl
Concen: 0.381 ng
RT: 13.051 min Scan# 1087
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54




Instrument :
BNA_N
ClientSampleId :
ICVBN071521


Tgt Ion:172	Resp:	48285
Ion Ratio	Lower	Upper
172	100	
171	35.5	30.6 45.8
170	23.9	20.2 30.4



#16
Acenaphthylene
Concen: 0.370 ng
RT: 14.142 min Scan# 1173
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54

Tgt Ion:152	Resp:	54115
Ion Ratio	Lower	Upper
152	100	
151	19.4	16.7 25.1
153	13.0	10.5 15.7









#17
Acenaphthene
Concen: 0.383 ng
RT: 14.497 min Scan# 1201
Delta R.T. 0.012 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54




Instrument :
BNA_N
ClientSampleId :
ICVBN071521


Tgt Ion:154 Resp: 37389
Ion Ratio Lower Upper
154 100
153 112.0 90.9 136.3
152 54.4 45.8 68.8



#18
Fluorene
Concen: 0.383 ng
RT: 15.474 min Scan# 1278
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54

Tgt Ion:166 Resp: 44942
Ion Ratio Lower Upper
166 100
165 97.7 79.2 118.8
167 13.5 10.7 16.1









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.176 min Scan# 1416
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54




Instrument :
BNA_N
ClientSampleId :
ICVBN071521


Tgt Ion:188	Resp:	56715
Ion Ratio	Lower	Upper
188	100	
94	0.0	0.0
80	10.8	9.0
		13.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.616 ng
RT: 15.563 min Scan# 1285
Delta R.T. -0.013 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54

Tgt Ion:198	Resp:	362
Ion Ratio	Lower	Upper
198	100	
182	31.4	34.5
77	0.0	0.0
		51.7#









#21
4-Bromophenyl-phenylether
Concen: 0.383 ng
RT: 16.372 min Scan# 1350
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54




Instrument :
BNA_N
ClientSampleId :
ICVBN071521


Tgt Ion	Ratio	Lower	Upper
248	100		
250	95.8	86.2	129.4
141	57.0	36.2	54.2#



#22
Hexachlorobenzene
Concen: 0.382 ng
RT: 16.494 min Scan# 1360
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54

Tgt Ion	Ratio	Lower	Upper
284	100		
142	34.5	22.2	33.2#
249	31.0	26.6	39.8









#23
Atrazine
Concen: 0.385 ng
RT: 16.640 min Scan# 1372
Delta R.T. 0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54




Instrument :
BNA_N
ClientSampleId :
ICVBN071521


Tgt Ion:	200	Resp:	9383
Ion Ratio	Lower	Upper	
200	100		
173	28.6	21.7	32.5
215	46.6	43.1	64.7



#24
Pentachlorophenol
Concen: 0.460 ng
RT: 16.835 min Scan# 1388
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54

Tgt Ion:	266	Resp:	1960
Ion Ratio	Lower	Upper	
266	100		
264	62.8	51.0	76.4
268	62.5	50.6	76.0









#25
Phenanthrene
Concen: 0.385 ng
RT: 17.224 min Scan# 1420
Delta R.T. 0.012 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54




Instrument :
BNA_N
ClientSampleId :
ICVBN071521


Tgt Ion:178 Resp: 70478
Ion Ratio Lower Upper
178 100
176 18.8 15.6 23.4
179 15.4 12.1 18.1



#26
Anthracene
Concen: 0.375 ng
RT: 17.310 min Scan# 1427
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54

Tgt Ion:178 Resp: 58988
Ion Ratio Lower Upper
178 100
176 18.1 15.1 22.7
179 15.2 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.368 ng
RT: 19.217 min Scan# 1727
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54




Instrument :
BNA_N
ClientSampleId :
ICVBN071521


Tgt Ion:212 Resp: 63151
Ion Ratio Lower Upper
212 100
106 14.3 7.0 10.6#
104 7.8 4.1 6.1#



#28
Fluoranthene
Concen: 0.387 ng
RT: 19.247 min Scan# 1733
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54

Tgt Ion:202 Resp: 76858
Ion Ratio Lower Upper
202 100
101 11.7 6.6 10.0#
203 17.2 13.9 20.9









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.376 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54




Instrument :
BNA_N
ClientSampleId :
ICVBN071521


Tgt Ion:240 Resp: 51912
Ion Ratio Lower Upper
240 100
120 16.0 4.2 6.4#
236 27.1 21.4 32.2



#30
Pyrene
Concen: 0.384 ng
RT: 19.609 min Scan# 1806
Delta R.T. -0.000 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54

Tgt Ion:202 Resp: 76679
Ion Ratio Lower Upper
202 100
200 20.7 16.9 25.3
203 17.5 13.9 20.9









#31
 Terphenyl-d14
 Concen: 0.380 ng
 RT: 19.822 min Scan# 1849
 Delta R.T. 0.005 min
 Lab File: BN015580.D
 Acq: 15 Jul 2021 15:54




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN071521


Tgt Ion:244	Resp:	46793
Ion Ratio	Lower	Upper
244	100	
212	7.5	6.3 9.5
122	11.9	6.2 9.2#



#32
 Benzo(a)anthracene
 Concen: 0.385 ng
 RT: 21.365 min Scan# 2014
 Delta R.T. 0.010 min
 Lab File: BN015580.D
 Acq: 15 Jul 2021 15:54

Tgt Ion:228	Resp:	69882
Ion Ratio	Lower	Upper
228	100	
226	26.8	23.0 34.6
229	19.7	15.7 23.5





#33

Chrysene

Concen: 0.385 ng

RT: 21.419 min Scan# 2019

Delta R.T. 0.010 min

Lab File: BN015580.D

Acq: 15 Jul 2021 15:54

Tgt Ion:228 Resp: 73912

Ion Ratio Lower Upper

228 100

226 29.2 25.0 37.6





229 19.5 15.9 23.9

Instrument :

BNA_N

ClientSampleId :

ICVBN071521



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.465 ng

RT: 21.302 min Scan# 2008

Delta R.T. -0.000 min

Lab File: BN015580.D

Acq: 15 Jul 2021 15:54




Tgt Ion:149 Resp: 36148


Ion Ratio Lower Upper

149 100

167 28.1 28.9 43.3#


279 3.8 4.9 7.3#








#35
Perylene-d12
Concen: 0.400 ng
RT: 23.734 min Scan# 2500
Delta R.T. 0.010 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54


Instrument :
BNA_N
ClientSampleId :
ICVBN071521





Tgt Ion:264 Resp: 46876
Ion Ratio Lower Upper
264 100
260 24.7 20.6 31.0
265 23.7 21.4 32.2




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.367 ng
RT: 26.144 min Scan# 3204
Delta R.T. 0.010 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54



Tgt Ion:276 Resp: 61510
Ion Ratio Lower Upper
276 100
138 28.3 10.0 15.0#
277 25.0 20.4 30.6









#37
Benzo(b)fluoranthene
Concen: 0.383 ng
RT: 23.019 min Scan# 2291
Delta R.T. 0.007 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54




Instrument :
BNA_N
ClientSampleId :
ICVBN071521


Tgt Ion:252 Resp: 70432
Ion Ratio Lower Upper
252 100
253 22.4 18.5 27.7
125 11.7 5.8 8.6#



#38
Benzo(k)fluoranthene
Concen: 0.387 ng
RT: 23.067 min Scan# 2305
Delta R.T. 0.010 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54

Tgt Ion:252 Resp: 71906
Ion Ratio Lower Upper
252 100
253 22.1 19.1 28.7
125 10.9 5.4 8.0#









#39
Benzo(a)pyrene
Concen: 0.379 ng
RT: 23.628 min Scan# 2469
Delta R.T. 0.010 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54




Instrument :
BNA_N
ClientSampleId :
ICVBN071521


Tgt Ion:252 Resp: 60036
Ion Ratio Lower Upper
252 100
253 22.3 19.4 29.2
125 14.3 7.4 11.2#



#40
Dibenzo(a,h)anthracene
Concen: 0.371 ng
RT: 26.168 min Scan# 3211
Delta R.T. 0.014 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54


Tgt Ion:278 Resp: 48733
Ion Ratio Lower Upper
278 100
139 18.6 9.0 13.4#
279 23.8 21.0 31.4





#41
Benzo(g,h,i)perylene
Concen: 0.362 ng
RT: 26.883 min Scan# 3420
Delta R.T. 0.010 min
Lab File: BN015580.D
Acq: 15 Jul 2021 15:54

Instrument :
BNA_N
ClientSampleId :
ICVBN071521



Tgt Ion	Ratio	Lower	Upper
276	100		
277	23.7	19.7	29.5
138	23.6	9.8	14.6

