

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071619\
 Data File : BN006855.D
 Acq On : 16 Jul 2019 08:21
 Operator : HP/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD0.429

Quant Time: Jul 16 08:58:16 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN062719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 12 06:39:20 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	90	0.01
2 I	Naphthalene-d8	0.400	0.400	0.0	83	0.00
3	Naphthalene	0.400	0.390	2.5	77	0.00
4 SURR	2-Methylnaphthalene-d10	0.400	0.369	7.8	72	0.00
5	2-Methylnaphthalene	0.400	0.357	10.8	70	0.01
6 I	Acenaphthene-d10	0.400	0.400	0.0	77	0.00
7	Acenaphthylene	0.400	0.357	10.8	65	0.00
8 C	Acenaphthene	0.400	0.383	4.3	69	0.00
9	Fluorene	0.400	0.363	9.3	66	0.00
10 I	Phenanthrene-d10	0.400	0.400	0.0	74	0.00
11	Pentachlorophenol	0.400	0.225	43.8#	49	0.00
12	Phenanthrene	0.400	0.371	7.3	65	0.00
13	Anthracene	0.400	0.333	16.8	58	0.00
14 SURR	Fluoranthene-d10	0.400	0.358	10.5	63	0.00
15 C	Fluoranthene	0.400	0.346	13.5	62	0.00
16 I	Chrysene-d12	0.400	0.400	0.0	64	0.00
17	Pyrene	0.400	0.411	-2.7	62	0.00
18	Benzo(a)anthracene	0.400	0.331	17.3	51	0.00
19	Chrysene	0.400	0.415	-3.7	64	0.00
20 I	Perylene-d12	0.400	0.400	0.0	67	0.00
21	Benzo(b)fluoranthene	0.400	0.372	7.0	58	0.00
22	Benzo(k)fluoranthene	0.400	0.399	0.3	64	0.00
23 C	Benzo(a)pyrene	0.400	0.371	7.3	60	0.00
24	Indeno(1,2,3-cd)pyrene	0.400	0.435	-8.7	70	0.00
25	Dibenzo(a,h)anthracene	0.400	0.419	-4.7	68	0.00
26	Benzo(g,h,i)perylene	0.400	0.471	-17.7	76	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0