




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071621\
Data File : BN015589.D
Acq On : 16 Jul 2021 09:29
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 16 11:40:12 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 15 05:41:44 2021
Response via : Initial Calibration



TIC: BN015589.D\data.ms

(77) Benzidine




19.428min (-0.018) 0.00 ng

response 1040955

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.84
183.00	13.20	11.73
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071621\
Data File : BN015589.D
Acq On : 16 Jul 2021 09:29
Operator : CG/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Jul 16 11:40:12 2021
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Jul 15 05:41:44 2021
Response via : Initial Calibration



TIC: BN015589.D\data.ms

(70) Pentachlorophenol (C)

16.828min (-0.012) 26241.00 ng

response 168453

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	66.24
264.00	61.60	67.98
0.00	0.00	0.00