

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071723\
 Data File : BN026538.D
 Acq On : 17 Jul 2023 15:53
 Operator : MA/JU
 Sample : 03457-13
 Misc :
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0C60

Quant Time: Jul 18 03:04:54 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN071223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 17 17:14:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.946	152	2165	0.400	ng/ul	0.00
4) Naphthalene-d8	10.748	136	5872	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.579	164	3823	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.326	188	8793	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.529	240	7360	0.400	ng/ul	# 0.00
23) Perylene-d12	23.935	264	8332	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.297	96	11544	4.049	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.359	152	2284	0.286	ng/ul	0.01
18) Fluoranthene-d10	19.359	212	9565	0.375	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071723\
Data File : BN026538.D
Acq On : 17 Jul 2023 15:53
Operator : MA/JU
Sample : 03457-13
Misc :
ALS Vial : 14 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0C60

Quant Time: Jul 18 03:04:54 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN071223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 17 17:14:21 2023
Response via : Initial Calibration

