


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN071819\
 Data File : BN006884.D
 Acq On : 18 Jul 2019 10:12
 Operator : HP/JU
 Sample : DFTPP07
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP07

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN062719.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Jul 19 03:57:16 2019



AutoFind: Scans 2456, 2457, 2458; Background Corrected with Scan 2450

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	56.3	109845	PASS
68	69	0.00	2	1.8	1641	PASS
69	198	0.00	100	47.3	92234	PASS
70	69	0.00	2	0.7	608	PASS
127	198	10	80	49.4	96328	PASS
197	198	0.00	2	0.7	1307	PASS
198	198	100	100	100.0	195008	PASS
199	198	5	9	7.0	13728	PASS
275	198	10	60	24.3	47400	PASS
365	198	1	100	2.9	5736	PASS
441	198	0.01	100	9.9	19350	PASS
442	198	50	100	62.2	121392	PASS
443	442	15	24	19.7	23910	PASS

Scan 2450 (17.398 min): BN006884.D
DFTPP07

m/z	Abundance
39.90	1244.0
41.00	417.0
43.00	372.0
44.00	4678.0
55.00	366.0
73.00	391.0
207.00	1223.0

Instrument :
BNA_N
ClientSampleId :
DFTPP07

m/z	Abundance
36.90	518.0
38.00	1170.0
39.00	8022.0
40.00	1439.0
41.00	783.0
43.00	605.0
44.00	4329.0
44.90	327.0
49.00	959.0
50.00	24648.0
51.00	94696.0
52.00	4654.0
53.00	339.0
55.00	867.0
56.00	2819.0
57.00	5707.0
57.90	370.0
61.00	1090.0
62.00	1152.0
63.00	3610.0
64.00	407.0
65.00	1634.0
68.00	1509.0
69.00	77624.0
70.00	586.0
73.00	1114.0
74.00	8290.0
75.00	13529.0
76.00	4442.0
77.00	93576.0
78.00	6745.0
79.00	5557.0
80.00	4399.0
81.00	6440.0
81.90	1645.0
83.00	1590.0
83.90	447.0
85.00	1011.0
86.00	1834.0
87.00	824.0
88.00	331.0
91.00	1513.0
92.00	1533.0
93.00	10136.0
93.90	628.0
95.10	302.0
96.00	570.0
97.10	330.0
98.00	7462.0
99.00	7430.0
100.00	644.0
100.90	3281.0
102.90	911.0
104.00	2050.0
105.00	1900.0
106.00	747.0
107.00	25344.0
108.00	3967.0
109.00	703.0
110.00	42568.0
111.00	6963.0
112.00	694.0
116.00	1325.0
117.00	22048.0
117.90	1804.0
119.00	441.0
120.00	339.0
122.00	1377.0
123.00	2329.0
123.90	962.0
125.00	1211.0
126.10	402.0
127.00	79472.0
128.00	5905.0
129.00	29008.0
130.00	2677.0
130.90	500.0
132.10	326.0
133.90	903.0
135.00	2345.0
135.90	950.0
136.90	1593.0
137.90	423.0
139.90	485.0
140.90	3779.0
141.90	1235.0
143.00	807.0
146.00	790.0
147.00	2181.0
148.00	4693.0
148.90	1335.0

Instrument :
BNA_N
ClientSampleId :
DFTPP07

150.00	334.0
151.10	562.0
151.80	354.0
153.00	1043.0
154.00	842.0
155.00	2330.0
156.00	3051.0
156.90	629.0
158.00	655.0
159.00	646.0
159.90	1285.0
160.90	1998.0
161.80	561.0
163.90	323.0
164.90	1322.0
166.00	1284.0
167.00	8048.0
168.00	6724.0
169.00	1083.0
169.90	356.0
170.90	359.0
172.00	669.0
172.90	791.0
174.00	1701.0
175.00	3287.0
175.90	803.0
177.00	1449.0
177.90	564.0
178.90	5381.0
180.00	4240.0
181.00	1928.0
181.90	370.0
184.00	556.0
185.00	3018.0
186.00	19280.0
187.00	5940.0
188.00	832.0
188.90	1336.0
191.00	772.0
191.90	1837.0
192.90	2055.0
194.00	371.0
196.00	3766.0
197.00	911.0
198.00	146112.0
199.00	10924.0
199.90	896.0
201.00	303.0
201.50	484.0
202.90	973.0
204.00	5589.0
205.00	8856.0
206.00	37792.0
207.00	6377.0
208.00	1452.0
209.00	565.0
209.90	588.0
210.90	1456.0
211.80	377.0
215.00	378.0
215.90	743.0
216.90	9548.0
217.90	1456.0
221.00	5773.0
221.90	1362.0
222.90	2049.0
224.00	19936.0
225.00	5591.0
226.00	744.0
227.00	8060.0
228.00	1290.0
229.00	1841.0
230.90	756.0
234.00	482.0
235.00	554.0
235.90	374.0
237.00	1054.0
238.90	333.0
240.00	453.0
241.00	496.0
241.90	987.0
243.00	1175.0
244.00	14985.0
245.00	1840.0
246.00	3003.0
247.00	748.0
248.80	522.0
252.90	431.0
254.00	618.0
255.00	72240.0
256.00	10559.0
257.00	1007.0
258.00	4216.0
258.90	649.0
264.90	1491.0

Instrument :
BNA_N
ClientSampleId :
DFTPP07

273.00	1959.0
274.00	5580.0
275.00	30576.0
276.00	4472.0
277.00	2458.0
278.00	499.0
285.00	564.0
292.90	724.0
296.00	8682.0
297.00	1200.0
303.00	912.0
304.00	313.0
314.00	374.0
314.90	809.0
316.00	519.0
323.00	2735.0
324.00	429.0
326.90	568.0
334.00	1505.0
335.10	361.0
345.90	589.0
351.90	790.0
353.00	503.0
354.00	854.0
365.00	2880.0
365.90	439.0
371.90	1262.0
372.90	310.0
383.00	506.0
402.00	489.0
402.90	606.0
404.00	313.0
421.00	598.0
422.00	502.0
423.00	4292.0
424.00	1780.0
424.90	352.0
441.00	9668.0
442.00	57488.0
443.00	12018.0
444.00	1171.0

Instrument :
BNA_N
ClientSampleId :
DFTPP07

m/z	Abundance
37.00	640.0
38.00	1527.0
39.00	9899.0
40.00	1552.0
41.00	712.0
43.00	539.0
44.00	4293.0
45.00	451.0
49.00	1058.0
50.00	33048.0
51.00	129600.0
52.00	6800.0
53.00	364.0
55.00	1175.0
56.00	3627.0
57.00	8353.0
57.90	407.0
61.00	1371.0
62.00	1799.0
63.00	4652.0
64.00	748.0
65.00	2405.0
66.90	359.0
68.00	1963.0
69.00	107848.0
70.00	668.0
73.00	1459.0
74.00	11610.0
75.00	17264.0
76.00	6301.0
77.00	128624.0
78.00	9074.0
79.00	7920.0
80.00	5774.0
81.00	8887.0
82.00	2267.0
83.00	2148.0
83.90	409.0
85.00	1466.0
86.00	2483.0
87.00	1096.0
87.90	450.0
91.00	2050.0
92.00	2041.0
93.00	13744.0
94.00	1022.0
94.90	455.0
96.00	574.0
97.00	339.0
98.00	10329.0
99.00	9210.0
99.90	819.0
101.00	4645.0
101.90	306.0
102.90	1543.0
103.90	2749.0
105.00	2694.0
106.00	993.0
107.00	35464.0
108.00	5944.0
109.00	1102.0
110.00	61184.0
111.00	9269.0
112.00	1360.0
112.90	390.0
116.00	1838.0
117.00	31360.0
118.00	2450.0
119.00	531.0
119.90	541.0
121.90	1985.0
123.00	3525.0
123.90	1486.0
125.00	1406.0
126.10	589.0
127.00	113280.0
128.00	8765.0
129.00	42472.0
130.00	3446.0
131.00	833.0
132.00	374.0
132.90	310.0
133.90	1330.0
135.00	3586.0
136.00	1349.0
137.00	2278.0
137.70	473.0
139.90	444.0
141.00	5195.0
141.90	1894.0
142.90	1217.0

Instrument :
BNA_N
ClientSampleId :
DFTPP07

143.90	332.0
144.90	355.0
146.00	1044.0
147.00	2804.0
148.00	7691.0
149.00	1469.0
149.90	387.0
151.00	784.0
151.80	601.0
152.90	1863.0
154.00	1291.0
155.00	2898.0
156.00	4602.0
157.00	851.0
157.90	993.0
159.00	1010.0
159.90	2049.0
161.00	2481.0
162.00	752.0
164.00	347.0
164.90	2222.0
165.90	1895.0
167.00	12461.0
168.00	9061.0
169.00	1052.0
170.00	304.0
170.90	517.0
172.00	1111.0
173.00	1289.0
174.00	2484.0
175.00	4604.0
176.00	1423.0
176.90	2179.0
177.80	720.0
178.90	7911.0
180.00	6709.0
181.00	2829.0
182.00	499.0
182.90	451.0
183.90	621.0
185.00	4368.0
186.00	32360.0
187.00	8813.0
188.00	1294.0
188.90	1947.0
190.00	348.0
190.90	932.0
192.00	2458.0
193.00	3137.0
194.00	494.0
195.00	351.0
196.00	6370.0
197.00	1452.0
197.90	227520.0
199.00	15828.0
200.00	996.0
201.50	927.0
203.00	1531.0
204.00	8348.0
205.00	14368.0
206.00	57928.0
207.00	9343.0
207.90	2188.0
209.00	890.0
210.00	930.0
211.00	2248.0
214.90	618.0
216.00	1334.0
216.90	15369.0
217.90	2059.0
221.00	9277.0
221.90	2058.0
223.00	3293.0
224.00	32648.0
225.00	8241.0
226.00	968.0
227.00	13198.0
228.00	1752.0
228.90	2892.0
230.00	457.0
231.00	1144.0
233.90	766.0
235.00	1006.0
235.90	684.0
236.90	1687.0
238.90	499.0
239.90	415.0
240.90	740.0
242.00	1742.0
243.00	2033.0
244.00	25744.0
245.00	3642.0
246.00	4608.0
246.90	991.0
248.90	995.0

Instrument :
BNA_N
ClientSampleId :
DFTPP07

252.90	627.0
254.10	916.0
255.00	128144.0
256.00	18448.0
256.90	1537.0
258.00	7207.0
258.90	1154.0
264.90	2622.0
265.90	503.0
272.00	377.0
273.00	4027.0
274.00	10526.0
275.00	56560.0
276.00	8054.0
277.00	4825.0
277.90	855.0
283.00	604.0
284.10	332.0
285.00	905.0
293.00	1227.0
294.00	381.0
295.10	350.0
296.00	15279.0
297.00	1935.0
303.00	1595.0
304.00	621.0
310.00	307.0
314.00	690.0
315.00	1757.0
315.90	993.0
320.90	464.0
322.10	305.0
323.00	5129.0
324.00	885.0
326.90	907.0
328.00	491.0
332.00	343.0
333.00	462.0
334.00	3310.0
335.00	891.0
341.00	545.0
346.00	985.0
352.00	1575.0
353.00	974.0
354.00	1634.0
364.90	6791.0
366.00	948.0
371.00	441.0
372.00	2464.0
373.00	395.0
383.00	754.0
401.90	915.0
403.00	1189.0
404.00	591.0
421.00	1351.0
422.00	1249.0
423.00	8554.0
424.00	2761.0
425.00	395.0
441.00	22776.0
442.00	139648.0
443.00	26776.0
444.00	2572.0

Instrument :
BNA_N
ClientSampleId :
DFTPP07

m/z	Abundance
35.90	320.0
36.90	463.0
38.00	1513.0
39.00	8334.0
40.00	1411.0
41.00	742.0
43.00	417.0
44.00	4390.0
44.90	359.0
49.00	898.0
50.00	26960.0
51.00	105240.0
52.00	5240.0
55.00	958.0
56.00	2805.0
57.00	6769.0
60.90	1152.0
62.00	1377.0
63.00	3909.0
64.00	444.0
65.00	2158.0
68.00	1453.0
69.00	91232.0
70.00	571.0
72.90	1143.0
74.00	9215.0
75.00	15046.0
76.00	5138.0
77.00	109792.0
78.00	7765.0
79.00	6416.0
80.00	5059.0
81.00	7598.0
82.00	1849.0
83.00	2091.0
83.90	493.0
84.90	1237.0
86.00	1938.0
87.00	1119.0
88.00	393.0
91.00	1765.0
92.00	1765.0
93.00	11360.0
94.00	737.0
95.00	326.0
96.00	677.0
97.20	354.0
98.00	8755.0
99.00	7432.0
100.00	622.0
101.00	3552.0
103.00	1055.0
103.90	2154.0
105.00	2507.0
106.00	814.0
107.00	30232.0
108.00	4735.0
109.00	893.0
110.00	51888.0
111.00	7686.0
112.00	1016.0
116.00	1358.0
117.00	27424.0
118.00	2210.0
119.00	370.0
120.00	448.0
121.90	1884.0
123.00	2637.0
123.90	1431.0
124.90	1232.0
127.00	96232.0
128.00	7538.0
129.00	35800.0
130.00	3225.0
130.90	595.0
132.00	445.0
133.90	1076.0
135.00	3309.0
136.00	1275.0
137.00	1983.0
137.90	325.0
140.00	488.0
141.00	4706.0
141.90	1615.0
143.00	1302.0
143.90	450.0
145.00	316.0
146.00	989.0
147.00	2605.0
147.90	6480.0
148.90	1581.0

Instrument :
BNA_N
ClientSampleId :
DFTPP07

149.90	483.0
150.90	538.0
151.50	455.0
153.00	1452.0
154.00	1141.0
155.00	3108.0
156.00	4291.0
157.00	940.0
158.00	774.0
159.00	870.0
159.90	1578.0
161.00	2593.0
162.00	711.0
164.00	344.0
165.00	2134.0
166.00	1315.0
167.00	11562.0
168.00	7424.0
168.90	1209.0
170.00	449.0
171.00	492.0
171.90	882.0
173.00	1421.0
174.00	2101.0
175.00	4028.0
176.00	1123.0
176.90	1943.0
177.90	623.0
178.90	7521.0
180.00	5148.0
181.00	2415.0
182.00	412.0
184.00	761.0
185.00	3803.0
186.00	28856.0
187.00	8388.0
188.00	1088.0
188.90	1866.0
191.00	949.0
192.00	2604.0
193.00	3315.0
194.00	643.0
194.90	427.0
196.00	5850.0
197.00	1560.0
197.90	211392.0
198.90	14432.0
200.00	1419.0
201.30	902.0
202.90	1235.0
204.00	7731.0
205.00	13407.0
206.00	55096.0
207.00	8218.0
208.00	1983.0
209.00	810.0
210.00	830.0
210.90	2362.0
214.90	694.0
216.00	1383.0
216.90	14549.0
218.00	1906.0
220.10	302.0
221.00	8813.0
221.80	1951.0
223.00	3258.0
224.00	30504.0
225.00	8043.0
226.10	891.0
227.00	12417.0
228.00	1703.0
228.90	2555.0
230.00	382.0
230.90	992.0
233.90	659.0
234.90	1013.0
235.90	630.0
237.00	1417.0
238.90	527.0
239.90	475.0
240.90	774.0
242.00	1561.0
243.00	1558.0
244.00	24840.0
245.00	3139.0
246.00	4498.0
246.90	986.0
249.00	806.0
252.90	535.0
254.10	1367.0
255.00	123888.0
256.00	17488.0
257.00	1716.0
258.00	6792.0
258.90	1044.0

Instrument :
BNA_N
ClientSampleId :
DFTPP07

265.00	2902.0
266.00	455.0
273.00	3814.0
274.00	10155.0
275.00	55064.0
276.00	7297.0
277.00	4293.0
277.90	854.0
283.00	569.0
284.00	321.0
285.00	833.0
293.00	1197.0
294.00	310.0
295.00	363.0
296.00	15115.0
297.00	2088.0
302.00	307.0
303.00	1811.0
303.90	507.0
314.00	894.0
314.90	1927.0
316.00	790.0
321.00	496.0
323.10	4953.0
323.90	923.0
326.90	998.0
328.00	570.0
332.00	358.0
332.90	431.0
334.00	3303.0
335.00	914.0
341.00	533.0
346.00	1144.0
352.00	1426.0
353.00	1173.0
354.00	1605.0
355.00	409.0
365.00	7539.0
366.00	976.0
371.00	558.0
372.00	2872.0
373.00	659.0
382.90	771.0
401.90	1023.0
402.90	1670.0
404.00	449.0
420.90	1370.0
422.00	1480.0
423.00	9923.0
424.00	2876.0
441.00	25608.0
442.00	167040.0
443.00	32936.0
444.00	2989.0

Instrument :
BNA_N
ClientSampleId :
DFTPP07