

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN071919\
 Data File : BN006929.D
 Acq On : 19 Jul 2019 18:03
 Operator : HP/JU
 Sample : K3823-12
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :

Quant Time: Jul 19 23:33:30 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN062719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 19 23:29:55 2019
 Response via : Initial Calibration


Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	2358	0.40	ng/ul	-0.02
2) Naphthalene-d8	10.35	136	8422	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.24	164	4251	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.09	188	913334	0.40	ng/ul	0.08
16) Chrysene-d12	0.00	240	0	0.00	ng/ul	-21.24
20) Perylene-d12	23.30	264	871030	0.40	ng/ul	-0.17
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.99	152	3891	0.30	ng/ul	0.00
14) Fluoranthene-d10	0.00	212	0	0.00	ng/ul	
Target Compounds						
9) Fluorene	15.23	166	1489	0.080	ng/ul#	15


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN071919\
Data File : BN006929.D
Acq On : 19 Jul 2019 18:03
Operator : HP/JU
Sample : K3823-12
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :

Quant Time: Jul 19 23:33:30 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN062719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 19 23:29:55 2019
Response via : Initial Calibration





#9

Fluorene

Concen: 0.080 ng/ul

RT: 15.23 min Scan# 3264

Delta R.T. -0.08 min

Lab File: BN006929.D

Acq: 19 Jul 2019 18:03

Instrument :

BNA_N

ClientSampleId :


Tot Ion:166 Resp: 1489

Ion Ratio Lower Upper

166 100

165 9.3 78.2 117.4#

167 35.6 11.3 16.9#



Abundance Ion 166.00 (165.70 to 166.70): E

Ion 165.00 (164.70 to 165.70): E

Ion 167.00 (166.70 to 167.70): E

