

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN071919\
 Data File : BN006931.D
 Acq On : 19 Jul 2019 19:13
 Operator : HP/JU
 Sample : K3823-16
 Misc :
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 A52R6

Manual Integrations
 APPROVED

mohammad
 7/22/2019 10:03:09 AM

Quant Time: Jul 20 00:16:36 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN062719.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 19 23:29:55 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.59	152	2418	0.40	ng/ul	-0.02
2) Naphthalene-d8	10.35	136	8662	0.40	ng/ul	-0.02
6) Acenaphthene-d10	14.24	164	4362	0.40	ng/ul	0.00
10) Phenanthrene-d10	17.00	188	9172m	0.40	ng/ul	-0.02
16) Chrysene-d12	21.24	240	7540m	0.40	ng/ul	0.00
20) Perylene-d12	23.46	264	6525m	0.40	ng/ul	-0.01
System Monitoring Compounds						
4) 2-Methylnaphthalene-d10	11.99	152	3568	0.27	ng/ul	0.00
14) Fluoranthene-d10	19.05	212	9901m	0.37	ng/ul	0.00
Target Compounds						
3) Naphthalene	10.40	128	708	0.028	ng/ul#	Ovalue 65

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN071919\
Data File : BN006931.D
Acq On : 19 Jul 2019 19:13
Operator : HP/JU
Sample : K3823-16
Misc :
ALS Vial : 12 Sample Multiplier: 1




Instrument :
BNA_N
ClientSampleId :
A52R6

Manual Integrations
APPROVED

mohammad
7/22/2019 10:03:09 AM

Quant Time: Jul 20 00:16:36 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\SOM-EPA-SIM-BN062719.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 19 23:29:55 2019
Response via : Initial Calibration





#3

Naphthalene

Concen: 0.028 ng/ul

RT: 10.40 min Scan# 2323

Delta R.T. -0.02 min

Lab File: BN006931.D

Acq: 19 Jul 2019 19:13

Instrument :

BNA_N

ClientSampleId :

A52R6

Tot Ion:128 Resp: 708

Ion Ratio Lower Upper

128 100

129 29.6 9.4 14.0#

127 23.6 10.6 16.0#

Manual Integrations

APPROVED

mohammad

7/22/2019 10:03:09 AM

