


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072018\
Data File : BN002090.D
Acq On : 20 Jul 2018 08:57
Operator : JU/SJ
Sample : DFTPP93
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP93

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN071318MA.M
Title : SVOA CALIBRATION
Last Update : Fri Jul 20 23:55:09 2018



AutoFind: Scans 2472, 2473, 2474; Background Corrected with Scan 2466

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	34.4	162376	PASS
68	69	0.00	2	1.6	2637	PASS
69	198	0.00	100	36.1	170056	PASS
70	69	0.00	2	0.5	802	PASS
127	198	10	80	45.6	214976	PASS
197	198	0.00	2	0.4	1778	PASS
198	198	100	100	100.0	471488	PASS
199	198	5	9	6.9	32432	PASS
275	198	10	60	27.2	128298	PASS
365	198	1	100	3.2	15320	PASS
441	198	0.01	100	15.2	71661	PASS
442	198	50	100	99.2	467690	PASS
443	442	15	24	19.7	91986	PASS

m/z	Abundance
37.10	764.0
38.00	1879.0
39.00	12902.0
40.00	1447.0
41.00	420.0
44.00	4697.0
45.10	465.0
49.10	1208.0
50.00	45968.0
51.00	177600.0
52.00	10034.0
53.00	494.0
55.00	1003.0
56.00	4944.0
57.00	12978.0
57.90	657.0
61.00	2243.0
62.00	3015.0
63.00	7469.0
63.90	1090.0
65.00	4029.0
65.90	416.0
68.00	3046.0
69.00	184704.0
70.00	986.0
73.00	1726.0
74.00	17760.0
75.00	29256.0
76.00	9745.0
77.00	205120.0
78.00	13090.0
79.00	13804.0
80.00	10345.0
81.00	13355.0
82.00	4005.0
83.00	3714.0
83.90	491.0
84.90	2615.0
86.00	3932.0
87.00	1850.0
88.00	471.0
88.80	379.0
90.90	2980.0
92.00	4240.0
93.00	22640.0
93.90	2034.0
96.00	1085.0
96.90	531.0
98.00	17376.0
99.00	14480.0
100.00	1113.0
101.00	7583.0
101.80	361.0
102.90	2825.0
104.00	5459.0
105.00	4490.0
106.00	1584.0
107.00	61096.0
108.00	9362.0
109.00	2024.0
110.00	114432.0
111.00	18440.0
112.00	1988.0
113.00	690.0
116.00	3246.0
117.00	44720.0
118.00	3180.0
118.90	533.0
119.90	889.0
120.80	383.0
122.00	4124.0
123.00	5748.0
123.90	2832.0
124.90	2807.0
126.10	1061.0
127.00	224000.0
128.00	16704.0
129.00	83312.0
129.90	6954.0
131.00	1452.0
132.10	642.0
132.70	329.0
133.90	2299.0
135.00	7335.0
135.90	2648.0
137.00	3723.0
138.90	378.0
139.90	913.0
140.90	11583.0
141.90	3892.0
142.90	2215.0

Instrument :
BNA_N
ClientSampleId :
DFTPP93

143.90	610.0
144.90	784.0
146.00	1727.0
146.90	5296.0
147.90	11193.0
149.00	2540.0
150.00	611.0
151.00	1502.0
151.80	892.0
152.90	3797.0
154.00	2597.0
155.00	6373.0
155.90	9573.0
157.00	1976.0
157.90	1539.0
159.00	1629.0
160.00	3114.0
160.90	4645.0
161.90	1392.0
162.80	497.0
164.10	682.0
164.90	3545.0
165.90	3250.0
167.00	20248.0
168.00	10402.0
169.00	1713.0
170.10	834.0
171.00	933.0
172.00	1806.0
172.90	2123.0
174.00	4880.0
175.00	9163.0
176.00	2840.0
176.90	4014.0
177.90	1509.0
179.00	15455.0
180.00	11861.0
181.00	5481.0
182.00	968.0
183.00	467.0
184.00	953.0
185.00	7785.0
186.00	60552.0
187.00	17592.0
188.00	1733.0
189.00	4334.0
189.90	566.0
191.00	1595.0
192.00	5213.0
193.00	5970.0
194.10	1031.0
195.00	977.0
196.00	11563.0
197.90	458752.0
199.00	31672.0
200.00	2617.0
201.50	2137.0
203.00	3119.0
204.00	16928.0
205.00	27216.0
206.00	116592.0
207.00	16440.0
208.00	3979.0
208.90	1231.0
210.00	1357.0
210.90	4484.0
213.00	356.0
215.00	1062.0
216.00	2439.0
217.00	29792.0
218.00	3646.0
219.00	457.0
221.10	17584.0
221.80	3658.0
222.90	6436.0
224.00	63600.0
225.00	15359.0
226.00	1655.0
227.00	26312.0
227.90	3645.0
229.00	5282.0
230.00	658.0
231.00	2172.0
232.80	406.0
234.00	1817.0
235.00	1791.0
235.90	1470.0
237.00	2571.0
237.90	421.0
238.80	836.0
239.90	902.0
240.90	1601.0
242.00	3829.0
243.00	3282.0
244.00	48160.0

Instrument :
BNA_N
ClientSampleId :
DFTPP93

245.00	7113.0
246.00	9794.0
247.00	2159.0
247.80	408.0
248.10	341.0
248.90	1576.0
250.00	308.0
251.00	354.0
251.90	434.0
252.90	972.0
255.00	253824.0
256.00	38728.0
257.00	3011.0
258.00	13238.0
259.00	2447.0
260.00	304.0
261.00	473.0
263.90	306.0
265.00	6139.0
266.00	934.0
271.10	578.0
271.80	596.0
273.00	7396.0
274.00	19384.0
275.00	115880.0
276.00	14404.0
276.90	9558.0
278.00	1532.0
280.90	343.0
281.80	362.0
283.00	1137.0
283.90	771.0
285.00	1877.0
288.80	351.0
290.90	306.0
292.00	439.0
293.00	1987.0
293.90	795.0
296.00	27680.0
297.00	4181.0
300.90	338.0
302.00	494.0
303.00	3401.0
304.00	1094.0
308.00	481.0
309.00	398.0
310.00	594.0
312.90	339.0
314.10	1462.0
315.00	3762.0
315.90	2017.0
317.00	504.0
321.00	1157.0
322.00	463.0
323.00	9944.0
323.90	1984.0
325.80	333.0
326.90	1679.0
327.90	923.0
329.10	370.0
331.90	505.0
333.00	923.0
334.00	5278.0
335.00	1395.0
341.00	1330.0
341.80	379.0
346.00	2147.0
346.90	468.0
352.00	2566.0
353.00	2775.0
353.90	2842.0
355.00	849.0
365.00	12947.0
366.00	1812.0
370.90	864.0
372.00	4975.0
373.00	1352.0
383.00	1973.0
384.00	522.0
390.00	469.0
390.90	455.0
400.90	461.0
402.00	2298.0
403.00	2780.0
404.00	1484.0
421.00	2799.0
422.00	2763.0
423.00	17824.0
424.00	4937.0
424.90	493.0
441.00	55024.0
442.00	355968.0
443.00	72488.0
444.00	7080.0
445.00	577.0

Instrument :
BNA_N
ClientSampleId :
DFTPP93

m/z	Abundance
37.10	739.0
38.00	1967.0
39.00	12719.0
40.00	1250.0
41.00	464.0
44.00	4868.0
44.90	452.0
49.10	1305.0
50.00	48688.0
51.00	194688.0
52.00	10208.0
53.00	470.0
55.00	1037.0
56.00	5970.0
57.00	13271.0
57.90	750.0
61.00	2598.0
62.00	2365.0
63.00	8612.0
64.10	920.0
65.00	4569.0
68.00	2705.0
69.00	200704.0
70.00	938.0
73.00	1629.0
74.00	18648.0
75.00	30032.0
76.00	10224.0
77.00	223104.0
78.00	14509.0
79.00	15457.0
80.00	11859.0
81.00	15853.0
81.90	3869.0
83.00	4345.0
84.00	344.0
85.00	3081.0
86.00	4570.0
87.00	1817.0
87.90	672.0
88.90	448.0
91.00	4058.0
92.00	3831.0
93.00	24304.0
94.10	1783.0
94.90	360.0
95.90	924.0
97.00	468.0
98.00	17392.0
99.00	14172.0
99.90	1372.0
101.00	9409.0
102.00	601.0
102.90	3142.0
103.90	5165.0
104.90	5287.0
106.00	1895.0
107.00	68728.0
108.00	11021.0
109.00	2606.0
110.00	128552.0
111.00	19616.0
111.90	2282.0
113.00	954.0
115.20	354.0
115.90	3521.0
117.00	48992.0
118.00	3698.0
118.90	586.0
119.90	873.0
120.90	373.0
121.90	4665.0
123.00	7026.0
124.00	3118.0
124.90	2824.0
127.00	257600.0
128.00	19744.0
129.00	96960.0
130.00	9020.0
131.00	1968.0
132.00	873.0
132.80	370.0
134.00	2610.0
135.00	7511.0
136.00	3079.0
136.90	3830.0
137.80	998.0
139.00	573.0
139.90	1144.0
140.90	11982.0
142.00	3566.0

Instrument :
BNA_N
ClientSampleId :
DFTPP93

142.90	2688.0
144.00	757.0
144.90	713.0
146.00	2376.0
147.00	6779.0
147.90	13745.0
149.00	3501.0
150.00	619.0
150.90	1665.0
151.30	1599.0
152.00	1053.0
153.00	4091.0
154.00	3386.0
155.00	6988.0
156.00	11793.0
157.00	2417.0
158.00	2359.0
159.00	1890.0
160.00	3919.0
161.00	6296.0
161.90	1576.0
162.90	510.0
164.10	439.0
165.00	4308.0
166.00	4084.0
167.00	23128.0
168.00	11920.0
169.00	2037.0
170.10	787.0
171.00	1053.0
171.90	2434.0
172.90	2413.0
174.00	5907.0
175.00	10888.0
175.90	2760.0
176.90	5038.0
177.90	1705.0
178.90	19760.0
180.00	13157.0
181.00	6899.0
182.00	927.0
183.10	747.0
183.90	2029.0
185.00	9216.0
186.00	72424.0
187.00	21504.0
188.00	2179.0
188.90	4200.0
189.90	684.0
191.00	2229.0
191.90	5693.0
193.00	7008.0
194.00	1074.0
194.90	793.0
196.00	15128.0
197.00	3017.0
197.90	555008.0
199.00	39584.0
200.00	2754.0
201.30	1973.0
201.70	1901.0
203.00	4073.0
204.00	19320.0
205.00	31904.0
206.00	141696.0
207.00	18872.0
208.00	4482.0
208.90	1611.0
209.90	2313.0
210.30	2082.0
211.00	5431.0
211.70	1413.0
212.90	539.0
214.90	1317.0
216.00	3460.0
216.90	36696.0
218.00	4676.0
219.00	602.0
219.70	502.0
221.00	22648.0
221.80	5009.0
223.00	8360.0
224.00	79952.0
225.00	20320.0
226.00	1992.0
227.00	30880.0
228.00	4738.0
229.00	6690.0
229.90	1197.0
231.00	3025.0
232.00	572.0
233.00	902.0
234.10	1674.0
235.00	2068.0
235.90	1491.0

Instrument :
BNA_N
ClientSampleId :
DFTPP93

237.00	2541.0
237.70	447.0
238.90	1217.0
240.00	1236.0
240.90	1824.0
242.00	3470.0
242.90	3878.0
244.00	60504.0
245.00	8202.0
246.00	11837.0
247.00	1847.0
247.90	532.0
249.00	2640.0
249.80	402.0
250.90	617.0
251.80	523.0
253.00	1074.0
255.00	334528.0
256.00	46800.0
257.00	4579.0
258.00	16160.0
258.90	2705.0
259.90	359.0
260.90	599.0
263.90	559.0
265.00	7449.0
265.90	1126.0
270.60	336.0
271.00	358.0
272.10	826.0
273.00	9547.0
274.00	25664.0
275.00	150272.0
276.00	18784.0
277.00	11934.0
277.90	1892.0
278.90	413.0
281.00	351.0
283.00	1334.0
284.00	985.0
285.00	2232.0
286.00	361.0
288.90	614.0
289.70	323.0
290.90	342.0
292.00	515.0
293.00	2924.0
294.00	829.0
294.90	610.0
296.00	38216.0
297.00	4492.0
297.90	563.0
301.10	653.0
301.90	702.0
303.00	5232.0
304.00	1717.0
308.00	431.0
308.90	340.0
310.00	660.0
313.10	396.0
314.00	2018.0
314.90	3893.0
316.00	2464.0
316.90	560.0
321.00	1232.0
322.00	831.0
323.00	12647.0
324.00	2844.0
327.10	2126.0
328.00	1352.0
332.10	821.0
332.90	1080.0
334.00	7806.0
335.00	2113.0
339.00	336.0
339.80	316.0
341.00	1657.0
341.90	589.0
346.00	2975.0
346.90	580.0
352.00	4581.0
353.00	2946.0
354.00	4566.0
355.00	1097.0
359.00	371.0
364.90	18080.0
365.90	2530.0
369.90	467.0
371.00	1273.0
372.00	7826.0
373.10	1799.0
382.90	1992.0
383.90	563.0
390.00	964.0
390.90	539.0

Instrument :
BNA_N
ClientSampleId :
DFTPP93

392.00	688.0
401.10	380.0
402.00	2721.0
403.00	3968.0
404.00	1576.0
421.00	3980.0
421.90	3484.0
423.00	27672.0
424.00	7146.0
425.00	668.0
441.00	83168.0
442.00	560256.0
443.00	107224.0
444.00	11017.0
445.00	644.0

Instrument :
BNA_N
ClientSampleId :
DFTPP93

m/z	Abundance
36.90	442.0
38.00	1286.0
39.00	8585.0
40.00	966.0
41.00	582.0
44.00	5418.0
49.10	779.0
50.00	29112.0
51.00	114840.0
52.00	5488.0
53.10	316.0
54.90	455.0
56.00	3739.0
57.00	8558.0
58.00	459.0
60.90	1700.0
62.00	2033.0
63.00	4918.0
64.00	589.0
65.00	2194.0
68.10	2162.0
69.00	124760.0
70.00	483.0
72.90	1082.0
74.00	12137.0
75.00	19992.0
76.10	6753.0
77.00	139584.0
78.00	9330.0
79.00	8828.0
80.00	7351.0
81.00	10061.0
82.00	2547.0
83.00	2870.0
83.80	385.0
85.00	1851.0
85.90	2963.0
87.00	1391.0
87.90	565.0
91.00	2404.0
92.00	2249.0
93.00	15501.0
94.00	1023.0
96.10	757.0
98.00	11511.0
99.00	9506.0
100.00	687.0
101.00	6031.0
101.90	376.0
103.00	1782.0
103.90	3546.0
104.90	3948.0
106.00	1357.0
107.00	45496.0
108.00	7094.0
109.00	1264.0
110.00	85216.0
111.00	12993.0
112.00	1578.0
113.10	510.0
115.90	2736.0
117.00	32736.0
118.00	2665.0
119.00	368.0
120.00	989.0
121.90	2308.0
123.00	5232.0
124.00	1981.0
125.00	1747.0
126.10	920.0
127.00	163328.0
128.00	12932.0
129.00	62680.0
129.90	5237.0
131.00	970.0
132.00	536.0
133.10	442.0
134.00	1938.0
135.00	5540.0
136.00	2363.0
137.00	2987.0
138.10	581.0
138.80	338.0
139.90	884.0
140.90	7806.0
142.00	2320.0
142.90	1917.0
144.00	400.0
144.80	355.0
145.20	362.0
146.00	1651.0

Instrument :
BNA_N
ClientSampleId :
DFTPP93

146.90	4089.0
147.90	8892.0
149.00	2068.0
149.80	664.0
151.00	1121.0
151.90	958.0
152.90	3172.0
153.90	1761.0
155.00	4856.0
156.00	7883.0
157.10	1363.0
157.80	1325.0
158.90	1469.0
159.90	2206.0
161.00	4214.0
161.90	1024.0
163.80	627.0
164.90	3182.0
166.00	2612.0
167.00	16160.0
168.00	7796.0
169.00	1381.0
169.90	769.0
171.00	696.0
172.00	1839.0
172.90	2095.0
174.00	4825.0
175.00	6998.0
176.10	2031.0
177.00	3289.0
177.90	998.0
178.90	12835.0
180.00	9423.0
181.00	4189.0
181.90	806.0
184.10	1126.0
185.00	6791.0
186.00	48928.0
187.00	14958.0
188.00	1631.0
188.90	3157.0
189.90	672.0
191.00	1583.0
192.00	4079.0
193.00	4419.0
194.00	1093.0
194.90	631.0
195.90	11999.0
196.90	2318.0
197.90	400704.0
199.00	26040.0
199.90	1995.0
201.40	1145.0
202.10	739.0
202.90	2583.0
204.00	13148.0
205.00	23112.0
206.00	104352.0
207.00	13204.0
208.00	3740.0
208.90	1150.0
210.10	1894.0
210.90	4252.0
214.90	937.0
216.00	2086.0
216.90	27392.0
217.90	3500.0
219.00	491.0
221.10	17256.0
221.80	3960.0
223.00	5497.0
224.00	56752.0
225.00	13737.0
226.10	1377.0
227.00	23632.0
228.10	3513.0
228.90	5014.0
229.90	680.0
231.00	2680.0
232.00	377.0
233.00	620.0
233.90	1633.0
235.00	1692.0
236.00	1505.0
237.00	2136.0
238.90	1076.0
239.90	773.0
241.00	1333.0
242.10	3446.0
243.00	3180.0
244.00	44752.0
245.00	5552.0
246.00	9899.0
247.00	1811.0
249.00	1798.0

Instrument :
BNA_N
ClientSampleId :
DFTPP93

249.80	373.0
250.90	412.0
251.90	615.0
252.90	917.0
254.00	1645.0
255.00	247680.0
256.00	36080.0
257.00	2880.0
257.90	12727.0
258.90	1965.0
259.90	330.0
260.90	308.0
264.00	511.0
265.00	5343.0
265.90	626.0
270.90	705.0
271.90	539.0
273.00	6586.0
274.00	18448.0
275.00	118744.0
276.00	15455.0
277.00	9316.0
277.90	1591.0
278.90	304.0
283.00	1199.0
284.00	944.0
285.00	1398.0
286.10	362.0
288.90	372.0
289.80	393.0
290.80	313.0
292.00	516.0
293.00	2107.0
294.00	600.0
296.00	30840.0
297.00	4439.0
298.00	323.0
300.90	453.0
301.90	573.0
303.00	4070.0
304.00	996.0
308.00	610.0
310.00	368.0
313.00	386.0
314.00	1629.0
314.90	3573.0
316.00	1774.0
317.00	554.0
321.00	1335.0
322.00	447.0
323.00	11161.0
324.00	1889.0
326.90	2001.0
328.00	997.0
332.10	850.0
333.00	956.0
334.00	7422.0
335.10	2093.0
341.00	1408.0
341.90	389.0
346.00	2658.0
346.80	515.0
352.00	3552.0
353.00	2120.0
354.00	3824.0
355.10	622.0
359.00	393.0
365.00	14934.0
365.90	2120.0
371.00	1210.0
372.00	6721.0
373.00	1855.0
383.00	1519.0
384.00	628.0
390.00	774.0
390.90	478.0
392.00	610.0
400.80	328.0
401.90	2584.0
402.90	3517.0
403.90	1265.0
404.90	303.0
421.00	3685.0
422.00	3843.0
423.00	26600.0
424.00	6172.0
424.90	700.0
441.00	76792.0
442.00	486848.0
443.00	96248.0
444.00	8889.0
445.00	631.0

Instrument :
BNA_N
ClientSampleId :
DFTPP93

Scan 2466 (17.492 min): BN002090.D
DFTPP93

m/z	Abundance
40.00	803.0
44.00	4450.0
73.00	312.0
190.90	313.0
207.10	945.0
208.00	341.0

Instrument :
BNA_N
ClientSampleId :
DFTPP93