

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072023\
 Data File : BN026647.D
 Acq On : 22 Jul 2023 03:37
 Operator : MA/JU
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD0.4370

Quant Time: Jul 22 04:12:46 2023
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN071223.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 21 19:02:16 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.844	152	4030	0.400	ng/ul	-0.01
4) Naphthalene-d8	10.636	136	14298	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.472	164	8131	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.223	188	15901	0.400	ng/ul	0.00
17) Chrysene-d12	21.414	240	10253	0.400	ng/ul	-0.01
23) Perylene-d12	23.787	264	11615	0.400	ng/ul	#-0.02
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.203	96	1383	0.261	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.231	152	8135	0.419	ng/ul	-0.01
18) Fluoranthene-d10	19.250	212	16919	0.476	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.237	88	1406	0.280	ng/ul#	81
5) Naphthalene	10.685	128	16088	0.410	ng/ul	98
7) 2-Methylnaphthalene	12.302	142	9607	0.431	ng/ul	100
8) 1-Methylnaphthalene	12.517	142	10489	0.440	ng/ul	100
10) Acenaphthylene	14.194	152	17387	0.437	ng/ul	98
11) Acenaphthene	14.532	153	12834	0.423	ng/ul	99
12) Fluorene	15.526	166	13731	0.431	ng/ul	99
14) Pentachlorophenol	16.890	266	1260	0.314	ng/ul	98
15) Phenanthrene	17.261	178	20943	0.414	ng/ul	98
16) Anthracene	17.359	178	19196	0.437	ng/ul	96
19) Fluoranthene	19.283	202	23227	0.483	ng/ul	99
20) Pyrene	19.645	202	24820	0.476	ng/ul	97
21) Benzo(a)anthracene	21.402	228	17035	0.452	ng/ul	99
22) Chrysene	21.454	228	17799	0.415	ng/ul	99
24) Benzo(b)fluoranthene	23.065	252	17536	0.388	ng/ul	91
25) Benzo(k)fluoranthene	23.115	252	19033	0.401	ng/ul#	92
26) Benzo(a)pyrene	23.685	252	16589	0.412	ng/ul#	90
27) Indeno(1,2,3-cd)pyrene	26.244	276	22009	0.417	ng/ul#	96
28) Dibenzo(a,h)anthracene	26.258	278	17209	0.425	ng/ul	90
29) Benzo(g,h,i)perylene	26.999	276	18565	0.398	ng/ul	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072023\
Data File : BN026647.D
Acq On : 22 Jul 2023 03:37
Operator : MA/JU
Sample : SSTDC00.4EC
Misc :
ALS Vial : 60 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD0.4370

Quant Time: Jul 22 04:12:46 2023
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN071223.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 21 19:02:16 2023
Response via : Initial Calibration

