

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN072220\
 Data File : BN011218.D
 Acq On : 21 Jul 2020 13:55
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 21 17:40:42 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 21 14:35:20 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
2	1,4-Dioxane	0.572	0.553	3.3	78	0.00
3	n-Nitrosodimethylamine	0.288	0.270	6.2	86	0.00
4 S	2-Fluorophenol	0.986	0.941	4.6	79	0.00
5 S	Phenol-d6	1.115	1.182	-6.0	90	0.00
6	bis(2-Chloroethyl)ether	1.021	1.100	-7.7	100	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.355	0.328	7.6	90	0.00
9	Naphthalene	1.190	1.111	6.6	92	0.00
10	Hexachlorobutadiene	0.295	0.267	9.5	86	0.00
11 SURR	2-Methylnaphthalene-d10	0.649	0.673	-3.7	97	0.00
12	2-Methylnaphthalene	0.749	0.769	-2.7	97	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	117	0.00
14 S	2,4,6-Tribromophenol	0.178	0.170	4.5	108	0.00
15 S	2-Fluorobiphenyl	1.874	1.638	12.6	95	0.00
16	Acenaphthylene	2.036	1.820	10.6	107	0.00
17	Acenaphthene	1.358	1.253	7.7	107	0.00
18	Fluorene	1.742	1.678	3.7	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	133	0.00
20	4,6-Dinitro-2-methylphenol	0.053	0.043	18.9	106	0.00
21	4-Bromophenyl-phenylether	0.272	0.243	10.7	119	0.00
22	Hexachlorobenzene	0.316	0.269	14.9	111	0.00
23	Atrazine	0.191	0.177	7.3	123	0.00
24	Pentachlorophenol	0.094	0.080	14.9	122	0.00
25	Phenanthrene	1.285	1.192	7.2	123	0.00
26	Anthracene	1.067	0.946	11.3	119	0.00
27 SURR	Fluoranthene-d10	1.261	1.152	8.6	119	0.00
28	Fluoranthene	1.541	1.383	10.3	114	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	106	0.00
30	Pyrene	1.527	1.780	-16.6	109	0.00
31 S	Terphenyl-d14	1.041	1.176	-13.0	113	0.00
32	Benzo(a)anthracene	1.296	1.241	4.2	105	0.00
33	Chrysene	1.584	1.554	1.9	105	0.00
34	Bis(2-ethylhexyl)phthalate	0.534	0.634	-18.7	128	0.00
35	Indeno(1,2,3-cd)pyrene	1.656	1.336	19.3	74	0.00
36 I	Perylene-d12	1.000	1.000	0.0	93	0.00
37	Benzo(b)fluoranthene	1.607	1.481	7.8	95	0.00
38	Benzo(k)fluoranthene	1.816	1.627	10.4	83	0.00
39 C	Benzo(a)pyrene	1.447	1.261	12.9	80	0.00
40	Dibenzo(a,h)anthracene	1.533	1.200	21.7	71	0.00
41	Benzo(g,h,i)perylene	1.683	1.451	13.8	78	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN072220\
Data File : BN011218.D
Acq On : 21 Jul 2020 13:55
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 21 17:40:42 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071120.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 21 14:35:20 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0