

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN072220\  
 Data File : BN011218.D  
 Acq On : 21 Jul 2020 13:55  
 Operator : CG/JU  
 Sample : SSTDCCC0.4  
 Misc :  
 ALS Vial : 2 Sample Multiplier: 1

Instrument :  
 BNA\_N  
 LabSampleId :  
 SSTDCCC0.4

Quant Time: Jul 21 17:40:42 2020  
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071120.M  
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
 QLast Update : Tue Jul 21 14:35:20 2020  
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
 Max. RRF Dev : 25% Max. Rel. Area : 150%

|         | Compound                   | Amount | Calc. | %Dev  | Area% | Dev(min) |
|---------|----------------------------|--------|-------|-------|-------|----------|
| 1 I     | 1,4-Dichlorobenzene-d4     | 0.400  | 0.400 | 0.0   | 87    | 0.00     |
| 2       | 1,4-Dioxane                | 0.400  | 0.387 | 3.3   | 78    | 0.00     |
| 3       | n-Nitrosodimethylamine     | 0.400  | 0.376 | 6.0   | 86    | 0.00     |
| 4 S     | 2-Fluorophenol             | 0.400  | 0.381 | 4.8   | 79    | 0.00     |
| 5 S     | Phenol-d6                  | 0.400  | 0.424 | -6.0  | 90    | 0.00     |
| 6       | bis(2-Chloroethyl)ether    | 0.400  | 0.431 | -7.7  | 100   | 0.00     |
| 7 I     | Naphthalene-d8             | 0.400  | 0.400 | 0.0   | 98    | 0.00     |
| 8 S     | Nitrobenzene-d5            | 0.400  | 0.371 | 7.3   | 90    | 0.00     |
| 9       | Naphthalene                | 0.400  | 0.374 | 6.5   | 92    | 0.00     |
| 10      | Hexachlorobutadiene        | 0.400  | 0.362 | 9.5   | 86    | 0.00     |
| 11 SURR | 2-Methylnaphthalene-d10    | 0.400  | 0.415 | -3.7  | 97    | 0.00     |
| 12      | 2-Methylnaphthalene        | 0.400  | 0.411 | -2.7  | 97    | 0.00     |
| 13 I    | Acenaphthene-d10           | 0.400  | 0.400 | 0.0   | 117   | 0.00     |
| 14 S    | 2,4,6-Tribromophenol       | 0.400  | 0.382 | 4.5   | 108   | 0.00     |
| 15 S    | 2-Fluorobiphenyl           | 0.400  | 0.350 | 12.5  | 95    | 0.00     |
| 16      | Acenaphthylene             | 0.400  | 0.358 | 10.5  | 107   | 0.00     |
| 17      | Acenaphthene               | 0.400  | 0.369 | 7.8   | 107   | 0.00     |
| 18      | Fluorene                   | 0.400  | 0.385 | 3.8   | 99    | 0.00     |
| 19 I    | Phenanthrene-d10           | 0.400  | 0.400 | 0.0   | 133   | 0.00     |
| 20      | 4,6-Dinitro-2-methylphenol | 0.400  | 0.363 | 9.3   | 106   | 0.00     |
| 21      | 4-Bromophenyl-phenylether  | 0.400  | 0.358 | 10.5  | 119   | 0.00     |
| 22      | Hexachlorobenzene          | 0.400  | 0.340 | 15.0  | 111   | 0.00     |
| 23      | Atrazine                   | 0.400  | 0.371 | 7.3   | 123   | 0.00     |
| 24      | Pentachlorophenol          | 0.400  | 0.340 | 15.0  | 122   | 0.00     |
| 25      | Phenanthrene               | 0.400  | 0.371 | 7.3   | 123   | 0.00     |
| 26      | Anthracene                 | 0.400  | 0.354 | 11.5  | 119   | 0.00     |
| 27 SURR | Fluoranthene-d10           | 0.400  | 0.365 | 8.8   | 119   | 0.00     |
| 28      | Fluoranthene               | 0.400  | 0.359 | 10.3  | 114   | 0.00     |
| 29 I    | Chrysene-d12               | 0.400  | 0.400 | 0.0   | 106   | 0.00     |
| 30      | Pyrene                     | 0.400  | 0.466 | -16.5 | 109   | 0.00     |
| 31 S    | Terphenyl-d14              | 0.400  | 0.452 | -13.0 | 113   | 0.00     |
| 32      | Benzo(a)anthracene         | 0.400  | 0.383 | 4.3   | 105   | 0.00     |
| 33      | Chrysene                   | 0.400  | 0.392 | 2.0   | 105   | 0.00     |
| 34      | Bis(2-ethylhexyl)phthalate | 0.400  | 0.475 | -18.7 | 128   | 0.00     |
| 35      | Indeno(1,2,3-cd)pyrene     | 0.400  | 0.313 | 21.8  | 74    | 0.00     |
| 36 I    | Perylene-d12               | 0.400  | 0.400 | 0.0   | 93    | 0.00     |
| 37      | Benzo(b)fluoranthene       | 0.400  | 0.369 | 7.8   | 95    | 0.00     |
| 38      | Benzo(k)fluoranthene       | 0.400  | 0.358 | 10.5  | 83    | 0.00     |
| 39 C    | Benzo(a)pyrene             | 0.400  | 0.349 | 12.8  | 80    | 0.00     |
| 40      | Dibenzo(a,h)anthracene     | 0.400  | 0.313 | 21.8  | 71    | 0.00     |
| 41      | Benzo(g,h,i)perylene       | 0.400  | 0.345 | 13.8  | 78    | 0.00     |

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN072220\  
Data File : BN011218.D  
Acq On : 21 Jul 2020 13:55  
Operator : CG/JU  
Sample : SSTDCCC0.4  
Misc :  
ALS Vial : 2 Sample Multiplier: 1

Instrument :  
BNA\_N  
LabSampleId :  
SSTDCCC0.4

Quant Time: Jul 21 17:40:42 2020  
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071120.M  
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION  
QLast Update : Tue Jul 21 14:35:20 2020  
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min  
Max. RRF Dev : 25% Max. Rel. Area : 150%

| Compound | Amount | Calc. | %Dev | Area% | Dev(min) |
|----------|--------|-------|------|-------|----------|
|----------|--------|-------|------|-------|----------|

(#) = Out of Range

SPCC's out = 0 CCC's out = 0