


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN072318\
Data File : BN002127.D
Acq On : 24 Jul 2018 08:33
Operator : JU/SJ
Sample : DFTPP95
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP95

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN071318MA.M
Title : SVOA CALIBRATION
Last Update : Tue Jul 24 12:25:22 2018



AutoFind: Scans 2471, 2472, 2473; Background Corrected with Scan 2465

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	38.0	198997	PASS
68	69	0.00	2	1.6	3134	PASS
69	198	0.00	100	38.5	201770	PASS
70	69	0.00	2	0.6	1151	PASS
127	198	10	80	46.8	245568	PASS
197	198	0.00	2	0.4	2053	PASS
198	198	100	100	100.0	524181	PASS
199	198	5	9	6.9	36013	PASS
275	198	10	60	26.2	137370	PASS
365	198	1	100	3.0	15890	PASS
441	198	0.01	100	13.6	71104	PASS
442	198	50	100	91.4	479232	PASS
443	442	15	24	19.4	93202	PASS

m/z	Abundance
37.00	567.0
38.10	1810.0
39.00	11312.0
40.00	1329.0
41.10	508.0
44.00	4344.0
45.00	451.0
49.00	1060.0
50.00	42256.0
51.00	172032.0
52.00	8695.0
52.90	452.0
55.00	1076.0
56.00	5213.0
57.00	12452.0
57.90	523.0
61.00	2091.0
62.00	2534.0
63.00	6879.0
64.00	824.0
65.00	3265.0
68.00	2857.0
69.00	169664.0
70.00	763.0
73.00	1233.0
74.00	15765.0
75.00	26656.0
76.00	7485.0
77.00	184320.0
78.00	12646.0
79.00	11676.0
80.00	9025.0
81.00	13181.0
82.00	3425.0
83.00	2649.0
83.90	462.0
85.00	2294.0
86.00	3485.0
87.00	1567.0
87.90	427.0
90.90	3203.0
91.90	3092.0
92.90	18840.0
94.00	1345.0
95.00	445.0
96.00	984.0
98.00	13790.0
99.00	12092.0
99.90	1464.0
101.00	7703.0
102.10	590.0
103.00	2303.0
104.00	5447.0
105.00	4371.0
105.90	1306.0
107.00	51424.0
108.00	8540.0
109.10	1456.0
110.00	101720.0
111.00	14639.0
112.00	1796.0
112.90	594.0
114.80	375.0
116.00	3020.0
117.00	35664.0
118.00	2828.0
119.00	381.0
119.80	707.0
121.00	371.0
121.90	3306.0
123.00	5619.0
123.90	2413.0
125.00	2038.0
126.10	971.0
127.00	189376.0
128.00	14871.0
129.00	71256.0
130.00	6013.0
131.00	1170.0
132.10	571.0
132.90	407.0
134.00	1914.0
135.00	5639.0
135.90	2253.0
137.00	3373.0
138.00	497.0
140.00	733.0
140.90	8776.0
142.00	3562.0
142.90	1821.0
143.90	733.0

Instrument :
BNA_N
ClientSampleId :
DFTPP95

145.00	357.0
145.90	1513.0
147.00	5401.0
147.90	8895.0
149.00	2367.0
150.00	734.0
150.90	1278.0
151.60	729.0
152.90	2824.0
153.90	2135.0
155.00	5210.0
156.00	7524.0
156.90	1356.0
158.10	1543.0
159.00	1407.0
160.00	3069.0
160.90	4302.0
161.90	1509.0
162.90	446.0
163.90	621.0
164.90	3487.0
166.00	2866.0
167.00	15734.0
168.00	10010.0
169.00	1385.0
170.00	736.0
170.90	625.0
171.90	1564.0
172.90	2188.0
174.00	3848.0
175.00	6775.0
176.00	2529.0
177.00	3742.0
177.90	1177.0
178.90	13645.0
180.00	10130.0
181.00	4258.0
181.90	723.0
182.90	457.0
184.00	1178.0
185.00	6295.0
186.00	52376.0
187.00	15012.0
187.90	1337.0
189.00	3397.0
189.90	638.0
190.90	2092.0
191.90	4354.0
193.00	5129.0
193.90	1163.0
195.00	729.0
196.00	11094.0
197.90	371968.0
198.90	25616.0
200.00	2492.0
201.20	1141.0
201.60	947.0
203.00	2461.0
204.00	12700.0
205.00	22096.0
206.00	95528.0
207.00	13651.0
208.00	3324.0
209.00	1302.0
209.90	1162.0
210.90	3836.0
212.00	567.0
213.00	343.0
215.00	931.0
216.00	1868.0
216.90	25760.0
218.00	3122.0
218.80	315.0
221.00	12116.0
222.90	4952.0
224.00	51688.0
225.00	12726.0
226.00	1458.0
227.00	20864.0
228.00	2720.0
228.90	4159.0
230.00	418.0
230.90	1948.0
232.00	513.0
233.90	903.0
234.90	1509.0
235.90	1164.0
237.00	2091.0
238.00	360.0
238.90	907.0
240.00	991.0
240.90	1205.0
242.00	2090.0
243.10	3146.0
244.00	38448.0

Instrument :
BNA_N
ClientSampleId :
DFTPP95

245.00	5452.0
246.00	7568.0
246.90	1444.0
247.90	556.0
249.00	1505.0
250.90	311.0
252.00	346.0
252.90	771.0
255.00	197184.0
256.00	31680.0
257.00	2895.0
257.90	9791.0
258.90	1764.0
259.90	493.0
261.00	456.0
264.90	3712.0
266.00	651.0
271.00	323.0
272.00	504.0
273.00	5774.0
274.00	15474.0
275.00	86416.0
276.00	12003.0
277.00	6920.0
278.00	1008.0
279.00	322.0
282.90	1047.0
283.90	629.0
284.90	1349.0
286.10	323.0
291.90	624.0
292.90	1691.0
293.90	400.0
294.90	442.0
296.00	21840.0
297.00	3191.0
297.90	302.0
301.00	400.0
301.90	591.0
303.10	2477.0
304.10	772.0
308.00	358.0
312.80	324.0
313.90	1111.0
315.00	2391.0
316.00	1488.0
316.90	398.0
320.90	748.0
322.10	333.0
323.00	7142.0
323.90	1678.0
326.90	1638.0
328.00	593.0
332.00	554.0
333.00	864.0
334.00	4989.0
335.00	1248.0
339.90	313.0
341.00	948.0
345.80	769.0
352.00	1896.0
353.00	1520.0
354.00	2546.0
355.10	580.0
365.00	9910.0
365.90	1693.0
370.90	697.0
372.10	4152.0
373.10	888.0
383.00	1187.0
389.90	492.0
390.90	439.0
400.80	328.0
401.90	1295.0
403.00	2150.0
404.00	933.0
405.00	333.0
421.00	2367.0
421.90	1962.0
423.00	12883.0
424.00	3543.0
425.10	511.0
441.00	38584.0
442.00	248768.0
443.00	50792.0
444.00	5016.0

Instrument :
BNA_N
ClientSampleId :
DFTPP95

m/z	Abundance
37.00	951.0
38.00	2777.0
39.00	17056.0
40.00	1334.0
41.00	565.0
44.00	3918.0
45.00	552.0
49.10	1455.0
50.00	56200.0
51.00	238016.0
52.00	12676.0
52.90	582.0
54.90	1389.0
55.90	7086.0
57.00	16300.0
57.90	810.0
61.00	2592.0
62.10	3955.0
63.00	9990.0
64.10	1456.0
65.00	5118.0
66.00	442.0
68.00	3332.0
69.00	243136.0
70.00	1326.0
70.90	321.0
73.00	1723.0
74.00	23648.0
75.00	38416.0
76.00	12204.0
77.00	268352.0
78.00	18016.0
79.00	18000.0
80.00	14045.0
81.00	17736.0
82.00	4785.0
83.00	4872.0
84.10	559.0
85.00	3514.0
86.00	4612.0
87.00	2407.0
87.90	826.0
89.00	629.0
91.00	4284.0
92.00	4412.0
93.00	29480.0
94.00	2171.0
95.00	635.0
96.10	1204.0
96.90	451.0
98.00	20720.0
98.90	16816.0
100.00	1634.0
101.00	10123.0
102.00	655.0
103.00	3693.0
104.00	6529.0
104.90	6522.0
106.00	2325.0
107.00	82576.0
108.00	13687.0
108.90	2324.0
110.00	152384.0
111.00	23592.0
112.00	2387.0
113.00	999.0
116.00	4915.0
117.00	56488.0
118.00	3989.0
119.00	748.0
120.00	1300.0
120.80	335.0
121.90	4997.0
123.00	7882.0
123.90	3558.0
125.00	3766.0
126.20	1698.0
127.00	295104.0
128.00	23440.0
129.00	110560.0
130.00	9017.0
130.90	1989.0
131.90	1435.0
133.10	553.0
133.90	2889.0
134.90	8901.0
136.00	3223.0
137.00	4390.0
137.90	1207.0
138.90	555.0
139.90	1636.0

Instrument :
BNA_N
ClientSampleId :
DFTPP95

141.00	12675.0
141.90	4372.0
143.00	3376.0
144.00	1101.0
144.90	953.0
145.90	2389.0
147.00	7239.0
148.00	14104.0
149.00	3975.0
150.00	948.0
151.10	2228.0
151.80	1197.0
152.90	5007.0
154.00	3110.0
155.00	8289.0
156.00	11312.0
157.00	2817.0
157.90	2483.0
158.90	2012.0
160.00	5094.0
161.00	6675.0
162.00	1935.0
163.00	631.0
163.90	759.0
164.90	5382.0
166.00	4333.0
167.00	26920.0
168.00	13539.0
169.00	2422.0
170.00	1240.0
170.90	895.0
172.00	3215.0
173.00	3261.0
174.00	6485.0
175.00	11843.0
175.90	3706.0
177.00	5830.0
178.00	2227.0
178.90	22144.0
180.00	14865.0
181.00	7174.0
182.00	1352.0
182.80	369.0
183.90	1734.0
185.00	11125.0
186.00	82520.0
187.00	24536.0
188.00	2815.0
188.90	4703.0
190.00	878.0
191.00	2667.0
192.00	6345.0
193.00	8007.0
194.00	1183.0
194.90	1069.0
196.00	17464.0
197.00	3083.0
197.90	621120.0
198.90	42832.0
200.00	3375.0
201.40	2889.0
203.00	4421.0
204.00	21224.0
205.00	36152.0
206.00	158784.0
207.00	20736.0
208.00	5286.0
208.90	1677.0
210.00	2407.0
210.90	7296.0
211.80	1312.0
213.00	427.0
214.90	1709.0
216.00	2962.0
217.00	42088.0
218.00	5252.0
219.00	564.0
220.10	1057.0
221.00	26584.0
221.80	6484.0
223.00	9027.0
224.00	85128.0
225.00	22888.0
226.00	2445.0
227.00	34456.0
228.00	4930.0
229.00	6943.0
229.90	1022.0
231.00	2815.0
231.90	620.0
232.90	911.0
234.00	2451.0
235.00	2745.0
236.00	1841.0
237.00	3001.0

Instrument :
BNA_N
ClientSampleId :
DFTPP95

237.90	455.0
239.00	1469.0
239.90	1023.0
241.00	1697.0
242.00	4252.0
243.10	4690.0
244.00	66264.0
245.00	9307.0
246.00	13436.0
247.00	2549.0
248.00	621.0
249.00	2223.0
249.90	672.0
251.10	842.0
252.00	842.0
253.00	1382.0
255.00	361088.0
256.00	50456.0
257.00	3754.0
258.00	19976.0
259.00	2712.0
259.80	533.0
261.00	813.0
264.10	370.0
265.00	7495.0
265.90	1181.0
269.90	608.0
271.00	441.0
271.90	815.0
272.90	10087.0
274.00	28416.0
275.00	162880.0
276.00	20720.0
277.00	13125.0
277.90	1967.0
279.00	570.0
281.00	560.0
281.70	341.0
282.90	1386.0
283.90	1257.0
285.00	2574.0
286.00	525.0
289.10	455.0
289.90	572.0
290.90	415.0
291.90	675.0
293.00	3269.0
294.00	1095.0
295.10	1239.0
296.00	41296.0
297.00	5340.0
301.10	315.0
302.00	997.0
303.00	5282.0
304.00	1213.0
307.80	634.0
309.00	382.0
310.00	794.0
313.00	474.0
313.90	2013.0
315.00	5036.0
315.90	2484.0
317.00	601.0
321.00	1382.0
322.00	886.0
323.00	14586.0
324.00	2672.0
326.00	355.0
326.90	2562.0
328.00	1405.0
328.80	303.0
331.90	930.0
332.90	1484.0
334.00	8614.0
335.00	2564.0
339.10	437.0
341.00	1533.0
341.90	384.0
346.00	2975.0
346.90	383.0
352.00	3911.0
353.00	2286.0
354.00	4421.0
355.00	803.0
359.00	401.0
364.90	17464.0
365.90	2857.0
369.90	382.0
371.00	1609.0
372.00	7390.0
373.10	1795.0
383.00	1678.0
384.00	795.0
384.90	327.0
389.90	653.0

Instrument :
BNA_N
ClientSampleId :
DFTPP95

390.90	400.0
391.20	495.0
392.00	679.0
400.90	373.0
401.90	2789.0
403.00	4512.0
404.00	1741.0
404.90	337.0
421.00	3712.0
422.00	3929.0
423.00	27904.0
424.00	6715.0
424.90	898.0
441.00	75760.0
442.00	538688.0
443.00	101936.0
444.00	10691.0
444.90	650.0

Instrument :
BNA_N
ClientSampleId :
DFTPP95

m/z	Abundance
37.10	666.0
38.10	1731.0
39.00	13611.0
39.90	1526.0
41.00	698.0
44.00	4820.0
45.10	411.0
49.10	1157.0
50.00	46744.0
51.00	186944.0
52.00	10507.0
53.00	602.0
55.00	1054.0
56.00	5769.0
57.00	13454.0
57.90	586.0
61.00	2585.0
62.00	2695.0
63.00	7681.0
64.00	1211.0
65.00	3847.0
68.00	3215.0
69.00	192512.0
70.00	1365.0
73.00	1585.0
74.00	19504.0
75.00	30088.0
76.00	10183.0
77.00	211904.0
78.00	14882.0
79.00	15739.0
80.00	11042.0
81.00	15318.0
81.90	3603.0
83.00	3966.0
83.90	325.0
85.00	3018.0
86.00	4609.0
87.00	1917.0
87.90	638.0
88.90	439.0
91.00	3842.0
92.00	4018.0
93.00	24080.0
94.00	1665.0
94.80	493.0
96.00	1117.0
97.00	529.0
98.00	17800.0
99.00	14175.0
100.00	1359.0
101.00	8816.0
101.90	421.0
103.00	3113.0
104.00	5295.0
104.90	5528.0
106.00	1690.0
107.00	68408.0
108.00	11660.0
109.00	1892.0
110.00	129712.0
111.00	19648.0
111.90	2119.0
113.00	770.0
114.80	361.0
116.00	3911.0
117.00	47816.0
117.90	3548.0
119.00	601.0
120.00	915.0
121.10	549.0
122.00	4298.0
123.00	7025.0
123.90	3181.0
125.00	2715.0
126.10	1417.0
127.00	252224.0
128.00	19840.0
129.00	96800.0
130.00	7895.0
130.90	1707.0
132.00	842.0
133.00	403.0
133.90	2907.0
135.00	7277.0
136.00	2966.0
137.00	3723.0
138.00	849.0
138.90	481.0
140.10	1406.0
141.00	14138.0

Instrument :
BNA_N
ClientSampleId :
DFTPP95

142.00	3512.0
142.90	3410.0
144.00	841.0
145.00	820.0
146.00	2270.0
146.90	6163.0
147.90	14263.0
149.00	2941.0
149.90	883.0
151.00	1577.0
151.80	997.0
152.90	3640.0
153.90	2882.0
155.00	6882.0
156.00	11762.0
157.00	1868.0
157.90	2304.0
159.00	1856.0
160.00	4096.0
160.90	5846.0
162.00	1662.0
163.10	405.0
163.90	381.0
164.20	384.0
164.90	5271.0
166.00	4158.0
167.00	22984.0
168.00	11510.0
169.00	1775.0
170.00	1053.0
170.90	1125.0
172.00	2080.0
173.10	2572.0
174.00	5553.0
175.00	9656.0
176.00	3415.0
176.90	5409.0
177.90	1424.0
179.00	19080.0
180.00	13725.0
180.90	6735.0
182.00	962.0
183.00	623.0
183.90	1897.0
185.00	10023.0
186.00	75560.0
187.00	22344.0
188.00	1976.0
189.00	4809.0
190.10	894.0
191.00	2282.0
191.90	6065.0
193.00	7063.0
194.00	1533.0
195.00	1164.0
196.00	15559.0
196.80	3078.0
197.90	579456.0
198.90	39592.0
200.00	3372.0
201.50	2728.0
202.90	3621.0
204.00	19792.0
205.00	33376.0
206.00	152512.0
207.00	21504.0
208.00	4727.0
208.90	1500.0
210.00	2306.0
210.90	5747.0
211.80	1317.0
212.90	454.0
214.90	1421.0
216.00	3250.0
216.90	40600.0
218.00	5360.0
218.90	520.0
221.00	24464.0
221.80	6568.0
223.00	8220.0
224.00	83616.0
225.00	22376.0
226.00	2198.0
227.00	31168.0
228.00	5563.0
229.00	7821.0
229.90	1021.0
231.00	2366.0
231.90	599.0
233.00	563.0
233.90	2438.0
235.00	2627.0
235.90	1524.0
237.00	3037.0
237.90	423.0

Instrument :
BNA_N
ClientSampleId :
DFTPP95

238.90	1495.0
240.00	1403.0
241.00	1380.0
241.90	4707.0
243.00	5001.0
244.00	64904.0
245.00	8082.0
246.00	13289.0
247.00	2311.0
247.90	730.0
248.90	2726.0
250.00	406.0
251.00	641.0
252.00	783.0
252.90	1413.0
255.00	354304.0
256.00	51512.0
257.00	3620.0
258.00	18952.0
259.00	2844.0
259.90	743.0
260.90	574.0
263.90	601.0
265.00	8224.0
265.90	942.0
270.10	323.0
271.00	824.0
271.80	666.0
272.20	752.0
273.00	10594.0
274.00	27760.0
275.00	162816.0
276.00	20768.0
277.00	12637.0
277.90	2024.0
279.00	658.0
280.90	504.0
281.90	382.0
283.00	1642.0
284.00	1014.0
284.90	2304.0
285.90	463.0
289.00	477.0
290.00	484.0
290.90	433.0
291.90	619.0
293.00	2936.0
294.00	676.0
295.10	1080.0
296.00	45352.0
297.00	6594.0
298.00	360.0
300.80	860.0
302.00	762.0
303.00	4674.0
304.00	1357.0
307.90	723.0
308.80	396.0
309.90	362.0
314.00	2015.0
315.00	5389.0
316.00	2899.0
316.80	346.0
320.90	1122.0
322.00	855.0
323.00	15485.0
324.00	3107.0
327.00	2333.0
327.90	1642.0
329.00	312.0
332.00	1022.0
333.00	1720.0
334.00	9666.0
335.00	2524.0
336.00	428.0
338.90	323.0
341.00	1714.0
346.00	2947.0
347.00	544.0
352.00	4458.0
353.00	3519.0
354.00	4886.0
355.00	909.0
358.90	576.0
363.90	302.0
364.90	20296.0
365.90	2683.0
371.10	1805.0
372.00	8402.0
373.00	1803.0
382.90	2373.0
384.00	630.0
390.10	1344.0
390.90	852.0
392.00	721.0

Instrument :
BNA_N
ClientSampleId :
DFTPP95

400.90	424.0
402.00	3520.0
403.00	5255.0
404.00	1673.0
405.00	340.0
421.00	5095.0
422.00	3932.0
423.00	32680.0
424.00	6623.0
424.90	901.0
441.00	98968.0
442.00	650240.0
443.00	126880.0
444.00	11575.0
445.10	760.0

Instrument :
BNA_N
ClientSampleId :
DFTPP95

Scan 2465 (17.486 min): BN002127.D
DFTPP95

m/z	Abundance
40.00	945.0
44.00	4517.0
206.90	1023.0

Instrument :
BNA_N
ClientSampleId :
DFTPP95