

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN072320\
 Data File : BN011224.D
 Acq On : 22 Jul 2020 18:57
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 23 03:55:41 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 22 19:29:31 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
2	1,4-Dioxane	0.572	0.493	13.8	70	0.00
3	n-Nitrosodimethylamine	0.288	0.263	8.7	84	0.00
4 S	2-Fluorophenol	0.986	0.903	8.4	77	0.00
5 S	Phenol-d6	1.115	1.174	-5.3	89	0.00
6	bis(2-Chloroethyl)ether	1.021	0.941	7.8	86	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	98	0.00
8 S	Nitrobenzene-d5	0.355	0.329	7.3	90	0.01
9	Naphthalene	1.190	1.122	5.7	93	0.00
10	Hexachlorobutadiene	0.295	0.279	5.4	89	0.00
11 SURR	2-Methylnaphthalene-d10	0.649	0.706	-8.8	101	0.00
12	2-Methylnaphthalene	0.749	0.760	-1.5	96	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
14 S	2,4,6-Tribromophenol	0.178	0.168	5.6	107	0.00
15 S	2-Fluorobiphenyl	1.874	1.659	11.5	96	0.00
16	Acenaphthylene	2.036	1.832	10.0	108	0.00
17	Acenaphthene	1.358	1.240	8.7	106	0.00
18	Fluorene	1.742	1.637	6.0	96	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	129	0.00
20	4,6-Dinitro-2-methylphenol	0.053	0.045	15.1	108	0.00
21	4-Bromophenyl-phenylether	0.272	0.246	9.6	116	0.00
22	Hexachlorobenzene	0.316	0.269	14.9	108	0.00
23	Atrazine	0.191	0.185	3.1	125	-0.01
24	Pentachlorophenol	0.094	0.083	11.7	122	0.00
25	Phenanthrene	1.285	1.202	6.5	121	0.00
26	Anthracene	1.067	0.961	9.9	117	0.00
27 SURR	Fluoranthene-d10	1.261	1.179	6.5	119	0.00
28	Fluoranthene	1.541	1.395	9.5	112	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	106	-0.01
30	Pyrene	1.527	1.751	-14.7	107	0.00
31 S	Terphenyl-d14	1.041	1.162	-11.6	112	0.00
32	Benzo(a)anthracene	1.296	1.215	6.2	102	0.00
33	Chrysene	1.584	1.582	0.1	106	-0.01
34	Bis(2-ethylhexyl)phthalate	0.534	0.643	-20.4	129	0.00
35	Indeno(1,2,3-cd)pyrene	1.656	1.457	12.0	80	-0.01
36 I	Perylene-d12	1.000	1.000	0.0	98	-0.01
37	Benzo(b)fluoranthene	1.607	1.435	10.7	97	0.00
38	Benzo(k)fluoranthene	1.816	1.770	2.5	95	-0.01
39 C	Benzo(a)pyrene	1.447	1.257	13.1	84	-0.01
40	Dibenzo(a,h)anthracene	1.533	1.198	21.9	75	-0.02
41	Benzo(g,h,i)perylene	1.683	1.386	17.6	78	-0.01

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN072320\
Data File : BN011224.D
Acq On : 22 Jul 2020 18:57
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 23 03:55:41 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071120.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 22 19:29:31 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)
----------	-------	------	-----------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0