

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN072320\
 Data File : BN011232.D
 Acq On : 22 Jul 2020 23:48
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 23 03:58:42 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071120.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Jul 22 19:29:31 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	86	0.00
2	1,4-Dioxane	0.400	0.350	12.5	70	0.00
3	n-Nitrosodimethylamine	0.400	0.354	11.5	80	0.00
4 S	2-Fluorophenol	0.400	0.369	7.8	76	0.00
5 S	Phenol-d6	0.400	0.404	-1.0	85	0.00
6	bis(2-Chloroethyl)ether	0.400	0.357	10.8	83	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	95	0.00
8 S	Nitrobenzene-d5	0.400	0.373	6.8	88	0.00
9	Naphthalene	0.400	0.382	4.5	92	0.00
10	Hexachlorobutadiene	0.400	0.369	7.8	85	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.434	-8.5	99	0.00
12	2-Methylnaphthalene	0.400	0.417	-4.2	96	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	115	0.00
14 S	2,4,6-Tribromophenol	0.400	0.401	-0.3	112	0.00
15 S	2-Fluorobiphenyl	0.400	0.354	11.5	95	0.00
16	Acenaphthylene	0.400	0.365	8.8	108	0.00
17	Acenaphthene	0.400	0.374	6.5	107	0.00
18	Fluorene	0.400	0.394	1.5	100	-0.01
19 I	Phenanthrene-d10	0.400	0.400	0.0	140	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.374	6.5	117	0.00
21	4-Bromophenyl-phenylether	0.400	0.343	14.2	120	0.00
22	Hexachlorobenzene	0.400	0.314	21.5	109	-0.01
23	Atrazine	0.400	0.379	5.3	132	-0.01
24	Pentachlorophenol	0.400	0.333	16.8	126	0.00
25	Phenanthrene	0.400	0.374	6.5	131	0.00
26	Anthracene	0.400	0.363	9.3	128	0.00
27 SURR	Fluoranthene-d10	0.400	0.383	4.3	132	0.00
28	Fluoranthene	0.400	0.374	6.5	125	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	122	-0.01
30	Pyrene	0.400	0.442	-10.5	119	0.00
31 S	Terphenyl-d14	0.400	0.444	-11.0	128	0.00
32	Benzo(a)anthracene	0.400	0.396	1.0	124	-0.01
33	Chrysene	0.400	0.389	2.8	119	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.469	-17.2	145	-0.01
35	Indeno(1,2,3-cd)pyrene	0.400	0.339	15.3	91	-0.02
36 I	Perylene-d12	0.400	0.400	0.0	114	-0.01
37	Benzo(b)fluoranthene	0.400	0.360	10.0	113	-0.01
38	Benzo(k)fluoranthene	0.400	0.383	4.3	108	-0.02
39 C	Benzo(a)pyrene	0.400	0.340	15.0	95	-0.02
40	Dibenzo(a,h)anthracene	0.400	0.317	20.8	88	-0.02
41	Benzo(g,h,i)perylene	0.400	0.319	20.3	88	-0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN072320\
Data File : BN011232.D
Acq On : 22 Jul 2020 23:48
Operator : CG/JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 23 03:58:42 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN071120.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Jul 22 19:29:31 2020
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area	% Dev(min)
----------	--------	-------	------	------	------------

(#) = Out of Range

SPCC's out = 0 CCC's out = 0