

DDT Breakdown

Date	Instrument Name	DFTPP Data File
7/24/2023	BNA_N	<u>BN026742.D</u>
Compound Name	Response	Retention Time
DDT	149430	20.692
DDD	119956	20.251
DDE	4352	19.739
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
124308	273738	45.41