

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072423\
 Data File : BN026727.D
 Acq On : 25 Jul 2023 00:35
 Operator : MA/JU
 Sample : PB154197BS
 Misc :
 ALS Vial : 63 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB154197BS

Quant Time: Jul 25 03:26:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN062723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 24 23:51:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.835	152	1502	0.400	ng	0.00
7) Naphthalene-d8	10.628	136	4641	0.400	ng	# 0.00
13) Acenaphthene-d10	14.458	164	3447	0.400	ng	0.00
19) Phenanthrene-d10	17.219	188	7304	0.400	ng	# 0.00
29) Chrysene-d12	21.417	240	5183	0.400	ng	# 0.00
35) Perylene-d12	23.779	264	6500	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.422	112	972	0.267	ng	0.00
5) Phenol-d6	7.047	99	1174	0.253	ng	0.00
8) Nitrobenzene-d5	9.037	82	1088	0.274	ng	0.00
11) 2-Methylnaphthalene-d10	12.233	152	4034	0.594	ng	0.00
14) 2,4,6-Tribromophenol	15.978	330	500	0.399	ng	0.00
15) 2-Fluorobiphenyl	13.101	172	4387	0.330	ng	0.00
27) Fluoranthene-d10	19.253	212	6448	0.417	ng	0.00
31) Terphenyl-d14	19.848	244	4930	0.372	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.227	88	497	0.313	ng	# 6
3) n-Nitrosodimethylamine	3.631	42	685	0.426	ng	# 62
6) bis(2-Chloroethyl)ether	7.307	93	606	0.158	ng	# 30
9) Naphthalene	10.682	128	4648	0.370	ng	# 94
10) Hexachlorobutadiene	10.938	225	1136	0.468	ng	# 99
12) 2-Methylnaphthalene	12.305	142	3277	0.371	ng	# 92
16) Acenaphthylene	14.191	152	6108	0.378	ng	100
17) Acenaphthene	14.523	154	3820	0.366	ng	98
18) Fluorene	15.519	166	4683	0.344	ng	98
20) 4,6-Dinitro-2-methylph...	16.127	198	10	0.067	ng	# 1
21) 4-Bromophenyl-phenylether	16.412	248	1538	0.369	ng	# 90
22) Hexachlorobenzene	16.524	284	1828	0.437	ng	96
23) Atrazine	16.685	200	1670	0.466	ng	92
24) Pentachlorophenol	16.896	266	1426	0.725	ng	98
25) Phenanthrene	17.256	178	7163	0.335	ng	99
26) Anthracene	17.356	178	6652	0.345	ng	99
28) Fluoranthene	19.281	202	8228	0.366	ng	# 95
30) Pyrene	19.643	202	8463	0.295	ng	99
32) Benzo(a)anthracene	21.399	228	6730	0.361	ng	98
33) Chrysene	21.443	228	7385	0.355	ng	97
34) Bis(2-ethylhexyl)phtha...	21.300	149	6023	0.362	ng	# 95
36) Indeno(1,2,3-cd)pyrene	26.241	276	10211	0.383	ng	# 86
37) Benzo(b)fluoranthene	23.060	252	7433	0.315	ng	# 90
38) Benzo(k)fluoranthene	23.107	252	7857	0.308	ng	# 88
39) Benzo(a)pyrene	23.680	252	7508	0.324	ng	# 87
40) Dibenzo(a,h)anthracene	26.252	278	8220	0.394	ng	# 90
41) Benzo(g,h,i)perylene	26.992	276	9239	0.379	ng	# 87


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072423\
Data File : BN026727.D
Acq On : 25 Jul 2023 00:35
Operator : MA/JU
Sample : PB154197BS
Misc :
ALS Vial : 63 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB154197BS

Quant Time: Jul 25 03:26:23 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN062723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 24 23:51:48 2023
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.835 min Scan# 61
 Delta R.T. -0.000 min
 Lab File: BN026727.D
 Acq: 25 Jul 2023 00:35




Instrument :
 BNA_N
 ClientSampleId :
 PB154197BS


Tgt Ion:152 Resp: 1502
 Ion Ratio Lower Upper
 152 100
 150 144.2 124.7 187.1
 115 70.5 48.2 72.4



#2
 1,4-Dioxane
 Concen: 0.313 ng
 RT: 3.227 min Scan# 48
 Delta R.T. -0.007 min
 Lab File: BN026727.D
 Acq: 25 Jul 2023 00:35

Tgt Ion: 88 Resp: 497
 Ion Ratio Lower Upper
 88 100
 43 153.1 24.2 36.4#
 58 93.8 53.3 79.9#









#3
 n-Nitrosodimethylamine
 Concen: 0.426 ng
 RT: 3.631 min Scan# 104
 Delta R.T. -0.000 min
 Lab File: BN026727.D
 Acq: 25 Jul 2023 00:35




Instrument :
 BNA_N
 ClientSampleId :
 PB154197BS


Tgt Ion: 42 Resp: 685
 Ion Ratio Lower Upper
 42 100
 74 74.3 93.8 140.8#
 44 0.0 3.0 4.4#



#4
 2-Fluorophenol
 Concen: 0.267 ng
 RT: 5.422 min Scan# 352
 Delta R.T. 0.007 min
 Lab File: BN026727.D
 Acq: 25 Jul 2023 00:35

Tgt Ion: 112 Resp: 972
 Ion Ratio Lower Upper
 112 100
 64 53.9 45.8 68.6
 63 41.2 27.4 41.2









#5
Phenol-d6
Concen: 0.253 ng
RT: 7.047 min Scan# 511
Delta R.T. 0.007 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35




Instrument :
BNA_N
ClientSampleId :
PB154197BS


Tgt Ion: 99 Resp: 1174
Ion Ratio Lower Upper
99 100
42 31.9 18.4 27.6#
71 43.9 30.6 46.0



#6
bis(2-Chloroethyl)ether
Concen: 0.158 ng
RT: 7.307 min Scan# 613
Delta R.T. 0.014 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35

Tgt Ion: 93 Resp: 606
Ion Ratio Lower Upper
93 100
63 0.0 58.8 88.2#
95 16.3 29.2 43.8#





#7

Naphthalene-d8

Concen: 0.400 ng

RT: 10.628 min Scan# 9

Delta R.T. -0.000 min

Lab File: BN026727.D

Acq: 25 Jul 2023 00:35

Instrument :

BNA_N

ClientSampleId :

PB154197BS

Tgt Ion:136 Resp: 4641





Ion Ratio Lower Upper

136 100

137 13.6 9.0 13.4#

54 11.7 7.8 11.6#

68 7.4 5.8 8.8



#8

Nitrobenzene-d5

Concen: 0.274 ng

RT: 9.037 min Scan# 820

Delta R.T. -0.000 min

Lab File: BN026727.D

Acq: 25 Jul 2023 00:35




Tgt Ion: 82 Resp: 1088


Ion Ratio Lower Upper

82 100

128 54.9 33.4 50.2#

54 70.5 41.7 62.5#









#9
Naphthalene
Concen: 0.370 ng
RT: 10.682 min Scan# 91
Delta R.T. -0.000 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35




Instrument :
BNA_N
ClientSampleId :
PB154197BS


Tgt Ion:128 Resp: 4648
Ion Ratio Lower Upper
128 100
129 14.0 9.1 13.7#
127 15.8 10.9 16.3



#10
Hexachlorobutadiene
Concen: 0.468 ng
RT: 10.938 min Scan# 998
Delta R.T. -0.000 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35

Tgt Ion:225 Resp: 1136
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.3 50.3 75.5









#11
2-Methylnaphthalene-d10
Concen: 0.594 ng
RT: 12.233 min Scan# 11
Delta R.T. 0.004 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35




Instrument :
BNA_N
ClientSampleId :
PB154197BS


Tgt Ion:152 Resp: 4034
Ion Ratio Lower Upper
152 100
151 13.4 16.2 24.2#



#12
2-Methylnaphthalene
Concen: 0.371 ng
RT: 12.305 min Scan# 1263
Delta R.T. 0.004 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35

Tgt Ion:142 Resp: 3277
Ion Ratio Lower Upper
142 100
141 87.9 71.7 107.5
115 54.5 33.4 50.0#









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.458 min Scan# 1720
Delta R.T. -0.000 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35




Instrument :
BNA_N
ClientSampleId :
PB154197BS


Tgt Ion:164 Resp: 3447
Ion Ratio Lower Upper
164 100
162 107.5 85.3 127.9
160 51.7 40.0 60.0



#14
2,4,6-Tribromophenol
Concen: 0.399 ng
RT: 15.978 min Scan# 1720
Delta R.T. -0.000 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35

Tgt Ion:330 Resp: 500
Ion Ratio Lower Upper
330 100
332 98.6 76.0 114.0
141 52.6 42.8 64.2





#15

2-Fluorobiphenyl

Concen: 0.330 ng

RT: 13.101 min Scan# 14

Delta R.T. -0.000 min

Lab File: BN026727.D

Acq: 25 Jul 2023 00:35

Instrument :

BNA_N

ClientSampleId :

PB154197BS





Tgt Ion:172 Resp: 4387

Ion Ratio Lower Upper

172 100

171 39.1 27.8 41.8

170 26.5 19.3 28.9



#16

Acenaphthylene

Concen: 0.378 ng

RT: 14.191 min Scan# 1559

Delta R.T. -0.000 min

Lab File: BN026727.D

Acq: 25 Jul 2023 00:35




Tgt Ion:152 Resp: 6108


Ion Ratio Lower Upper

152 100

151 19.9 15.9 23.9

153 13.2 10.6 15.8









#17
Acenaphthene
Concen: 0.366 ng
RT: 14.523 min Scan# 1590
Delta R.T. -0.000 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35




Instrument :
BNA_N
ClientSampleId :
PB154197BS


Tgt Ion:154 Resp: 3820
Ion Ratio Lower Upper
154 100
153 115.6 94.4 141.6
152 55.4 46.1 69.1



#18
Fluorene
Concen: 0.344 ng
RT: 15.519 min Scan# 1683
Delta R.T. -0.000 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35

Tgt Ion:166 Resp: 4683
Ion Ratio Lower Upper
166 100
165 101.1 79.7 119.5
167 14.0 10.6 16.0





#19

Phenanthrene-d10

Concen: 0.400 ng

RT: 17.219 min Scan# 1732

Delta R.T. -0.000 min

Lab File: BN026727.D

Acq: 25 Jul 2023 00:35

Instrument :

BNA_N

ClientSampleId :

PB154197BS





Tgt Ion:188 Resp: 7304

Ion Ratio Lower Upper

188 100

94 0.0 0.0 0.0

80 6.3 8.6 13.0#



#20

4,6-Dinitro-2-methylphenol

Concen: 0.067 ng

RT: 16.127 min Scan# 1732

Delta R.T. -0.025 min

Lab File: BN026727.D

Acq: 25 Jul 2023 00:35




Tgt Ion:198 Resp: 10


Ion Ratio Lower Upper

198 100

51 215.7 85.8 128.8#

105 164.7 48.2 72.2#









#21
4-Bromophenyl-phenylether
Concen: 0.369 ng
RT: 16.412 min Scan# 1755
Delta R.T. -0.000 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35




Instrument :
BNA_N
ClientSampleId :
PB154197BS


Tgt Ion:248 Resp: 1538
Ion Ratio Lower Upper
248 100
250 100.1 79.9 119.9
141 72.5 41.6 62.4#



#22
Hexachlorobenzene
Concen: 0.437 ng
RT: 16.524 min Scan# 1764
Delta R.T. -0.000 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35

Tgt Ion:284 Resp: 1828
Ion Ratio Lower Upper
284 100
142 46.2 38.7 58.1
249 32.7 28.5 42.7





#23

Atrazine

Concen: 0.466 ng

RT: 16.685 min Scan# 11

Delta R.T. -0.000 min

Lab File: BN026727.D

Acq: 25 Jul 2023 00:35

Instrument :

BNA_N

ClientSampleId :

PB154197BS





Tgt Ion:200 Resp: 1670

Ion Ratio Lower Upper

200 100

173 30.7 22.9 34.3

215 51.8 35.5 53.3



#24

Pentachlorophenol

Concen: 0.725 ng

RT: 16.896 min Scan# 1794

Delta R.T. -0.000 min

Lab File: BN026727.D

Acq: 25 Jul 2023 00:35




Tgt Ion:266 Resp: 1426


Ion Ratio Lower Upper

266 100

264 61.3 49.3 73.9

268 63.1 48.3 72.5









#25
 Phenanthrene
 Concen: 0.335 ng
 RT: 17.256 min Scan# 1823
 Delta R.T. -0.000 min
 Lab File: BN026727.D
 Acq: 25 Jul 2023 00:35




Instrument :
 BNA_N
 ClientSampleId :
 PB154197BS


Tgt Ion:	178	Resp:	7163
Ion	Ratio	Lower	Upper
178	100		
176	20.1	15.6	23.4
179	15.1	12.3	18.5



#26
 Anthracene
 Concen: 0.345 ng
 RT: 17.356 min Scan# 1831
 Delta R.T. -0.000 min
 Lab File: BN026727.D
 Acq: 25 Jul 2023 00:35

Tgt Ion:	178	Resp:	6652
Ion	Ratio	Lower	Upper
178	100		
176	18.6	15.0	22.4
179	14.7	12.4	18.6









#27
Fluoranthene-d10
Concen: 0.417 ng
RT: 19.253 min Scan# 2084
Delta R.T. 0.004 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35




Instrument :
BNA_N
ClientSampleId :
PB154197BS


Tgt Ion:212 Resp: 6448
Ion Ratio Lower Upper
212 100
106 6.5 11.6 17.4#
104 4.2 7.2 10.8#



#28
Fluoranthene
Concen: 0.366 ng
RT: 19.281 min Scan# 2090
Delta R.T. -0.000 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35

Tgt Ion:202 Resp: 8228
Ion Ratio Lower Upper
202 100
101 5.5 8.3 12.5#
203 17.2 13.5 20.3









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.417 min Scan# 2168
Delta R.T. 0.009 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35




Instrument :
BNA_N
ClientSampleId :
PB154197BS


Tgt Ion:240 Resp: 5183
Ion Ratio Lower Upper
240 100
120 12.4 14.6 22.0#
236 31.2 24.2 36.2



#30
Pyrene
Concen: 0.295 ng
RT: 19.643 min Scan# 2168
Delta R.T. -0.000 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35

Tgt Ion:202 Resp: 8463
Ion Ratio Lower Upper
202 100
200 20.9 17.0 25.4
203 18.0 14.2 21.4









#31
 Terphenyl-d14
 Concen: 0.372 ng
 RT: 19.848 min Scan# 2198
 Delta R.T. -0.000 min
 Lab File: BN026727.D
 Acq: 25 Jul 2023 00:35




Instrument :
 BNA_N
 ClientSampleId :
 PB154197BS


Tgt Ion:244 Resp: 4930
 Ion Ratio Lower Upper
 244 100
 212 12.3 8.7 13.1
 122 10.7 13.7 20.5#



#32
 Benzo(a)anthracene
 Concen: 0.361 ng
 RT: 21.399 min Scan# 2505
 Delta R.T. -0.000 min
 Lab File: BN026727.D
 Acq: 25 Jul 2023 00:35

Tgt Ion:228 Resp: 6730
 Ion Ratio Lower Upper
 228 100
 226 29.2 23.0 34.6
 229 21.6 15.9 23.9









#33
Chrysene
Concen: 0.355 ng
RT: 21.443 min Scan# 2149
Delta R.T. -0.000 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35




Instrument :
BNA_N
ClientSampleId :
PB154197BS


Tgt Ion:228 Resp: 7385
Ion Ratio Lower Upper
228 100
226 32.8 24.8 37.2
229 21.2 15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.362 ng
RT: 21.300 min Scan# 2494
Delta R.T. -0.000 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35

Tgt Ion:149 Resp: 6023
Ion Ratio Lower Upper
149 100
167 27.5 20.2 30.4
279 5.0 2.4 3.6





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.779 min Scan# 3086

Delta R.T. -0.003 min

Lab File: BN026727.D

Acq: 25 Jul 2023 00:35

Instrument :

BNA_N

ClientSampleId :

PB154197BS





Tgt Ion:264 Resp: 6500

Ion Ratio Lower Upper

264 100

260 27.0 21.9 32.9

265 96.2 80.3 120.5



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.383 ng

RT: 26.241 min Scan# 3928

Delta R.T. -0.000 min

Lab File: BN026727.D

Acq: 25 Jul 2023 00:35




Tgt Ion:276 Resp: 10211


Ion Ratio Lower Upper

276 100

138 13.6 21.4 32.2#


277 24.0 19.6 29.4








#37
Benzo(b)fluoranthene
Concen: 0.315 ng
RT: 23.060 min Scan# 2840
Delta R.T. -0.000 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35

Instrument :
BNA_N
ClientSampleId :
PB154197BS







Tgt Ion:252 Resp: 7433
Ion Ratio Lower Upper
252 100
253 34.5 24.1 36.1
125 12.8 15.4 23.0#



#38
Benzo(k)fluoranthene
Concen: 0.308 ng
RT: 23.107 min Scan# 2856
Delta R.T. -0.000 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35

Tgt Ion:252 Resp: 7857
Ion Ratio Lower Upper
252 100
253 34.8 23.6 35.4
125 12.9 15.8 23.8#





#39

Benzo(a)pyrene

Concen: 0.324 ng

RT: 23.680 min Scan# 3089

Delta R.T. -0.000 min

Lab File: BN026727.D

Acq: 25 Jul 2023 00:35

Instrument :

BNA_N

ClientSampleId :

PB154197BS





Tgt Ion:252 Resp: 7508

Ion Ratio Lower Upper

252 100

253 39.1 26.2 39.4

125 16.1 18.7 28.1#



#40

Dibenzo(a,h)anthracene

Concen: 0.394 ng

RT: 26.252 min Scan# 3932

Delta R.T. -0.000 min

Lab File: BN026727.D

Acq: 25 Jul 2023 00:35




Tgt Ion:278 Resp: 8220


Ion Ratio Lower Upper

278 100

139 16.5 18.1 27.1#


279 36.3 25.4 38.0





#41
Benzo(g,h,i)perylene
Concen: 0.379 ng
RT: 26.992 min Scan# 41
Delta R.T. -0.000 min
Lab File: BN026727.D
Acq: 25 Jul 2023 00:35

Instrument :
BNA_N
ClientSampleId :
PB154197BS



Tgt Ion:	276	Resp:	9239
Ion	Ratio	Lower	Upper
276	100		
277	29.2	20.3	30.5
138	17.3	21.5	32.3

