

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072423\
 Data File : BN026739.D
 Acq On : 25 Jul 2023 07:54
 Operator : MA/JU
 Sample : 03606-03
 Misc :
 ALS Vial : 75 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GWBR-04-293-308-071223-FD

Quant Time: Jul 25 08:26:35 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN062723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 24 23:51:48 2023
 Response via : Initial Calibration

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/25/2023
 Supervised By :mohammad ahmed 07/26/2023

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.820	152	1904	0.400	ng	-0.01
7) Naphthalene-d8	10.618	136	6365	0.400	ng	#-0.01
13) Acenaphthene-d10	14.459	164	4550	0.400	ng	0.00
19) Phenanthrene-d10	17.207	188	9326	0.400	ng	#-0.01
29) Chrysene-d12	21.408	240	7033	0.400	ng	# 0.00
35) Perylene-d12	23.782	264	7959	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.408	112	741	0.160	ng	0.00
5) Phenol-d6	7.026	99	503	0.085	ng	-0.01
8) Nitrobenzene-d5	9.016	82	1165m	0.214	ng	-0.02
11) 2-Methylnaphthalene-d10	12.215	152	3241	0.348	ng	-0.02
14) 2,4,6-Tribromophenol	15.966	330	722	0.436	ng	-0.01
15) 2-Fluorobiphenyl	13.080	172	5595	0.319	ng	-0.02
27) Fluoranthene-d10	19.244	212	8405	0.425	ng	0.00
31) Terphenyl-d14	19.839	244	7338	0.408	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.227	88	7431	3.690	ng	# 88
9) Naphthalene	10.671	128	551	0.032	ng	# 54
25) Phenanthrene	17.256	178	670	0.025	ng	# 94
34) Bis(2-ethylhexyl)phtha...	21.300	149	14228	0.630	ng	# 97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072423\
Data File : BN026739.D
Acq On : 25 Jul 2023 07:54
Operator : MA/JU
Sample : 03606-03
Misc :
ALS Vial : 75 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GWBR-04-293-308-071223-FD

Quant Time: Jul 25 08:26:35 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN062723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 24 23:51:48 2023
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/25/2023
Supervised By :mohammad ahmed 07/26/2023

