

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072423\
 Data File : BN026749.D
 Acq On : 25 Jul 2023 16:52
 Operator : MA/JU
 Sample : PB154354BL
 Misc :
 ALS Vial : 85 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB154354BL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/26/2023
 Supervised By :mohammad ahmed 07/26/2023

Quant Time: Jul 25 17:40:05 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN062723.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 24 23:51:48 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.871	152	1125m	0.400	ng	0.04
7) Naphthalene-d8	10.671	136	3270	0.400	ng	# 0.04
13) Acenaphthene-d10	14.480	164	2908	0.400	ng	0.02
19) Phenanthrene-d10	17.244	188	5105	0.400	ng	# 0.02
29) Chrysene-d12	21.426	240	4060	0.400	ng	0.02
35) Perylene-d12	23.794	264	4709	0.400	ng	0.01
System Monitoring Compounds						
4) 2-Fluorophenol	5.430	112	999	0.366	ng	0.01
5) Phenol-d6	7.156	99	937m	0.269	ng	0.12
8) Nitrobenzene-d5	9.208	82	745m	0.266	ng	0.17
11) 2-Methylnaphthalene-d10	12.279	152	1841	0.385	ng	0.05
14) 2,4,6-Tribromophenol	16.015	330	320	0.302	ng	0.04
15) 2-Fluorobiphenyl	13.133	172	3070m	0.273	ng	0.03
27) Fluoranthene-d10	19.263	212	4923	0.455	ng	0.01
31) Terphenyl-d14	19.853	244	3861	0.372	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072423\
Data File : BN026749.D
Acq On : 25 Jul 2023 16:52
Operator : MA/JU
Sample : PB154354BL
Misc :
ALS Vial : 85 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB154354BL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/26/2023
Supervised By :mohammad ahmed 07/26/2023

Quant Time: Jul 25 17:40:05 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN062723.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 24 23:51:48 2023
Response via : Initial Calibration

