


Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN072518\
 Data File : BN002131.D
 Acq On : 25 Jul 2018 12:17
 Operator : JU/SJ
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-BN072518.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Wed Jul 25 20:08:02 2018



AutoFind: Scans 2470, 2471, 2472; Background Corrected with Scan 2464

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	37.9	91493	PASS
68	69	0.00	2	1.6	1520	PASS
69	198	0.00	100	38.3	92616	PASS
70	69	0.00	2	0.7	606	PASS
127	198	10	80	46.5	112293	PASS
197	198	0.00	2	0.4	880	PASS
198	198	100	100	100.0	241578	PASS
199	198	5	9	6.5	15808	PASS
275	198	10	60	26.2	63336	PASS
365	198	1	100	3.2	7746	PASS
441	198	0.01	100	13.3	32040	PASS
442	198	50	100	90.7	218994	PASS
443	442	15	24	18.9	41333	PASS