



Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072723\
Data File : BN026765.D
Acq On : 27 Jul 2023 13:21
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN062723.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Thu Jul 27 18:29:30 2023






AutoFind: Scans 2505, 2506, 2507; Background Corrected with Scan 2499

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	46.8	72208	PASS
68	69	0.00	2	2.0	1587	PASS
69	198	0.00	100	51.3	79197	PASS
70	69	0.00	2	0.8	618	PASS
127	198	10	80	45.1	69549	PASS
197	198	0.00	2	0.8	1251	PASS
198	198	100	100	100.0	154283	PASS
199	198	5	9	7.1	10958	PASS
275	198	10	60	30.5	47072	PASS
365	198	1	100	7.0	10751	PASS
441	198	0.01	100	13.2	20374	PASS
442	442	50	100	100.0	146432	PASS
443	442	15	24	18.9	27744	PASS

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072723\
 Data File : BN026765.D
 Acq On : 27 Jul 2023 13:21
 Operator : MA/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP

Quant Time: Jul 28 04:59:12 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 28 04:59:07 2023
 Response via : Initial Calibration



TIC: BN026765.D\data.ms

(70) Pentachlorophenol (C)

16.857min (0.000) 14480.09 ng




response 146136

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	60.98
264.00	61.60	61.98
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072723\
Data File : BN026765.D
Acq On : 27 Jul 2023 13:21
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP

Quant Time: Jul 28 04:59:12 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 28 04:59:07 2023
Response via : Initial Calibration



TIC: BN026765.D\data.ms

(77) Benzidine

19.463min (0.000) 0.00 ng

response 902039

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.46
183.00	13.20	11.62
0.00	0.00	0.00

DDT Breakdown

Date	Instrument Name	DFTPP Data File
7/27/2023	BNA_N	<u>BN026765.D</u>
Compound Name	Response	Retention Time
DDT	639647	20.686
DDD	14470	20.251
DDE	546	19.739
SUM(DDD+DDE)	SUM(DDT+DDD+DDE)	% Breakdown Of DDT
15016	654663	2.29

Instrument :
BNA_N
ClientSampleId :
DFTPP