

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072724\
 Data File : BN033041.D
 Acq On : 27 Jul 2024 20:15
 Operator : MA/JU
 Sample : P3375-02
 Misc :
 ALS Vial : 57 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GWBR-05-258-278-072524

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 07/29/2024
 Supervised By :mohammad ahmed 07/30/2024

Quant Time: Jul 29 01:07:52 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jul 21 23:33:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.589	152	1523m	0.400	ng	0.06
7) Naphthalene-d8	10.436	136	5664m	0.400	ng	0.13
13) Acenaphthene-d10	14.201	164	4151m	0.400	ng	0.02
19) Phenanthrene-d10	16.970	188	8884	0.400	ng	# 0.02
29) Chrysene-d12	21.169	240	8591	0.400	ng	0.00
35) Perylene-d12	23.361	264	9427	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.249	112	723m	0.188	ng	0.07
5) Phenol-d6	7.207	99	444m	0.093	ng	0.43
8) Nitrobenzene-d5	9.080	82	1096m	0.257	ng	0.35
11) 2-Methylnaphthalene-d10	12.030	152	2568m	0.296	ng	0.11
14) 2,4,6-Tribromophenol	15.766	330	430	0.242	ng	0.04
15) 2-Fluorobiphenyl	12.864	172	3928m	0.254	ng	0.05
27) Fluoranthene-d10	18.992	212	9349	0.407	ng	0.00
31) Terphenyl-d14	19.582	244	8102	0.460	ng	-0.01
Target Compounds						Qvalue
2) 1,4-Dioxane	3.097	88	546	0.289	ng	# 67
33) Chrysene	21.214	228	1331m	0.041	ng	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072724\
Data File : BN033041.D
Acq On : 27 Jul 2024 20:15
Operator : MA/JU
Sample : P3375-02
Misc :
ALS Vial : 57 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GWBR-05-258-278-072524

Quant Time: Jul 29 01:07:52 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Jul 21 23:33:21 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 07/29/2024
Supervised By :mohammad ahmed 07/30/2024

