

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072724\
 Data File : BN033050.D
 Acq On : 28 Jul 2024 01:40
 Operator : MA/JU
 Sample : PB162161BL
 Misc :
 ALS Vial : 66 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB162161BL

Manual Integrations
 APPROVED

Reviewed By :Jagrut Upadhyay 07/29/2024
 Supervised By :mohammad ahmed 07/30/2024

Quant Time: Jul 29 01:10:57 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sun Jul 21 23:33:21 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.575	152	3452m	0.400	ng	0.04
7) Naphthalene-d8	10.415	136	12933m	0.400	ng	0.11
13) Acenaphthene-d10	14.201	164	8660	0.400	ng	0.02
19) Phenanthrene-d10	16.982	188	17606	0.400	ng	0.03
29) Chrysene-d12	21.178	240	17027	0.400	ng	0.01
35) Perylene-d12	23.356	264	17299	0.400	ng	# 0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.221	112	2949m	0.338	ng	0.04
5) Phenol-d6	7.055	99	4721m	0.436	ng	0.28
8) Nitrobenzene-d5	8.995	82	3042m	0.312	ng	0.26
11) 2-Methylnaphthalene-d10	12.026	152	6323m	0.320	ng	0.11
14) 2,4,6-Tribromophenol	15.791	330	550	0.149	ng	0.07
15) 2-Fluorobiphenyl	12.886	172	9222m	0.286	ng	0.08
27) Fluoranthene-d10	19.006	212	16444	0.361	ng	0.01
31) Terphenyl-d14	19.596	244	13741	0.394	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072724\
Data File : BN033050.D
Acq On : 28 Jul 2024 01:40
Operator : MA/JU
Sample : PB162161BL
Misc :
ALS Vial : 66 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB162161BL

Manual Integrations
APPROVED

Reviewed By :Jagrut Upadhyay 07/29/2024
Supervised By :mohammad ahmed 07/30/2024

Quant Time: Jul 29 01:10:57 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sun Jul 21 23:33:21 2024
Response via : Initial Calibration

