

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072924\
 Data File : BN033063.D
 Acq On : 29 Jul 2024 10:51
 Operator : MA/JU
 Sample : PB162236BL
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 PB162236BL

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 07/30/2024
 Supervised By :mohammad ahmed 07/30/2024

Quant Time: Jul 29 11:35:21 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 29 10:37:08 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.611	152	2248m	0.400	ng	0.09
7) Naphthalene-d8	10.522	136	7922m	0.400	ng	0.19
13) Acenaphthene-d10	14.244	164	4941m	0.400	ng	0.07
19) Phenanthrene-d10	17.032	188	11397m	0.400	ng	0.07
29) Chrysene-d12	21.196	240	9263	0.400	ng	# 0.04
35) Perylene-d12	23.382	264	9302	0.400	ng	# 0.03
System Monitoring Compounds						
4) 2-Fluorophenol	5.242	112	2088	0.367	ng	0.00
5) Phenol-d6	7.120	99	2901m	0.411	ng	0.22
8) Nitrobenzene-d5	9.123	82	2053m	0.344	ng	0.27
11) 2-Methylnaphthalene-d10	12.120	152	4010m	0.331	ng	0.17
14) 2,4,6-Tribromophenol	15.840	330	387m	0.183	ng	0.10
15) 2-Fluorobiphenyl	12.961	172	6031m	0.328	ng	0.14
27) Fluoranthene-d10	19.034	212	11535m	0.391	ng	0.04
31) Terphenyl-d14	19.619	244	7142	0.376	ng	0.02

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN072924\
Data File : BN033063.D
Acq On : 29 Jul 2024 10:51
Operator : MA/JU
Sample : PB162236BL
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
PB162236BL

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 07/30/2024
Supervised By :mohammad ahmed 07/30/2024

Quant Time: Jul 29 11:35:21 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN071124.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 29 10:37:08 2024
Response via : Initial Calibration

