

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN073018\
 Data File : BN002185.D
 Acq On : 30 Jul 2018 13:57
 Operator : JU/SJ
 Sample : SSTDICC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6

Quant Time: Jul 30 14:38:27 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 30 14:09:58 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	3126	0.40	ng	0.00
7) Naphthalene-d8	10.47	136	12895	0.40	ng	0.00
13) Acenaphthene-d10	14.32	164	7241	0.40	ng	0.00
19) Phenanthrene-d10	17.07	188	14173	0.40	ng	0.00
27) Chrysene-d12	21.27	240	12429	0.40	ng	0.00
34) Perylene-d12	23.50	264	10573	0.40	ng	0.00

System Monitoring Compounds	R.T.	QIon	Response	Conc	Units	Dev(Min)
4) 2-Fluorophenol	5.31	112	11797	1.79	ng	0.00
5) Phenol-d6	6.87	99	14819	1.77	ng	0.00
8) Nitrobenzene-d5	8.83	82	16079	2.10	ng	-0.01
11) 2-Methylnaphthalene-d10	12.06	152	30303	1.56	ng	0.00
14) 2,4,6-Tribromophenol	15.81	330	5037	2.19	ng	-0.01
15) 2-Fluorobiphenyl	12.94	172	43986	1.28	ng	0.00
25) Fluoranthene-d10	19.10	212	53205	1.40	ng	0.00
29) Terphenyl-d14	19.71	244	40983	1.88	ng	0.00


Target Compounds	R.T.	QIon	Response	Conc	Units	Qvalue
2) 1,4-Dioxane	3.29	88	5237	1.617	ng	98
3) n-Nitrosodimethylamine	3.59	42	4557	2.069	ng	# 95
6) bis(2-Chloroethyl)ether	7.13	93	13864	1.301	ng	# 87
9) Naphthalene	10.52	128	46262	1.228	ng	97
10) Hexachlorobutadiene	10.80	225	9753	1.525	ng	# 55
12) 2-Methylnaphthalene	12.13	142	32275	1.307	ng	98
16) Acenaphthylene	14.03	152	51683	1.514	ng	100
17) Acenaphthene	14.38	154	32127	1.194	ng	97
18) Fluorene	15.37	166	38989	1.162	ng	# 80
20) 4-Bromophenyl-phenylether	16.26	248	13825	1.693	ng	# 84
21) Hexachlorobenzene	16.38	284	14705	1.346	ng	95
22) Pentachlorophenol	16.73	266	3424	1.173	ng	98
23) Phenanthrene	17.11	178	57933	1.162	ng	99
24) Anthracene	17.20	178	50275	1.329	ng	97
26) Fluoranthene	19.14	202	59735	1.155	ng	# 97
28) Pyrene	19.50	202	58208	1.193	ng	99
30) Benzo(a)anthracene	21.25	228	49784	1.271	ng	96
31) Chrysene	21.30	228	53336	1.166	ng	99
32) Bis(2-ethylhexyl)phthalate	21.18	149	20589	1.413	ng	# 94
33) Indeno(1,2,3-cd)pyrene	25.72	276	56520	1.323	ng	# 94
35) Benzo(b)fluoranthene	22.83	252	48949	1.014	ng	# 87
36) Benzo(k)fluoranthene	22.87	252	51503	1.091	ng	96
37) Benzo(a)pyrene	23.40	252	45702	1.162	ng	# 92
38) Dibenzo(a,h)anthracene	25.73	278	47214	1.348	ng	# 93
39) Benzo(g,h,i)perylene	26.40	276	47522	1.150	ng	# 90


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN073018\
 Data File : BN002185.D
 Acq On : 30 Jul 2018 13:57
 Operator : JU/SJ
 Sample : SSTDIC1.6
 Misc :
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 BNA_N
 Client Sampled :
 SSTDIC1.6

Quant Time: Jul 30 14:38:27 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 30 14:09:58 2018
 Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.69 min Scan# 579
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion	Resp	Lower	Upper
152	3126		
152	100		
150	154.8	117.2	175.8
115	61.2	57.0	85.6



#2
 1,4-Dioxane
 Concen: 1.617 ng
 RT: 3.29 min Scan# 32
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion	Resp	Lower	Upper
88	5237		
88	100		
58	88.3	72.0	108.0
43	35.4	28.0	42.0









#3
 n-Nitrosodimethylamine
 Concen: 2.069 ng
 RT: 3.59 min Scan# 70
 Delta R.T. -0.02 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Instrument :
 BNA_N
ClientSampleId :
 SSTDICC1.6

Tgt Ion: 42 Resp: 4557




Ion	Ratio	Lower	Upper
42	100		
74	111.1	88.0	132.0
44	6.8	16.0	24.0




#4
 2-Fluorophenol
 Concen: 1.787 ng
 RT: 5.31 min Scan# 283
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion: 112 Resp: 11797

Ion	Ratio	Lower	Upper
112	100		
64	64.7	60.1	90.1
63	35.0	116.8	175.2








#5
 Phenol-d6
 Concen: 1.767 ng
 RT: 6.87 min Scan# 477
 Delta R.T. -0.01 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57





Instrument :
 BNA_N
ClientSampled :
 SSTDICC1.6


Tgt Ion	Resp	Lower	Upper
99	14819		
42	22.9	20.2	30.2
71	36.1	28.4	42.6



#6
 bis(2-Chloroethyl)ether
 Concen: 1.301 ng
 RT: 7.13 min Scan# 509
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion	Resp	Lower	Upper
93	13864		
63	84.2	56.0	84.0#
95	32.8	25.2	37.8









#7
 Naphthalene-d8
 Concen: 0.400 ng
 RT: 10.47 min Scan# 835
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57




Instrument :
 BNA_N
 ClientSampled :
 SSTDICC1.6


Tgt Ion	Resp	Lower	Upper
136	12895		
137	11.7	9.5	14.3
54	6.9	29.1	43.7#
68	4.8	6.2	9.2#



#8
 Nitrobenzene-d5
 Concen: 2.102 ng
 RT: 8.83 min Scan# 706
 Delta R.T. -0.01 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion	Resp	Lower	Upper
82	16079		
128	34.5	150.0	225.0#
54	61.9	154.2	231.4#









#9
 Naphthalene
 Concen: 1.228 ng
 RT: 10.52 min Scan# 839
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57




Instrument :
 BNA_N
 ClientSampled :
 SSTDICC1.6


Tgt Ion	Resp	Lower	Upper
128	46262		
129	11.2	8.0	12.0
127	12.9	9.6	14.4



#10
 Hexachlorobutadiene
 Concen: 1.525 ng
 RT: 10.80 min Scan# 861
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion	Resp	Lower	Upper
225	9753		
223	0.0	52.0	78.0#
227	64.0	48.0	72.0








#11
 2-Methylnaphthalene-d10
 Concen: 1.561 ng
 RT: 12.06 min Scan# 1041
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57





Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion:152 Resp: 30303
 Ion Ratio Lower Upper
 152 100
 151 19.0 15.4 23.0



#12
 2-Methylnaphthalene
 Concen: 1.307 ng
 RT: 12.13 min Scan# 1057
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion:142 Resp: 32275
 Ion Ratio Lower Upper
 142 100
 141 88.9 70.4 105.6
 115 36.3 31.7 47.5









#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.32 min Scan# 1332
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Instrument :
 BNA_N
 ClientSampled :
 SSTDICC1.6

Tgt Ion:164 Resp: 7241




Ion	Ratio	Lower	Upper
164	100		
162	101.4	83.1	124.7
160	47.1	37.1	55.7




#14
 2,4,6-Tribromophenol
 Concen: 2.191 ng
 RT: 15.81 min Scan# 1466
 Delta R.T. -0.01 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion:330 Resp: 5037

Ion	Ratio	Lower	Upper
330	100		
332	0.0	152.7	229.1#
141	34.5	33.7	50.5









#15
 2-Fluorobiphenyl
 Concen: 1.281 ng
 RT: 12.94 min Scan# 1208
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion	Resp	Lower	Upper
172	43986		
171	34.2	29.9	44.9
170	22.5	21.8	32.8



#16
 Acenaphthylene
 Concen: 1.514 ng
 RT: 14.03 min Scan# 1306
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion	Resp	Lower	Upper
152	51683		
151	19.4	15.7	23.5
153	13.1	10.5	15.7









#17
 Acenaphthene
 Concen: 1.194 ng
 RT: 14.38 min Scan# 1337
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion	Resp	Lower	Upper
154	100		
153	113.8	89.2	133.8
152	55.5	42.0	63.0



#18
 Fluorene
 Concen: 1.162 ng
 RT: 15.37 min Scan# 1426
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion	Resp	Lower	Upper
166	100		
165	97.3	77.8	116.6
167	13.6	42.4	63.6#









#19
 Phenanthrene-d10
 Concen: 0.400 ng
 RT: 17.07 min Scan# 1583
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57




Instrument :
 BNA_N
 ClientSampled :
 SSTDICC1.6


Tgt Ion	Resp	Lower	Upper
188	14173		
94	0.0	3.9	5.9#
80	8.6	3.6	5.4#



#20
 4-Bromophenyl-phenylether
 Concen: 1.693 ng
 RT: 16.26 min Scan# 1507
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion	Resp	Lower	Upper
248	13825		
250	92.5	62.7	94.1
141	72.7	48.2	72.2#









#21
Hexachlorobenzene
Concen: 1.346 ng
RT: 16.38 min Scan# 1518
Delta R.T. -0.00 min
Lab File: BN002185.D
Acq: 30 Jul 2018 13:57




Instrument :
BNA_N
ClientSampled :
SSTDICC1.6


Tgt Ion	Resp	Lower	Upper
284	14705		
142	35.5	32.0	48.0
249	30.8	24.0	36.0



#22
Pentachlorophenol
Concen: 1.173 ng
RT: 16.73 min Scan# 1551
Delta R.T. -0.01 min
Lab File: BN002185.D
Acq: 30 Jul 2018 13:57

Tgt Ion	Resp	Lower	Upper
266	3424		
264	63.1	48.0	72.0
268	65.2	52.0	78.0









#23
 Phenanthrene
 Concen: 1.162 ng
 RT: 17.11 min Scan# 1587
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion	Resp	Lower	Upper
178	57933		
176	18.8	15.1	22.7
179	15.3	11.8	17.6



#24
 Anthracene
 Concen: 1.329 ng
 RT: 17.20 min Scan# 1595
 Delta R.T. -0.01 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion	Resp	Lower	Upper
178	50275		
176	18.1	12.8	19.2
179	15.5	13.0	19.6









#25
 Fluoranthene-d10
 Concen: 1.405 ng
 RT: 19.10 min Scan# 1891
 Delta R.T. -0.01 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57




Instrument :
 BNA_N
 ClientSampled :
 SSTDICC1.6


Tgt Ion	Resp	Lower	Upper
212	100		
106	11.4	2.8	4.2#
104	6.4	1.6	2.4#



#26
 Fluoranthene
 Concen: 1.155 ng
 RT: 19.14 min Scan# 1898
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion	Resp	Lower	Upper
202	100		
101	9.8	9.8	14.8#
203	17.3	13.7	20.5









#27
 Chrysene-d12
 Concen: 0.400 ng
 RT: 21.27 min Scan# 2192
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion	Resp	Lower	Upper
240	12429		
120	6.7	5.5	8.3
236	26.1	20.7	31.1



#28
 Pyrene
 Concen: 1.193 ng
 RT: 19.50 min Scan# 1971
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion	Resp	Lower	Upper
202	58208		
200	20.5	16.6	25.0
203	17.4	13.8	20.8









#29
 Terphenyl-d14
 Concen: 1.877 ng
 RT: 19.71 min Scan# 2013
 Delta R.T. -0.01 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57




Instrument :
 BNA_N
 ClientSampled :
 SSTDICC1.6


Tgt Ion	Resp	Lower	Upper
244	40983		
212	7.6	25.4	38.0#
122	10.8	8.1	12.1



#30
 Benzo(a)anthracene
 Concen: 1.271 ng
 RT: 21.25 min Scan# 2190
 Delta R.T. -0.01 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion	Resp	Lower	Upper
228	49784		
226	26.9	24.0	36.0
229	19.3	16.0	24.0









#31
 Chrysene
 Concen: 1.166 ng
 RT: 21.30 min Scan# 2195
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57




Instrument : BNA_N
 ClientSampled : SSTDICC1.6


Tgt Ion	Resp	Lower	Upper
228	100		
226	29.7	24.0	36.0
229	19.3	16.0	24.0



#32
 Bis(2-ethylhexyl)phthalate
 Concen: 1.413 ng
 RT: 21.18 min Scan# 2184
 Delta R.T. -0.01 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion	Resp	Lower	Upper
149	100		
167	27.9	20.0	30.0
279	4.3	1.6	2.4#









#33
 Indeno(1,2,3-cd)pyrene
 Concen: 1.323 ng
 RT: 25.72 min Scan# 3267
 Delta R.T. -0.01 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion	Resp	Lower	Upper
276	56520		
138	22.5	22.5	33.7#
277	24.9	19.9	29.9



#34
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.50 min Scan# 2618
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion	Resp	Lower	Upper
264	10573		
260	24.9	20.5	30.7
265	33.3	22.8	34.2









#35
 Benzo(b)fluoranthene
 Concen: 1.014 ng
 RT: 22.83 min Scan# 2423
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC1.6


Tgt Ion	Resp	Lower	Upper
252	48949		
253	22.3	20.0	30.0
125	9.4	16.0	24.0#



#36
 Benzo(k)fluoranthene
 Concen: 1.091 ng
 RT: 22.87 min Scan# 2435
 Delta R.T. -0.01 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion	Resp	Lower	Upper
252	51503		
253	22.1	16.0	24.0
125	9.3	8.0	12.0









#37
 Benzo(a)pyrene
 Concen: 1.162 ng
 RT: 23.40 min Scan# 2589
 Delta R.T. -0.00 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57




Instrument :
 BNA_N
 ClientSampled :
 SSTDICC1.6


Tgt Ion	Resp	Lower	Upper
252	45702		
253	22.3	16.0	24.0
125	10.5	12.0	18.0



#38
 Dibenzo(a,h)anthracene
 Concen: 1.348 ng
 RT: 25.73 min Scan# 3271
 Delta R.T. -0.01 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Tgt Ion	Resp	Lower	Upper
278	47214		
139	15.1	16.0	24.0
279	23.4	20.0	30.0





#39
 Benzo(g,h,i)perylene
 Concen: 1.150 ng
 RT: 26.40 min Scan# 3465
 Delta R.T. -0.01 min
 Lab File: BN002185.D
 Acq: 30 Jul 2018 13:57

Instrument :
 BNA_N
ClientSampleId :
 SSTDICC1.6

Tgt Ion: 276 Resp: 47522

Ion	Ratio	Lower	Upper
276	100		
277	23.7	16.0	24.0
138	19.2	20.0	30.0#

