

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN073018\
 Data File : BN002188.D
 Acq On : 30 Jul 2018 16:19
 Operator : JU/SJ
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN073018

Quant Time: Jul 30 16:49:10 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN073018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 30 16:17:29 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
2	1,4-Dioxane	0.450	0.437	2.9	92	0.00
3	n-Nitrosodimethylamine	0.302	0.266	11.9	84	0.00
4 S	2-Fluorophenol	0.933	0.908	2.7	87	0.00
5 S	Phenol-d6	1.135	1.104	2.7	89	0.00
6	bis(2-Chloroethyl)ether	1.091	1.067	2.2	84	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
8 S	Nitrobenzene-d5	0.301	0.300	0.3	89	0.00
9	Naphthalene	0.902	0.925	-2.5	92	0.00
10	Hexachlorobutadiene	0.195	0.200	-2.6	92	0.00
11 SURR	2-Methylnaphthalene-d10	0.589	0.596	-1.2	92	0.00
12	2-Methylnaphthalene	0.610	0.624	-2.3	91	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	84	0.00
14 S	2,4,6-Tribromophenol	0.171	0.148	13.5	75	0.00
15 S	2-Fluorobiphenyl	1.536	1.560	-1.6	91	0.00
16	Acenaphthylene	1.709	1.677	1.9	87	0.01
17	Acenaphthene	1.099	1.105	-0.5	88	0.00
18	Fluorene	1.352	1.356	-0.3	85	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	71	0.00
20	4-Bromophenyl-phenylether	0.226	0.233	-3.1	80	0.01
21	Hexachlorobenzene	0.250	0.265	-6.0	81	0.00
22	Pentachlorophenol	0.046	0.038	17.4	62	0.00
23	Phenanthrene	0.995	1.007	-1.2	75	0.00
24	Anthracene	0.823	0.789	4.1	72	0.00
25 SURR	Fluoranthene-d10	0.988	0.900	8.9	64	0.00
26	Fluoranthene	1.083	0.984	9.1	63	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	52	0.00
28	Pvrene	1.183	1.305	-10.3	62	0.00
29 S	Terphenyl-d14	0.855	0.905	-5.8	60	0.00
30	Benzo(a)anthracene	1.006	0.987	1.9	54	0.00
31	Chrysene	1.126	1.161	-3.1	56	0.00
32	Bis(2-ethylhexyl)phthalate	0.460	0.408	11.3	47#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.061	1.188	-12.0	69	0.00
34 I	Perylene-d12	1.000	1.000	0.0	58	0.00
35	Benzo(b)fluoranthene	1.176	1.158	1.5	58	0.00
36	Benzo(k)fluoranthene	1.221	1.223	-0.2	61	0.00
37 C	Benzo(a)pyrene	1.066	1.081	-1.4	62	0.00
38	Dibenzo(a,h)anthracene	1.037	1.093	-5.4	70	0.00
39	Benzo(a,h,i)perylene	1.060	1.130	-6.6	70	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN073018\
Data File : BN002188.D
Acq On : 30 Jul 2018 16:19
Operator : JU/SJ
Sample : SSTDICV0.4
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN073018

Quant Time: Jul 30 16:49:10 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN073018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 30 16:17:29 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			