

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN073018\
 Data File : BN002199.D
 Acq On : 30 Jul 2018 23:07
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 31 01:26:40 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN073018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 30 16:17:29 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	86	0.00
2	1,4-Dioxane	0.450	0.419	6.9	89	0.00
3	n-Nitrosodimethylamine	0.302	0.232	23.2	74	0.00
4 S	2-Fluorophenol	0.933	0.845	9.4	82	0.00
5 S	Phenol-d6	1.135	1.004	11.5	82	0.00
6	bis(2-Chloroethyl)ether	1.091	1.015	7.0	80	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
8 S	Nitrobenzene-d5	0.301	0.283	6.0	84	0.00
9	Naphthalene	0.902	0.878	2.7	87	0.00
10	Hexachlorobutadiene	0.195	0.194	0.5	89	0.00
11 SURR	2-Methylnaphthalene-d10	0.589	0.555	5.8	85	0.00
12	2-Methylnaphthalene	0.610	0.578	5.2	84	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	79	0.00
14 S	2,4,6-Tribromophenol	0.171	0.139	18.7	66	0.00
15 S	2-Fluorobiphenyl	1.536	1.537	-0.1	84	0.00
16	Acenaphthylene	1.709	1.603	6.2	78	0.01
17	Acenaphthene	1.099	1.070	2.6	81	0.00
18	Fluorene	1.352	1.256	7.1	74	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	66	0.00
20	4-Bromophenyl-phenylether	0.226	0.221	2.2	70	0.01
21	Hexachlorobenzene	0.250	0.256	-2.4	72	0.00
22	Pentachlorophenol	0.046	0.036	21.7	54	0.01
23	Phenanthrene	0.995	0.947	4.8	65	0.00
24	Anthracene	0.823	0.739	10.2	62	0.00
25 SURR	Fluoranthene-d10	0.988	0.864	12.6	56	0.00
26	Fluoranthene	1.083	0.939	13.3	55	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	51	0.00
28	Pvrene	1.183	1.184	-0.1	55	0.00
29 S	Terphenyl-d14	0.855	0.828	3.2	53	0.00
30	Benzo(a)anthracene	1.006	0.934	7.2	50#	0.00
31	Chrysene	1.126	1.106	1.8	52	0.00
32	Bis(2-ethylhexyl)phthalate	0.460	0.391	15.0	44#	-0.01
33	Indeno(1,2,3-cd)pyrene	1.061	1.186	-11.8	68	0.00
34 I	Perylene-d12	1.000	1.000	0.0	59	0.00
35	Benzo(b)fluoranthene	1.176	1.124	4.4	57	0.00
36	Benzo(k)fluoranthene	1.221	1.159	5.1	59	0.00
37 C	Benzo(a)pyrene	1.066	1.048	1.7	60	0.00
38	Dibenzo(a,h)anthracene	1.037	1.047	-1.0	68	0.00
39	Benzo(a,h,i)perylene	1.060	1.081	-2.0	68	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN073018\
Data File : BN002199.D
Acq On : 30 Jul 2018 23:07
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 4 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 31 01:26:40 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN073018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 30 16:17:29 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			