

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN073021\
 Data File : BN015721.D
 Acq On : 30 Jul 2021 16:52
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Jul 30 18:46:40 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 30 18:43:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.790	152	30776	0.400	ng	0.00
7) Naphthalene-d8	10.571	136	137038	0.400	ng	0.00
13) Acenaphthene-d10	14.421	164	71047	0.400	ng	0.00
19) Phenanthrene-d10	17.164	188	134703	0.400	ng	0.00
29) Chrysene-d12	21.376	240	122225	0.400	ng	0.00
35) Perylene-d12	23.721	264	122721	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.383	112	9825	0.119	ng	0.00
5) Phenol-d6	6.951	99	12186	0.116	ng	0.00
8) Nitrobenzene-d5	8.937	82	8927	0.097	ng	0.00
11) 2-Methylnaphthalene-d10	12.167	152	21036	0.098	ng	0.00
14) 2,4,6-Tribromophenol	15.918	330	3025	0.105	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	32991	0.122	ng	0.00
27) Fluoranthene-d10	19.212	212	34805	0.091	ng	0.00
31) Terphenyl-d14	19.817	244	35589	0.114	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.355	88	914	0.085	ng	# 78
3) n-Nitrosodimethylamine	3.742	42	1292	0.061	ng	98
6) bis(2-Chloroethyl)ether	7.218	93	8989	0.109	ng	99
9) Naphthalene	10.622	128	39118	0.110	ng	100
10) Hexachlorobutadiene	10.914	225	6807	0.099	ng	# 100
12) 2-Methylnaphthalene	12.238	142	25474	0.106	ng	99
16) Acenaphthylene	14.129	152	32259	0.100	ng	100
17) Acenaphthene	14.484	154	21741	0.108	ng	99
18) Fluorene	15.474	166	26501	0.105	ng	100
20) 4,6-Dinitro-2-methylph...	15.550	198	612	0.061	ng	# 68
21) 4-Bromophenyl-phenylether	16.360	248	8231	0.103	ng	# 64
22) Hexachlorobenzene	16.482	284	9445	0.121	ng	99
23) Atrazine	16.640	200	5455	0.076	ng	98
24) Pentachlorophenol	16.823	266	2726	0.095	ng	99
25) Phenanthrene	17.212	178	42624	0.115	ng	99
26) Anthracene	17.297	178	36138	0.101	ng	99
28) Fluoranthene	19.242	202	44372	0.104	ng	100
30) Pyrene	19.604	202	45686	0.109	ng	100
32) Benzo(a)anthracene	21.366	228	40656	0.098	ng	97
33) Chrysene	21.419	228	44637	0.113	ng	97
34) Bis(2-ethylhexyl)phtha...	21.312	149	15739	0.066	ng	99
36) Indeno(1,2,3-cd)pyrene	26.127	276	44244	0.099	ng	99
37) Benzo(b)fluoranthene	23.012	252	41179	0.099	ng	# 95
38) Benzo(k)fluoranthene	23.012	252	41179	0.100	ng	97
39) Benzo(a)pyrene	23.618	252	37892	0.101	ng	# 93
40) Dibenzo(a,h)anthracene	26.147	278	37139	0.098	ng	# 93
41) Benzo(g,h,i)perylene	26.856	276	41495	0.109	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN073021\
Data File : BN015721.D
Acq On : 30 Jul 2021 16:52
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Jul 30 18:46:40 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 30 18:43:48 2021
Response via : Initial Calibration









1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.790 min Scan# 511
 Delta R.T. -0.000 min
 Lab File: BN015721.D
 Acq: 30 Jul 2021 16:52




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:152 Resp: 30776
 Ion Ratio Lower Upper
 152 100
 150 155.2 123.9 185.9
 115 59.7 47.8 71.6



1,4-Dioxane
 Concen: 0.085 ng
 RT: 3.355 min Scan# 30
 Delta R.T. 0.000 min
 Lab File: BN015721.D
 Acq: 30 Jul 2021 16:52

Tgt Ion: 88 Resp: 914
 Ion Ratio Lower Upper
 88 100
 43 38.0 17.2 25.8#
 58 93.5 62.2 93.2#









#3
n-Nitrosodimethylamine
Concen: 0.061 ng
RT: 3.742 min Scan# 72
Delta R.T. 0.028 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:	42	Resp:	1292
Ion	Ratio	Lower	Upper
42	100		
74	143.7	116.6	174.8
44	7.9	6.6	10.0



#4
2-Fluorophenol
Concen: 0.119 ng
RT: 5.383 min Scan# 250
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Tgt Ion:	112	Resp:	9825
Ion	Ratio	Lower	Upper
112	100		
64	53.2	42.1	63.1
63	28.8	21.9	32.9









Phenol-d6
 Concen: 0.116 ng
 RT: 6.951 min Scan# 420
 Delta R.T. 0.000 min
 Lab File: BN015721.D
 Acq: 30 Jul 2021 16:52




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	99	Resp:	12186
Ion	Ratio	Lower	Upper
99	100		
42	16.5	12.9	19.3
71	30.2	23.8	35.6




bis(2-Chloroethyl)ether
 Concen: 0.109 ng
 RT: 7.218 min Scan# 449
 Delta R.T. 0.000 min
 Lab File: BN015721.D
 Acq: 30 Jul 2021 16:52

Tgt Ion:	93	Resp:	8989
Ion	Ratio	Lower	Upper
93	100		
63	74.3	58.6	87.8
95	33.5	26.9	40.3






Abundance Scan 746 (10.571 min): BN015723.D\data.ms (-730) (-)




Naphthalene-d8
Concen: 0.400 ng
RT: 10.571 min Scan# 746
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:	136	Resp:	137038
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	6.4	5.5	8.3
68	5.1	4.5	6.7






Abundance Scan 617 (8.936 min): BN015723.D\data.ms (-614) (-)




Nitrobenzene-d5
Concen: 0.097 ng
RT: 8.937 min Scan# 617
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Tgt Ion:	82	Resp:	8927
Ion	Ratio	Lower	Upper
82	100		
128	42.2	33.2	49.8
54	49.9	38.4	57.6






Abundance Scan 750 (10.622 min): BN015723.D\data.ms (-749) (-)




Naphthalene
Concen: 0.110 ng
RT: 10.622 min Scan# 750
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:	128	Resp:	39118
Ion	Ratio	Lower	Upper
128	100		
129	11.3	8.9	13.3
127	13.0	10.4	15.6






Abundance Scan 773 (10.913 min): BN015723.D\data.ms (-769) (-)




Hexachlorobutadiene
Concen: 0.099 ng
RT: 10.914 min Scan# 773
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Tgt Ion:	225	Resp:	6807
Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.0	51.0	76.4




Abundance Scan 968 (12.167 min): BN015723.D\data.ms (-95) #11




2-Methylnaphthalene-d10
Concen: 0.098 ng
RT: 12.167 min Scan# 968
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:152 Resp: 21036
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.4



Abundance Scan 968 (12.167 min): BN015721.D\data.ms (-923) (-)





Abundance Scan 984 (12.238 min): BN015723.D\data.ms (-97) #12




2-Methylnaphthalene
Concen: 0.106 ng
RT: 12.238 min Scan# 984
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Tgt Ion:142 Resp: 25474
Ion Ratio Lower Upper
142 100
141 88.5 70.5 105.7
115 35.0 27.3 40.9




Abundance Scan 984 (12.238 min): BN015721.D\data.ms (-961) (-)






Abundance Scan 1195 (14.421 min): BN015723.D\data.ms (-1195) (-)

Acenaphthene-d10
Concen: 0.400 ng
RT: 14.421 min Scan# 1195
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1







Tgt Ion:	164	Resp:	71047
Ion	Ratio	Lower	Upper
164	100		
162	99.3	79.8	119.6
160	44.3	35.8	53.6




Abundance Scan 1313 (15.918 min): BN015723.D\data.ms (-1313) (-)

2,4,6-Tribromophenol
Concen: 0.105 ng
RT: 15.918 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Tgt Ion:	330	Resp:	3025
Ion	Ratio	Lower	Upper
330	100		
332	0.0	0.0	0.0
141	36.1	28.8	43.2









2-Fluorobiphenyl
Concen: 0.122 ng
RT: 13.051 min Scan# 1087
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:	172	Resp:	32991
Ion	Ratio	Lower	Upper
172	100		
171	33.4	26.9	40.3
170	21.8	17.4	26.0



Acenaphthylene
Concen: 0.100 ng
RT: 14.129 min Scan# 1172
Delta R.T. -0.012 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Tgt Ion:	152	Resp:	32259
Ion	Ratio	Lower	Upper
152	100		
151	19.7	15.7	23.5
153	13.3	10.4	15.6





Abundance Scan 1200 (14.484 min): BN015723.D\data.ms (-1197) (-)

Acenaphthene
Concen: 0.108 ng
RT: 14.484 min Scan# 1200
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion:154 Resp: 21741
Ion Ratio Lower Upper
154 100
153 113.4 89.7 134.5
152 56.2 43.8 65.6






Abundance Scan 1200 (14.484 min): BN015721.D\data.ms (-1192) (-)




Abundance Scan 1278 (15.474 min): BN015723.D\data.ms (-1213) (-)

Fluorene
Concen: 0.105 ng
RT: 15.474 min Scan# 1278
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Tgt Ion:166 Resp: 26501
Ion Ratio Lower Upper
166 100
165 97.8 78.1 117.1
167 13.3 10.8 16.2




Abundance Scan 1278 (15.474 min): BN015721.D\data.ms (-1262) (-)





Abundance Scan 1415 (17.163 min): BN015723.D\data.ms (-1419) (-)

Phenanthrene-d10
Concen: 0.400 ng
RT: 17.164 min Scan# 1415
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion:	188	Resp:	134703
Ion	Ratio	Lower	Upper
188	100		
94	0.0	0.0	0.0
80	12.1	10.2	15.4





Abundance Scan 1415 (17.164 min): BN015721.D\data.ms (-1399) (-)




Abundance Scan 1284 (15.550 min): BN015723.D\data.ms (-1278) (-)

4,6-Dinitro-2-methylphenol
Concen: 0.061 ng
RT: 15.550 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Tgt Ion:	198	Resp:	612
Ion	Ratio	Lower	Upper
198	100		
182	24.0	9.4	14.0#
77	0.0	0.0	0.0



Abundance Scan 1284 (15.550 min): BN015721.D\data.ms (-1276) (-)






Abundance Scan 1350 (16.372 min): BN015723.D\data.ms (-1325) (-)




4-Bromophenyl-phenylether
Concen: 0.103 ng
RT: 16.360 min Scan# 1349
Delta R.T. -0.012 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:	248	Resp:	8231
Ion	Ratio	Lower	Upper
248	100		
250	90.1	83.9	125.9
141	91.8	26.9	40.3#






Abundance Scan 1359 (16.482 min): BN015723.D\data.ms (-1322) (-)



Hexachlorobenzene
Concen: 0.121 ng
RT: 16.482 min Scan# 1359
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Tgt Ion:	284	Resp:	9445
Ion	Ratio	Lower	Upper
284	100		
142	37.6	29.4	44.0
249	30.0	23.8	35.6





Abundance Scan 1372 (16.640 min): BN015723.D\data.ms (-1325) (-)

Atrazine
Concen: 0.076 ng
RT: 16.640 min Scan# 1372
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion:	200	Resp:	5455
Ion	Ratio	Lower	Upper
200	100		
173	24.2	18.9	28.3
215	49.3	38.1	57.1





Abundance Scan 1372 (16.640 min): BN015721.D\data.ms (-1331) (-)




Abundance Scan 1387 (16.823 min): BN015723.D\data.ms (-1321) (-)

Pentachlorophenol
Concen: 0.095 ng
RT: 16.823 min Scan# 1387
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Tgt Ion:	266	Resp:	2726
Ion	Ratio	Lower	Upper
266	100		
264	62.7	50.2	75.2
268	64.9	51.1	76.7



Abundance Scan 1387 (16.823 min): BN015721.D\data.ms (-1379) (-)





Abundance Scan 1419 (17.212 min): BN015723.D\data.ms (-1425) (-)




Phenanthrene
Concen: 0.115 ng
RT: 17.212 min Scan# 1419
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Tgt Ion:	178	Resp:	42624
Ion	Ratio	Lower	Upper
178	100		
176	19.0	15.3	22.9
179	16.1	12.4	18.6


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Abundance Scan 1419 (17.212 min): BN015721.D\data.ms (-1394) (-)





Abundance Scan 1426 (17.297 min): BN015723.D\data.ms (-1425) (-)




Anthracene
Concen: 0.101 ng
RT: 17.297 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Tgt Ion:	178	Resp:	36138
Ion	Ratio	Lower	Upper
178	100		
176	18.6	14.6	22.0
179	15.2	12.2	18.4



Abundance Scan 1426 (17.297 min): BN015721.D\data.ms (-1418) (-)






Abundance Scan 1726 (19.212 min): BN015723.D\data.ms (-1705) (-)




Fluoranthene-d10
Concen: 0.091 ng
RT: 19.212 min Scan# 1726
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:212 Resp: 34805
Ion Ratio Lower Upper
212 100
106 13.5 10.6 16.0
104 7.5 5.9 8.9






Abundance Scan 1732 (19.241 min): BN015723.D\data.ms (-1724) (-)




Fluoranthene
Concen: 0.104 ng
RT: 19.242 min Scan# 1732
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Tgt Ion:202 Resp: 44372
Ion Ratio Lower Upper
202 100
101 11.5 9.0 13.6
203 17.2 13.8 20.8






Abundance Scan 2015 (21.376 min): BN015723.D\data.ms (-1996) (-)




Chrysene-d12
Concen: 0.400 ng
RT: 21.376 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:	240	Resp:	122225
Ion	Ratio	Lower	Upper
240	100		
120	10.2	7.4	11.0
236	26.4	20.9	31.3






Abundance Scan 1805 (19.604 min): BN015723.D\data.ms (-1734) (-)



Pyrene
Concen: 0.109 ng
RT: 19.604 min Scan# 1805
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Tgt Ion:	202	Resp:	45686
Ion	Ratio	Lower	Upper
202	100		
200	20.7	16.5	24.7
203	17.6	14.0	21.0






Abundance Scan 1848 (19.817 min): BN015723.D\data.ms (-1818) (-)

Terphenyl-d14
Concen: 0.114 ng
RT: 19.817 min Scan# 1848
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Instrument :
BNA_N
ClientSampled :
SSTDIC0.1




Tgt Ion	Ratio	Lower	Upper
244	100		
212	7.8	5.9	8.9
122	12.6	9.6	14.4






Abundance Scan 2013 (21.355 min): BN015723.D\data.ms (-2036) (-)


Benzo(a)anthracene
Concen: 0.098 ng
RT: 21.366 min Scan# 2014
Delta R.T. 0.011 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52



Tgt Ion	Ratio	Lower	Upper
228	100		
226	25.8	21.9	32.9
229	20.3	15.4	23.0






Abundance Scan 2018 (21.408 min): BN015723.D\data.ms (-2018) (-)




Chrysene
Concen: 0.113 ng
RT: 21.419 min Scan# 2019
Delta R.T. 0.011 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:	228	Resp:	44637
Ion	Ratio	Lower	Upper
228	100		
226	28.4	24.3	36.5
229	20.2	15.3	22.9






Abundance Scan 2009 (21.312 min): BN015723.D\data.ms (-2009) (-)



Bis(2-ethylhexyl)phthalate
Concen: 0.066 ng
RT: 21.312 min Scan# 2009
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Tgt Ion:	149	Resp:	15739
Ion	Ratio	Lower	Upper
149	100		
167	26.3	21.5	32.3
279	4.2	3.2	4.8





Abundance Scan 2496 (23.720 min): BN015723.D\data.ms (-2452) (-)

Perylene-d12
Concen: 0.400 ng
RT: 23.721 min Scan# 2496
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1



Tgt Ion:	264	Resp:	122721
Ion	Ratio	Lower	Upper
264	100		
260	24.7	19.4	29.2
265	26.2	19.9	29.9






Abundance Scan 2496 (23.721 min): BN015721.D\data.ms (-2452) (-)




Abundance Scan 3199 (26.127 min): BN015723.D\data.ms (-3170) (-)

Indeno(1,2,3-cd)pyrene
Concen: 0.099 ng
RT: 26.127 min Scan# 3199
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Tgt Ion:	276	Resp:	44244
Ion	Ratio	Lower	Upper
276	100		
138	26.8	20.6	30.8
277	25.2	20.1	30.1



Abundance Scan 3199 (26.127 min): BN015721.D\data.ms (-3170) (-)






Abundance Scan 2289 (23.012 min): BN015723.D\data.ms (-2231) (-)




Benzo(b)fluoranthene
Concen: 0.099 ng
RT: 23.012 min Scan# 2289
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Tgt Ion:252 Resp: 41179
Ion Ratio Lower Upper
252 100
253 24.8 18.1 27.1
125 12.7 8.3 12.5#






Abundance Scan 2302 (23.056 min): BN015723.D\data.ms (-2236) (-)



Benzo(k)fluoranthene
Concen: 0.100 ng
RT: 23.012 min Scan# 2289
Delta R.T. -0.044 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Tgt Ion:252 Resp: 41179
Ion Ratio Lower Upper
252 100
253 24.8 18.1 27.1
125 12.7 10.4 15.6






Abundance Scan 2466 (23.618 min): BN015723.D\data.ms (-2408) (-)

Benzo(a)pyrene
Concen: 0.101 ng
RT: 23.618 min Scan# 2466
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:	252	Resp:	37892
Ion	Ratio	Lower	Upper
252	100		
253	25.9	18.1	27.1
125	14.6	9.2	13.8#






Abundance Scan 3205 (26.147 min): BN015723.D\data.ms (-3182) (-)


Dibenzo(a,h)anthracene
Concen: 0.098 ng
RT: 26.147 min Scan# 3205
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52



Tgt Ion:	278	Resp:	37139
Ion	Ratio	Lower	Upper
278	100		
139	25.0	16.0	24.0#
279	25.9	19.2	28.8




Abundance Scan 3413 (26.859 min): BN015723.D\data.ms (-3351) (-)



Benzo(g,h,i)perylene
Concen: 0.109 ng
RT: 26.856 min Scan# 3412
Delta R.T. -0.003 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Abundance Scan 3412 (26.856 min): BN015721.D\data.ms



Tgt Ion:	276	Resp:	41495
Ion	Ratio	Lower	Upper
276	100		
277	24.4	19.2	28.8
138	22.8	16.8	25.2

Abundance Scan 3412 (26.856 min): BN015721.D\data.ms (-3355) (-)

