

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN073021\
 Data File : BN015724.D
 Acq On : 30 Jul 2021 18:43
 Operator : CG/JU
 Sample : SSTDIC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDIC0.8

Quant Time: Jul 31 00:32:17 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 30 18:55:25 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.790	152	34412	0.40	ng	0.00
7) Naphthalene-d8	10.572	136	150139	0.40	ng	0.00
13) Acenaphthene-d10	14.421	164	80027	0.40	ng	0.00
19) Phenanthrene-d10	17.164	188	150077	0.40	ng	0.00
29) Chrysene-d12	21.376	240	145057	0.40	ng	0.00
35) Perylene-d12	23.717	264	144540	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.393	112	73405	0.79	ng	0.00
5) Phenol-d6	6.951	99	88526	0.76	ng	0.00
8) Nitrobenzene-d5	8.937	82	70864	0.70	ng	0.00
11) 2-Methylnaphthalene-d10	12.167	152	184641	0.78	ng	0.00
14) 2,4,6-Tribromophenol	15.918	330	27056	0.83	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	235269	0.77	ng	0.00
27) Fluoranthene-d10	19.212	212	323365	0.76	ng	0.00
31) Terphenyl-d14	19.818	244	269317	0.73	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.355	88	11175	0.93	ng	Qvalue # 84
3) n-Nitrosodimethylamine	3.705	42	18657	0.79	ng	# 99
6) bis(2-Chloroethyl)ether	7.218	93	74855	0.81	ng	99
9) Naphthalene	10.622	128	312059	0.80	ng	100
10) Hexachlorobutadiene	10.914	225	54918	0.73	ng	# 100
12) 2-Methylnaphthalene	12.238	142	205260	0.78	ng	100
16) Acenaphthylene	14.142	152	283111	0.78	ng	100
17) Acenaphthene	14.485	154	182715	0.81	ng	99
18) Fluorene	15.474	166	227037	0.80	ng	100
20) 4,6-Dinitro-2-methylph...	15.550	198	9113	0.82	ng	# 93
21) 4-Bromophenyl-phenylether	16.360	248	69872	0.78	ng	# 63
22) Hexachlorobenzene	16.482	284	77080	0.89	ng	100
23) Atrazine	16.640	200	53688	0.67	ng	99
24) Pentachlorophenol	16.823	266	28112	0.88	ng	100
25) Phenanthrene	17.212	178	342442	0.83	ng	100
26) Anthracene	17.298	178	314914	0.79	ng	100
28) Fluoranthene	19.237	202	377803	0.80	ng	100
30) Pyrene	19.604	202	376198	0.76	ng	100
32) Benzo(a)anthracene	21.355	228	368664	0.75	ng	100
33) Chrysene	21.408	228	371303	0.79	ng	99
34) Bis(2-ethylhexyl)phtha...	21.302	149	181000	0.64	ng	100
36) Indeno(1,2,3-cd)pyrene	26.127	276	423473	0.80	ng	99
37) Benzo(b)fluoranthene	23.009	252	377114	0.77	ng	99
38) Benzo(k)fluoranthene	23.057	252	406038	0.84	ng	98
39) Benzo(a)pyrene	23.615	252	343696	0.78	ng	99
40) Dibenzo(a,h)anthracene	26.144	278	358543	0.81	ng	97
41) Benzo(g,h,i)perylene	26.856	276	362852	0.81	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN073021\
Data File : BN015724.D
Acq On : 30 Jul 2021 18:43
Operator : CG/JU
Sample : SSTDIC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8

Quant Time: Jul 31 00:32:17 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 30 18:55:25 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.790 min Scan# 511
 Delta R.T. 0.000 min
 Lab File: BN015724.D
 Acq: 30 Jul 2021 18:43




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:152 Resp: 34412
 Ion Ratio Lower Upper
 152 100
 150 156.7 123.9 185.9
 115 60.0 47.8 71.6



#2
 1,4-Dioxane
 Concen: 0.93 ng
 RT: 3.355 min Scan# 30
 Delta R.T. 0.000 min
 Lab File: BN015724.D
 Acq: 30 Jul 2021 18:43

Tgt Ion: 88 Resp: 11175
 Ion Ratio Lower Upper
 88 100
 43 22.4 17.2 25.8
 58 60.4 62.2 93.2#









#3
n-Nitrosodimethylamine
Concen: 0.79 ng
RT: 3.705 min Scan# 68
Delta R.T. -0.009 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 42 Resp: 18657
Ion Ratio Lower Upper
42 100
74 147.1 116.6 174.8
44 6.3 6.6 10.0#



#4
2-Fluorophenol
Concen: 0.79 ng
RT: 5.393 min Scan# 251
Delta R.T. 0.009 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43

Tgt Ion: 112 Resp: 73405
Ion Ratio Lower Upper
112 100
64 52.2 42.1 63.1
63 27.3 21.9 32.9









#5
Phenol-d6
Concen: 0.76 ng
RT: 6.951 min Scan# 420
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	99	Resp:	88526
Ion	Ratio	Lower	Upper
99	100		
42	15.9	12.9	19.3
71	29.6	23.8	35.6



#6
bis(2-Chloroethyl)ether
Concen: 0.81 ng
RT: 7.218 min Scan# 449
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43

Tgt Ion:	93	Resp:	74855
Ion	Ratio	Lower	Upper
93	100		
63	72.5	58.6	87.8
95	33.0	26.9	40.3









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.572 min Scan# 746
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:	136	Resp:	150139
Ion Ratio	Lower	Upper	
136	100		
137	10.9	8.7	13.1
54	6.9	5.5	8.3
68	5.5	4.5	6.7



#8
Nitrobenzene-d5
Concen: 0.70 ng
RT: 8.937 min Scan# 617
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43

Tgt Ion:	82	Resp:	70864
Ion Ratio	Lower	Upper	
82	100		
128	42.6	33.2	49.8
54	47.6	38.4	57.6









#9
Naphthalene
Concen: 0.80 ng
RT: 10.622 min Scan# 750
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:128 Resp: 312059
Ion Ratio Lower Upper
128 100
129 10.9 8.9 13.3
127 12.9 10.4 15.6



#10
Hexachlorobutadiene
Concen: 0.73 ng
RT: 10.914 min Scan# 773
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43

Tgt Ion:225 Resp: 54918
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 51.0 76.4









#11
2-Methylnaphthalene-d10
Concen: 0.78 ng
RT: 12.167 min Scan# 968
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:152 Resp: 184641
Ion Ratio Lower Upper
152 100
151 20.2 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.78 ng
RT: 12.238 min Scan# 984
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43

Tgt Ion:142 Resp: 205260
Ion Ratio Lower Upper
142 100
141 88.3 70.5 105.7
115 33.9 27.3 40.9









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.421 min Scan# 1195
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:164 Resp: 80027
Ion Ratio Lower Upper
164 100
162 100.5 79.8 119.6
160 44.6 35.8 53.6



#14
2,4,6-Tribromophenol
Concen: 0.83 ng
RT: 15.918 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43

Tgt Ion:330 Resp: 27056
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 35.3 28.8 43.2









#15
2-Fluorobiphenyl
Concen: 0.77 ng
RT: 13.051 min Scan# 1087
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:172	Resp: 235269
Ion Ratio	Lower Upper
172	100
171	33.2 26.9 40.3
170	21.4 17.4 26.0



#16
Acenaphthylene
Concen: 0.78 ng
RT: 14.142 min Scan# 1173
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43

Tgt Ion:152	Resp: 283111
Ion Ratio	Lower Upper
152	100
151	19.5 15.7 23.5
153	13.1 10.4 15.6









#17
Acenaphthene
Concen: 0.81 ng
RT: 14.485 min Scan# 1200
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:154 Resp: 182715
Ion Ratio Lower Upper
154 100
153 110.9 89.7 134.5
152 53.9 43.8 65.6



#18
Fluorene
Concen: 0.80 ng
RT: 15.474 min Scan# 1278
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43

Tgt Ion:166 Resp: 227037
Ion Ratio Lower Upper
166 100
165 97.1 78.1 117.1
167 13.5 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.164 min Scan# 1415
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:188 Resp: 150077
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.1 10.2 15.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.82 ng
RT: 15.550 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43

Tgt Ion:198 Resp: 9113
Ion Ratio Lower Upper
198 100
182 9.0 9.4 14.0#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.78 ng
RT: 16.360 min Scan# 1349
Delta R.T. -0.012 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:248 Resp: 69872
Ion Ratio Lower Upper
248 100
250 89.9 83.9 125.9
141 94.7 26.9 40.3#



#22
Hexachlorobenzene
Concen: 0.89 ng
RT: 16.482 min Scan# 1359
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43

Tgt Ion:284 Resp: 77080
Ion Ratio Lower Upper
284 100
142 36.8 29.4 44.0
249 29.8 23.8 35.6









#23
Atrazine
Concen: 0.67 ng
RT: 16.640 min Scan# 1372
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:200 Resp: 53688
Ion Ratio Lower Upper
200 100
173 23.2 18.9 28.3
215 48.4 38.1 57.1



#24
Pentachlorophenol
Concen: 0.88 ng
RT: 16.823 min Scan# 1387
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43

Tgt Ion:266 Resp: 28112
Ion Ratio Lower Upper
266 100
264 62.6 50.2 75.2
268 63.5 51.1 76.7









#25
Phenanthrene
Concen: 0.83 ng
RT: 17.212 min Scan# 1419
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:178 Resp: 342442
Ion Ratio Lower Upper
178 100
176 18.9 15.3 22.9
179 15.4 12.4 18.6



#26
Anthracene
Concen: 0.79 ng
RT: 17.298 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43

Tgt Ion:178 Resp: 314914
Ion Ratio Lower Upper
178 100
176 18.4 14.6 22.0
179 15.3 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.76 ng
RT: 19.212 min Scan# 1726
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:212 Resp: 323365
Ion Ratio Lower Upper
212 100
106 13.3 10.6 16.0
104 7.4 5.9 8.9



#28
Fluoranthene
Concen: 0.80 ng
RT: 19.237 min Scan# 1731
Delta R.T. -0.005 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43

Tgt Ion:202 Resp: 377803
Ion Ratio Lower Upper
202 100
101 11.1 9.0 13.6
203 17.3 13.8 20.8





#29

Chrysene-d12

Concen: 0.40 ng

RT: 21.376 min Scan# 2015

Delta R.T. 0.000 min

Lab File: BN015724.D

Acq: 30 Jul 2021 18:43

Tgt Ion:240 Resp: 145057

Ion Ratio Lower Upper

240 100

120 7.8 7.4 11.0





236 25.7 20.9 31.3

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8



#30

Pyrene

Concen: 0.76 ng

RT: 19.604 min Scan# 1805

Delta R.T. 0.000 min

Lab File: BN015724.D

Acq: 30 Jul 2021 18:43




Tgt Ion:202 Resp: 376198


Ion Ratio Lower Upper

202 100

200 20.8 16.5 24.7

203 17.5 14.0 21.0









#31
 Terphenyl-d14
 Concen: 0.73 ng
 RT: 19.818 min Scan# 1848
 Delta R.T. 0.000 min
 Lab File: BN015724.D
 Acq: 30 Jul 2021 18:43




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:244 Resp: 269317
 Ion Ratio Lower Upper
 244 100
 212 7.3 5.9 8.9
 122 11.3 9.6 14.4



#32
 Benzo(a)anthracene
 Concen: 0.75 ng
 RT: 21.355 min Scan# 2013
 Delta R.T. 0.000 min
 Lab File: BN015724.D
 Acq: 30 Jul 2021 18:43

Tgt Ion:228 Resp: 368664
 Ion Ratio Lower Upper
 228 100
 226 27.4 21.9 32.9
 229 19.1 15.4 23.0





#33

Chrysene

Concen: 0.79 ng

RT: 21.408 min Scan# 2018

Delta R.T. 0.000 min

Lab File: BN015724.D

Acq: 30 Jul 2021 18:43

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.8





Tgt Ion:228 Resp: 371303

Ion Ratio Lower Upper

228 100

226 29.9 24.3 36.5

229 19.1 15.3 22.9



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.64 ng

RT: 21.302 min Scan# 2008

Delta R.T. -0.010 min

Lab File: BN015724.D

Acq: 30 Jul 2021 18:43




Tgt Ion:149 Resp: 181000


Ion Ratio Lower Upper

149 100

167 26.6 21.5 32.3

279 4.0 3.2 4.8









#35
 Perylene-d12
 Concen: 0.40 ng
 RT: 23.717 min Scan# 2495
 Delta R.T. -0.003 min
 Lab File: BN015724.D
 Acq: 30 Jul 2021 18:43




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:264 Resp: 144540
 Ion Ratio Lower Upper
 264 100
 260 24.6 19.4 29.2
 265 25.0 19.9 29.9



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.80 ng
 RT: 26.127 min Scan# 3199
 Delta R.T. 0.000 min
 Lab File: BN015724.D
 Acq: 30 Jul 2021 18:43

Tgt Ion:276 Resp: 423473
 Ion Ratio Lower Upper
 276 100
 138 25.0 20.6 30.8
 277 25.1 20.1 30.1









#37
Benzo(b)fluoranthene
Concen: 0.77 ng
RT: 23.009 min Scan# 2288
Delta R.T. -0.003 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8


Tgt Ion:252 Resp: 377114
Ion Ratio Lower Upper
252 100
253 22.1 18.1 27.1
125 10.2 8.3 12.5



#38
Benzo(k)fluoranthene
Concen: 0.84 ng
RT: 23.057 min Scan# 2302
Delta R.T. 0.000 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43


Tgt Ion:252 Resp: 406038
Ion Ratio Lower Upper
252 100
253 22.2 18.1 27.1
125 11.2 10.4 15.6








#39
Benzo(a)pyrene
Concen: 0.78 ng
RT: 23.615 min Scan# 2465
Delta R.T. -0.003 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8





Tgt Ion:252 Resp: 343696
Ion Ratio Lower Upper
252 100
253 22.1 18.1 27.1
125 11.2 9.2 13.8




#40
Dibenzo(a,h)anthracene
Concen: 0.81 ng
RT: 26.144 min Scan# 3204
Delta R.T. -0.003 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43




Tgt Ion:278 Resp: 358543
Ion Ratio Lower Upper
278 100
139 17.6 16.0 24.0
279 23.7 19.2 28.8





#41
Benzo(g,h,i)perylene
Concen: 0.81 ng
RT: 26.856 min Scan# 3412
Delta R.T. -0.003 min
Lab File: BN015724.D
Acq: 30 Jul 2021 18:43

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.8



Tgt Ion: 276 Resp: 362852
Ion Ratio Lower Upper
276 100
277 23.8 19.2 28.8
138 21.0 16.8 25.2

