

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN073021\
 Data File : BN015728.D
 Acq On : 30 Jul 2021 21:10
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN073021

Quant Time: Jul 31 04:56:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 31 04:53:33 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.790	152	38780	0.40	ng	0.00
7) Naphthalene-d8	10.571	136	173981	0.40	ng	0.00
13) Acenaphthene-d10	14.421	164	94102	0.40	ng	0.00
19) Phenanthrene-d10	17.163	188	179334	0.40	ng	0.00
29) Chrysene-d12	21.376	240	169261	0.40	ng	# 0.00
35) Perylene-d12	23.717	264	172191	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.383	112	40724	0.39	ng	0.00
5) Phenol-d6	6.951	99	48493	0.38	ng	0.00
8) Nitrobenzene-d5	8.936	82	41029	0.38	ng	0.00
11) 2-Methylnaphthalene-d10	12.167	152	103208	0.39	ng	0.00
14) 2,4,6-Tribromophenol	15.918	330	13625	0.34	ng	0.00
15) 2-Fluorobiphenyl	13.050	172	133046	0.37	ng	0.00
27) Fluoranthene-d10	19.207	212	178910	0.38	ng	0.00
31) Terphenyl-d14	19.817	244	149757	0.36	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.355	88	6073	0.46	ng	# 85
3) n-Nitrosodimethylamine	3.714	42	9003	0.35	ng	96
6) bis(2-Chloroethyl)ether	7.218	93	41245	0.40	ng	99
9) Naphthalene	10.622	128	176387	0.39	ng	100
10) Hexachlorobutadiene	10.913	225	31294	0.39	ng	# 100
12) 2-Methylnaphthalene	12.238	142	115730	0.39	ng	99
16) Acenaphthylene	14.142	152	151528	0.37	ng	100
17) Acenaphthene	14.484	154	100722	0.37	ng	100
18) Fluorene	15.474	166	123538	0.38	ng	100
20) 4,6-Dinitro-2-methylph...	15.550	198	4243	0.38	ng	99
21) 4-Bromophenyl-phenylether	16.360	248	37595	0.36	ng	# 65
22) Hexachlorobenzene	16.482	284	43905	0.38	ng	99
23) Atrazine	16.640	200	27265	0.34	ng	98
24) Pentachlorophenol	16.823	266	13703	0.32	ng	100
25) Phenanthrene	17.212	178	191050	0.37	ng	100
26) Anthracene	17.297	178	167276	0.36	ng	100
28) Fluoranthene	19.236	202	209321	0.38	ng	100
30) Pyrene	19.604	202	208568	0.37	ng	100
32) Benzo(a)anthracene	21.355	228	201053	0.37	ng	100
33) Chrysene	21.408	228	203440	0.37	ng	99
34) Bis(2-ethylhexyl)phtha...	21.302	149	86101	0.40	ng	100
36) Indeno(1,2,3-cd)pyrene	26.120	276	231940	0.37	ng	100
37) Benzo(b)fluoranthene	23.008	252	214629	0.39	ng	100
38) Benzo(k)fluoranthene	23.056	252	222205	0.37	ng	99
39) Benzo(a)pyrene	23.611	252	187216	0.37	ng	99
40) Dibenzo(a,h)anthracene	26.140	278	196890	0.38	ng	99
41) Benzo(g,h,i)perylene	26.852	276	197232	0.37	ng	99


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN073021\
Data File : BN015728.D
Acq On : 30 Jul 2021 21:10
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN073021

Quant Time: Jul 31 04:56:53 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Jul 31 04:53:33 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.40 ng
 RT: 7.790 min Scan# 511
 Delta R.T. -0.000 min
 Lab File: BN015728.D
 Acq: 30 Jul 2021 21:10




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN073021


Tgt Ion:152 Resp: 38780
 Ion Ratio Lower Upper
 152 100
 150 156.7 123.9 185.9
 115 59.4 47.8 71.6



#2
 1,4-Dioxane
 Concen: 0.46 ng
 RT: 3.355 min Scan# 30
 Delta R.T. -0.000 min
 Lab File: BN015728.D
 Acq: 30 Jul 2021 21:10

Tgt Ion: 88 Resp: 6073
 Ion Ratio Lower Upper
 88 100
 43 22.3 17.2 25.8
 58 61.9 62.2 93.2#









#3
n-Nitrosodimethylamine
Concen: 0.35 ng
RT: 3.714 min Scan# 69
Delta R.T. 0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10




Instrument :
BNA_N
ClientSampleId :
ICVBN073021


Tgt Ion:	42	Resp:	9003
Ion	Ratio	Lower	Upper
42	100		
74	151.4	116.6	174.8
44	7.9	6.6	10.0



#4
2-Fluorophenol
Concen: 0.39 ng
RT: 5.383 min Scan# 250
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Tgt Ion:	112	Resp:	40724
Ion	Ratio	Lower	Upper
112	100		
64	51.8	42.1	63.1
63	26.8	21.9	32.9









#5
Phenol-d6
Concen: 0.38 ng
RT: 6.951 min Scan# 420
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10




Instrument :
BNA_N
ClientSampleId :
ICVBN073021


Tgt Ion:	99	Resp:	48493
Ion	Ratio	Lower	Upper
99	100		
42	15.6	12.9	19.3
71	29.4	23.8	35.6



#6
bis(2-Chloroethyl)ether
Concen: 0.40 ng
RT: 7.218 min Scan# 449
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Tgt Ion:	93	Resp:	41245
Ion	Ratio	Lower	Upper
93	100		
63	72.4	58.6	87.8
95	33.1	26.9	40.3









#7
Naphthalene-d8
Concen: 0.40 ng
RT: 10.571 min Scan# 746
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10




Instrument :
BNA_N
ClientSampleId :
ICVBN073021


Tgt Ion:	136	Resp:	173981
Ion Ratio	Lower	Upper	
136	100		
137	10.9	8.7	13.1
54	6.5	5.5	8.3
68	5.3	4.5	6.7



#8
Nitrobenzene-d5
Concen: 0.38 ng
RT: 8.936 min Scan# 617
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Tgt Ion:	82	Resp:	41029
Ion Ratio	Lower	Upper	
82	100		
128	43.2	33.2	49.8
54	47.6	38.4	57.6









#9
Naphthalene
Concen: 0.39 ng
RT: 10.622 min Scan# 750
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10




Instrument :
BNA_N
ClientSampleId :
ICVBN073021


Tgt Ion:128 Resp: 176387
Ion Ratio Lower Upper
128 100
129 11.0 8.9 13.3
127 13.0 10.4 15.6



#10
Hexachlorobutadiene
Concen: 0.39 ng
RT: 10.913 min Scan# 773
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Tgt Ion:225 Resp: 31294
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 63.8 51.0 76.4









#11
2-Methylnaphthalene-d10
Concen: 0.39 ng
RT: 12.167 min Scan# 968
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10




Instrument :
BNA_N
ClientSampleId :
ICVBN073021


Tgt Ion:152 Resp: 103208
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.39 ng
RT: 12.238 min Scan# 984
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Tgt Ion:142 Resp: 115730
Ion Ratio Lower Upper
142 100
141 89.6 70.5 105.7
115 34.5 27.3 40.9









#13
Acenaphthene-d10
Concen: 0.40 ng
RT: 14.421 min Scan# 1195
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10




Instrument :
BNA_N
ClientSampleId :
ICVBN073021


Tgt Ion:164	Resp:	94102
Ion Ratio	Lower	Upper
164	100	
162	99.3	79.8 119.6
160	44.4	35.8 53.6



#14
2,4,6-Tribromophenol
Concen: 0.34 ng
RT: 15.918 min Scan# 1313
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Tgt Ion:330	Resp:	13625
Ion Ratio	Lower	Upper
330	100	
332	0.0	0.0 0.0
141	34.5	28.8 43.2









#15
2-Fluorobiphenyl
Concen: 0.37 ng
RT: 13.050 min Scan# 1087
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10




Instrument :
BNA_N
ClientSampleId :
ICVBN073021


Tgt Ion:172	Resp:	133046
Ion Ratio	Lower	Upper
172	100	
171	33.3	26.9 40.3
170	21.5	17.4 26.0



#16
Acenaphthylene
Concen: 0.37 ng
RT: 14.142 min Scan# 1173
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Tgt Ion:152	Resp:	151528
Ion Ratio	Lower	Upper
152	100	
151	19.5	15.7 23.5
153	13.1	10.4 15.6









#17
Acenaphthene
Concen: 0.37 ng
RT: 14.484 min Scan# 1200
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10




Instrument :
BNA_N
ClientSampleId :
ICVBN073021


Tgt Ion:154 Resp: 100722
Ion Ratio Lower Upper
154 100
153 111.9 89.7 134.5
152 54.9 43.8 65.6



#18
Fluorene
Concen: 0.38 ng
RT: 15.474 min Scan# 1278
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Tgt Ion:166 Resp: 123538
Ion Ratio Lower Upper
166 100
165 97.8 78.1 117.1
167 13.6 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.40 ng
RT: 17.163 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10




Instrument :
BNA_N
ClientSampleId :
ICVBN073021


Tgt Ion:188 Resp: 179334
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 11.9 10.2 15.4



#20
4,6-Dinitro-2-methylphenol
Concen: 0.38 ng
RT: 15.550 min Scan# 1284
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Tgt Ion:198 Resp: 4243
Ion Ratio Lower Upper
198 100
182 11.9 9.4 14.0
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.36 ng
RT: 16.360 min Scan# 1349
Delta R.T. -0.012 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10




Instrument :
BNA_N
ClientSampleId :
ICVBN073021


Tgt Ion:248 Resp: 37595
Ion Ratio Lower Upper
248 100
250 90.7 83.9 125.9
141 91.0 26.9 40.3#



#22
Hexachlorobenzene
Concen: 0.38 ng
RT: 16.482 min Scan# 1359
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Tgt Ion:284 Resp: 43905
Ion Ratio Lower Upper
284 100
142 35.8 29.4 44.0
249 29.3 23.8 35.6









#23
Atrazine
Concen: 0.34 ng
RT: 16.640 min Scan# 1372
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10




Instrument :
BNA_N
ClientSampleId :
ICVBN073021


Tgt Ion:200 Resp: 27265
Ion Ratio Lower Upper
200 100
173 23.2 18.9 28.3
215 49.2 38.1 57.1



#24
Pentachlorophenol
Concen: 0.32 ng
RT: 16.823 min Scan# 1387
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Tgt Ion:266 Resp: 13703
Ion Ratio Lower Upper
266 100
264 62.8 50.2 75.2
268 63.9 51.1 76.7









#25
Phenanthrene
Concen: 0.37 ng
RT: 17.212 min Scan# 1419
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10




Instrument :
BNA_N
ClientSampleId :
ICVBN073021


Tgt Ion:178 Resp: 191050
Ion Ratio Lower Upper
178 100
176 19.0 15.3 22.9
179 15.4 12.4 18.6



#26
Anthracene
Concen: 0.36 ng
RT: 17.297 min Scan# 1426
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Tgt Ion:178 Resp: 167276
Ion Ratio Lower Upper
178 100
176 18.3 14.6 22.0
179 15.3 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.38 ng
RT: 19.207 min Scan# 1725
Delta R.T. -0.005 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10




Instrument :
BNA_N
ClientSampleId :
ICVBN073021


Tgt Ion:212 Resp: 178910
Ion Ratio Lower Upper
212 100
106 13.1 10.6 16.0
104 7.3 5.9 8.9



#28
Fluoranthene
Concen: 0.38 ng
RT: 19.236 min Scan# 1731
Delta R.T. -0.005 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Tgt Ion:202 Resp: 209321
Ion Ratio Lower Upper
202 100
101 11.2 9.0 13.6
203 17.3 13.8 20.8









#29
Chrysene-d12
Concen: 0.40 ng
RT: 21.376 min Scan# 2015
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10




Instrument :
BNA_N
ClientSampleId :
ICVBN073021


Tgt Ion:240 Resp: 169261
Ion Ratio Lower Upper
240 100
120 7.2 7.4 11.0#
236 25.6 20.9 31.3



#30
Pyrene
Concen: 0.37 ng
RT: 19.604 min Scan# 1805
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Tgt Ion:202 Resp: 208568
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 17.4 14.0 21.0









#31
 Terphenyl-d14
 Concen: 0.36 ng
 RT: 19.817 min Scan# 1848
 Delta R.T. -0.000 min
 Lab File: BN015728.D
 Acq: 30 Jul 2021 21:10




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN073021


Tgt Ion:244	Resp:	149757
Ion Ratio	Lower	Upper
244	100	
212	7.3	5.9 8.9
122	11.2	9.6 14.4



#32
 Benzo(a)anthracene
 Concen: 0.37 ng
 RT: 21.355 min Scan# 2013
 Delta R.T. -0.000 min
 Lab File: BN015728.D
 Acq: 30 Jul 2021 21:10

Tgt Ion:228	Resp:	201053
Ion Ratio	Lower	Upper
228	100	
226	27.1	21.9 32.9
229	19.3	15.4 23.0









#33
Chrysene
Concen: 0.37 ng
RT: 21.408 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10




Instrument :
BNA_N
ClientSampleId :
ICVBN073021


Tgt Ion:228 Resp: 203440
Ion Ratio Lower Upper
228 100
226 30.0 24.3 36.5
229 19.4 15.3 22.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.40 ng
RT: 21.302 min Scan# 2008
Delta R.T. -0.011 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10


Tgt Ion:149 Resp: 86101
Ion Ratio Lower Upper
149 100
167 26.8 21.5 32.3
279 4.0 3.2 4.8








#35
Perylene-d12
Concen: 0.40 ng
RT: 23.717 min Scan# 2495
Delta R.T. -0.004 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Instrument :
BNA_N
ClientSampleId :
ICVBN073021







Tgt Ion:264 Resp: 172191
Ion Ratio Lower Upper
264 100
260 24.5 19.4 29.2
265 24.3 19.9 29.9



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.37 ng
RT: 26.120 min Scan# 3197
Delta R.T. -0.007 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Tgt Ion:276 Resp: 231940
Ion Ratio Lower Upper
276 100
138 25.3 20.6 30.8
277 25.1 20.1 30.1









#37
Benzo(b)fluoranthene
Concen: 0.39 ng
RT: 23.008 min Scan# 2288
Delta R.T. -0.004 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10




Instrument :
BNA_N
ClientSampleId :
ICVBN073021


Tgt Ion:252 Resp: 214629
Ion Ratio Lower Upper
252 100
253 22.4 18.1 27.1
125 10.3 8.3 12.5



#38
Benzo(k)fluoranthene
Concen: 0.37 ng
RT: 23.056 min Scan# 2302
Delta R.T. -0.000 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Tgt Ion:252 Resp: 222205
Ion Ratio Lower Upper
252 100
253 22.7 18.1 27.1
125 12.2 10.4 15.6





#39

Benzo(a)pyrene

Concen: 0.37 ng

RT: 23.611 min Scan# 2464

Delta R.T. -0.007 min

Lab File: BN015728.D

Acq: 30 Jul 2021 21:10





Tgt Ion:252 Resp: 187216

Ion Ratio Lower Upper

252 100

253 22.1 18.1 27.1

125 11.3 9.2 13.8



#40

Dibenzo(a,h)anthracene

Concen: 0.38 ng

RT: 26.140 min Scan# 3203

Delta R.T. -0.007 min

Lab File: BN015728.D

Acq: 30 Jul 2021 21:10




Tgt Ion:278 Resp: 196890


Ion Ratio Lower Upper

278 100

139 19.1 16.0 24.0

279 24.1 19.2 28.8





#41
Benzo(g,h,i)perylene
Concen: 0.37 ng
RT: 26.852 min Scan# 3411
Delta R.T. -0.007 min
Lab File: BN015728.D
Acq: 30 Jul 2021 21:10

Instrument :
BNA_N
ClientSampleId :
ICVBN073021

Tgt Ion:276 Resp: 197232
Ion Ratio Lower Upper
276 100
277 23.7 19.2 28.8
138 20.5 16.8 25.2

