

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN073021\
 Data File : BN015728.D
 Acq On : 30 Jul 2021 21:10
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN073021

Quant Time: Jul 31 04:56:53 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Sat Jul 31 04:53:33 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	110	0.00
2	1,4-Dioxane	0.400	0.389	2.8	102	0.00
3	n-Nitrosodimethylamine	0.400	0.349	12.8	110	0.00
4 S	2-Fluorophenol	0.400	0.391	2.3	111	0.00
5 S	Phenol-d6	0.400	0.385	3.8	110	0.00
6	bis(2-Chloroethyl)ether	0.400	0.400	0.0	110	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	110	0.00
8 S	Nitrobenzene-d5	0.400	0.375	6.3	121	0.00
9	Naphthalene	0.400	0.389	2.8	110	0.00
10	Hexachlorobutadiene	0.400	0.392	2.0	112	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.394	1.5	112	0.00
12	2-Methylnaphthalene	0.400	0.393	1.8	111	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.400	0.335	16.3	114	0.00
15 S	2-Fluorobiphenyl	0.400	0.369	7.8	112	0.00
16	Acenaphthylene	0.400	0.367	8.3	112	0.00
17	Acenaphthene	0.400	0.374	6.5	112	0.00
18	Fluorene	0.400	0.377	5.8	113	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	114	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.384	4.0	127	0.00
21	4-Bromophenyl-phenylether	0.400	0.359	10.3	111	-0.01
22	Hexachlorobenzene	0.400	0.379	5.3	115	0.00
23	Atrazine	0.400	0.341	14.8	114	0.00
24	Pentachlorophenol	0.400	0.316	21.0	114	0.00
25	Phenanthrene	0.400	0.372	7.0	114	0.00
26	Anthracene	0.400	0.360	10.0	113	0.00
27 SURR	Fluoranthene-d10	0.400	0.379	5.3	116	0.00
28	Fluoranthene	0.400	0.381	4.8	115	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	116	0.00
30	Pyrene	0.400	0.369	7.8	115	0.00
31 S	Terphenyl-d14	0.400	0.364	9.0	116	0.00
32	Benzo(a)anthracene	0.400	0.373	6.8	118	0.00
33	Chrysene	0.400	0.371	7.3	115	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.405	-1.3	115	-0.01
35 I	Perylene-d12	0.400	0.400	0.0	117	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.375	6.3	119	0.00
37	Benzo(b)fluoranthene	0.400	0.388	3.0	120	0.00
38	Benzo(k)fluoranthene	0.400	0.372	7.0	117	0.00
39 C	Benzo(a)pyrene	0.400	0.370	7.5	117	0.00
40	Dibenzo(a,h)anthracene	0.400	0.376	6.0	119	0.00
41	Benzo(g,h,i)perylene	0.400	0.366	8.5	116	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN073021