

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN073021\
 Data File : BN015721.D
 Acq On : 30 Jul 2021 16:52
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Manual Integrations
 APPROVED

mohammad
 8/2/2021 3:47:23 PM

Quant Time: Jul 30 18:46:40 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 30 18:43:48 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.790	152	30776	0.400	ng	0.00
7) Naphthalene-d8	10.571	136	137038	0.400	ng	0.00
13) Acenaphthene-d10	14.421	164	71047	0.400	ng	0.00
19) Phenanthrene-d10	17.164	188	134703	0.400	ng	0.00
29) Chrysene-d12	21.376	240	122225	0.400	ng	0.00
35) Perylene-d12	23.721	264	122721	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.383	112	9825	0.119	ng	0.00
5) Phenol-d6	6.951	99	12186	0.116	ng	0.00
8) Nitrobenzene-d5	8.937	82	8927	0.097	ng	0.00
11) 2-Methylnaphthalene-d10	12.167	152	21036	0.098	ng	0.00
14) 2,4,6-Tribromophenol	15.918	330	3025	0.105	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	32991	0.122	ng	0.00
27) Fluoranthene-d10	19.212	212	34805	0.091	ng	0.00
31) Terphenyl-d14	19.817	244	35589	0.114	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.355	88	800m	0.074	ng	
6) bis(2-Chloroethyl)ether	7.218	93	8989	0.109	ng	99
9) Naphthalene	10.622	128	39118	0.110	ng	100
10) Hexachlorobutadiene	10.914	225	6807	0.099	ng	# 100
12) 2-Methylnaphthalene	12.238	142	25474	0.106	ng	99
16) Acenaphthylene	14.129	152	32259	0.100	ng	100
17) Acenaphthene	14.484	154	21741	0.108	ng	99
18) Fluorene	15.474	166	26501	0.105	ng	100
20) 4,6-Dinitro-2-methylph...	15.550	198	612	0.061	ng	# 68
21) 4-Bromophenyl-phenylether	16.360	248	8231	0.103	ng	# 64
22) Hexachlorobenzene	16.482	284	9445	0.121	ng	99
23) Atrazine	16.640	200	5455	0.076	ng	98
24) Pentachlorophenol	16.823	266	2726	0.095	ng	99
25) Phenanthrene	17.212	178	42624	0.115	ng	99
26) Anthracene	17.297	178	36138	0.101	ng	99
28) Fluoranthene	19.242	202	44372	0.104	ng	100
30) Pyrene	19.604	202	45686	0.109	ng	100
32) Benzo(a)anthracene	21.366	228	40656	0.098	ng	97
33) Chrysene	21.419	228	44637	0.113	ng	97
36) Indeno(1,2,3-cd)pyrene	26.127	276	44244	0.099	ng	99
37) Benzo(b)fluoranthene	23.012	252	41179	0.099	ng	# 95
38) Benzo(k)fluoranthene	23.060	252	48591m	0.119	ng	
39) Benzo(a)pyrene	23.618	252	37892	0.101	ng	# 93
40) Dibenzo(a,h)anthracene	26.147	278	37139	0.098	ng	# 93
41) Benzo(g,h,i)perylene	26.856	276	41495	0.109	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN073021\
Data File : BN015721.D
Acq On : 30 Jul 2021 16:52
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Manual Integrations
APPROVED

mohammad
8/2/2021 3:47:23 PM


Quant Time: Jul 30 18:46:40 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 30 18:43:48 2021
Response via : Initial Calibration





#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.790 min Scan# 511
Delta R.T. -0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52



Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Tgt Ion:152 Resp: 30776
Ion Ratio Lower Upper
152 100
150 155.2 123.9 185.9
115 59.7 47.8 71.6

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:23 PM




#2
1,4-Dioxane
Concen: 0.074 ng m
RT: 3.355 min Scan# 30
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Tgt Ion: 88 Resp: 800
Ion Ratio Lower Upper
88 100
43 43.4 17.2 25.8#
58 106.9 62.2 93.2#





#4
2-Fluorophenol
Concen: 0.119 ng
RT: 5.383 min Scan# 250
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion: 112 Resp: 9825
Ion Ratio Lower Upper
112 100
64 53.2 42.1 63.1
63 28.8 21.9 32.9

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:23 PM




#5
Phenol-d6
Concen: 0.116 ng
RT: 6.951 min Scan# 420
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52



Tgt Ion: 99 Resp: 12186
Ion Ratio Lower Upper
99 100
42 16.5 12.9 19.3
71 30.2 23.8 35.6









#6
bis(2-Chloroethyl)ether
Concen: 0.109 ng
RT: 7.218 min Scan# 449
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Tgt Ion:	93	Resp:	8989
Ion	Ratio	Lower	Upper
93	100		
63	74.3	58.6	87.8
95	33.5	26.9	40.3

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
8/2/2021 3:47:23 PM



#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.571 min Scan# 746
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Tgt Ion:	136	Resp:	137038
Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	6.4	5.5	8.3
68	5.1	4.5	6.7





#8
Nitrobenzene-d5
Concen: 0.097 ng
RT: 8.937 min Scan# 617
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion: 82 Resp: 8927
Ion Ratio Lower Upper
82 100
128 42.2 33.2 49.8
54 49.9 38.4 57.6

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:23 PM




#9
Naphthalene
Concen: 0.110 ng
RT: 10.622 min Scan# 750
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52



Tgt Ion: 128 Resp: 39118
Ion Ratio Lower Upper
128 100
129 11.3 8.9 13.3
127 13.0 10.4 15.6









#10
Hexachlorobutadiene
Concen: 0.099 ng
RT: 10.914 min Scan# 773
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Tgt Ion: 225 Resp: 6807
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 51.0 76.4

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
8/2/2021 3:47:23 PM



#11
2-Methylnaphthalene-d10
Concen: 0.098 ng
RT: 12.167 min Scan# 968
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Tgt Ion: 152 Resp: 21036
Ion Ratio Lower Upper
152 100
151 20.5 16.2 24.4





#12
2-Methylnaphthalene
Concen: 0.106 ng
RT: 12.238 min Scan# 984
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:142 Resp: 25474
Ion Ratio Lower Upper
142 100
141 88.5 70.5 105.7
115 35.0 27.3 40.9

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:23 PM




#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.421 min Scan# 1195
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Tgt Ion:164 Resp: 71047
Ion Ratio Lower Upper
164 100
162 99.3 79.8 119.6
160 44.3 35.8 53.6





#14
2,4,6-Tribromophenol
Concen: 0.105 ng
RT: 15.918 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52



Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Tgt Ion:330 Resp: 3025
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 36.1 28.8 43.2

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:23 PM




#15
2-Fluorobiphenyl
Concen: 0.122 ng
RT: 13.051 min Scan# 1087
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52



Tgt Ion:172 Resp: 32991
Ion Ratio Lower Upper
172 100
171 33.4 26.9 40.3
170 21.8 17.4 26.0









#16
Acenaphthylene
Concen: 0.100 ng
RT: 14.129 min Scan# 1172
Delta R.T. -0.012 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Tgt Ion:152 Resp: 32259
Ion Ratio Lower Upper
152 100
151 19.7 15.7 23.5
153 13.3 10.4 15.6

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
8/2/2021 3:47:23 PM



#17
Acenaphthene
Concen: 0.108 ng
RT: 14.484 min Scan# 1200
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Tgt Ion:154 Resp: 21741
Ion Ratio Lower Upper
154 100
153 113.4 89.7 134.5
152 56.2 43.8 65.6





#18
Fluorene
Concen: 0.105 ng
RT: 15.474 min Scan# 1278
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:166 Resp: 26501
Ion Ratio Lower Upper
166 100
165 97.8 78.1 117.1
167 13.3 10.8 16.2

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:23 PM




#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.164 min Scan# 1415
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Tgt Ion:188 Resp: 134703
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.1 10.2 15.4





#20
4,6-Dinitro-2-methylphenol
Concen: 0.061 ng
RT: 15.550 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:198 Resp: 612
Ion Ratio Lower Upper
198 100
182 24.0 9.4 14.0#
77 0.0 0.0 0.0

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:23 PM




#21
4-Bromophenyl-phenylether
Concen: 0.103 ng
RT: 16.360 min Scan# 1349
Delta R.T. -0.012 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Tgt Ion:248 Resp: 8231
Ion Ratio Lower Upper
248 100
250 90.1 83.9 125.9
141 91.8 26.9 40.3#





#22
Hexachlorobenzene
Concen: 0.121 ng
RT: 16.482 min Scan# 1359
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Tgt Ion:284 Resp: 9445
Ion Ratio Lower Upper
284 100
142 37.6 29.4 44.0
249 30.0 23.8 35.6

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:23 PM




#23
Atrazine
Concen: 0.076 ng
RT: 16.640 min Scan# 1372
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Tgt Ion:200 Resp: 5455
Ion Ratio Lower Upper
200 100
173 24.2 18.9 28.3
215 49.3 38.1 57.1





#24
Pentachlorophenol
Concen: 0.095 ng
RT: 16.823 min Scan# 1387
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:266 Resp: 2726
Ion Ratio Lower Upper
266 100
264 62.7 50.2 75.2
268 64.9 51.1 76.7

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:23 PM




#25
Phenanthrene
Concen: 0.115 ng
RT: 17.212 min Scan# 1419
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Tgt Ion:178 Resp: 42624
Ion Ratio Lower Upper
178 100
176 19.0 15.3 22.9
179 16.1 12.4 18.6





#26
 Anthracene
 Concen: 0.101 ng
 RT: 17.297 min Scan# 1426
 Delta R.T. 0.000 min
 Lab File: BN015721.D
 Acq: 30 Jul 2021 16:52

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1





Tgt Ion:178 Resp: 36138


Ion	Ratio	Lower	Upper
178	100		
176	18.6	14.6	22.0
179	15.2	12.2	18.4

Manual Integrations
 APPROVED

mohammad
 8/2/2021 3:47:23 PM






#27
 Fluoranthene-d10
 Concen: 0.091 ng
 RT: 19.212 min Scan# 1726
 Delta R.T. 0.000 min
 Lab File: BN015721.D
 Acq: 30 Jul 2021 16:52



Tgt Ion:212 Resp: 34805


Ion	Ratio	Lower	Upper
212	100		
106	13.5	10.6	16.0
104	7.5	5.9	8.9





#28
Fluoranthene
Concen: 0.104 ng
RT: 19.242 min Scan# 1732
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:202 Resp: 44372
Ion Ratio Lower Upper
202 100
101 11.5 9.0 13.6
203 17.2 13.8 20.8

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:23 PM




#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.376 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52



Tgt Ion:240 Resp: 122225
Ion Ratio Lower Upper
240 100
120 10.2 7.4 11.0
236 26.4 20.9 31.3









#30
Pyrene
Concen: 0.109 ng
RT: 19.604 min Scan# 1805
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52

Tgt Ion:	202	Resp:	45686
Ion Ratio	Lower	Upper	
202	100		
200	20.7	16.5	24.7
203	17.6	14.0	21.0

Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Manual Integrations
APPROVED


mohammad
8/2/2021 3:47:23 PM



#31
Terphenyl-d14
Concen: 0.114 ng
RT: 19.817 min Scan# 1848
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52


Tgt Ion:	244	Resp:	35589
Ion Ratio	Lower	Upper	
244	100		
212	7.8	5.9	8.9
122	12.6	9.6	14.4





#32
Benzo(a)anthracene
Concen: 0.098 ng
RT: 21.366 min Scan# 2014
Delta R.T. 0.011 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1




Tgt Ion:228 Resp: 40656
Ion Ratio Lower Upper
228 100
226 25.8 21.9 32.9
229 20.3 15.4 23.0

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:23 PM




#33
Chrysene
Concen: 0.113 ng
RT: 21.419 min Scan# 2019
Delta R.T. 0.011 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Tgt Ion:228 Resp: 44637
Ion Ratio Lower Upper
228 100
226 28.4 24.3 36.5
229 20.2 15.3 22.9





#35
Perylene-d12
Concen: 0.400 ng
RT: 23.721 min Scan# 2496
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1




Tgt Ion:264 Resp: 122721
Ion Ratio Lower Upper
264 100
260 24.7 19.4 29.2
265 26.2 19.9 29.9

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:23 PM




#36
Indeno(1,2,3-cd)pyrene
Concen: 0.099 ng
RT: 26.127 min Scan# 3199
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Tgt Ion:276 Resp: 44244
Ion Ratio Lower Upper
276 100
138 26.8 20.6 30.8
277 25.2 20.1 30.1






#37
Benzo(b)fluoranthene
Concen: 0.099 ng
RT: 23.012 min Scan# 2289
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Tgt Ion:252 Resp: 41179
Ion Ratio Lower Upper
252 100
253 24.8 18.1 27.1
125 12.7 8.3 12.5#




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:23 PM




#38
Benzo(k)fluoranthene
Concen: 0.119 ng m
RT: 23.060 min Scan# 2303
Delta R.T. 0.004 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Tgt Ion:252 Resp: 48591
Ion Ratio Lower Upper
252 100
253 25.5 18.1 27.1
125 17.4 10.4 15.6#







#39
Benzo(a)pyrene
Concen: 0.101 ng
RT: 23.618 min Scan# 2466
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Tgt Ion:252 Resp: 37892
Ion Ratio Lower Upper
252 100
253 25.9 18.1 27.1
125 14.6 9.2 13.8#




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:23 PM




#40
Dibenzo(a,h)anthracene
Concen: 0.098 ng
RT: 26.147 min Scan# 3205
Delta R.T. 0.000 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52




Tgt Ion:278 Resp: 37139
Ion Ratio Lower Upper
278 100
139 25.0 16.0 24.0#
279 25.9 19.2 28.8







#41
Benzo(g,h,i)perylene
Concen: 0.109 ng
RT: 26.856 min Scan# 3412
Delta R.T. -0.003 min
Lab File: BN015721.D
Acq: 30 Jul 2021 16:52



Tgt Ion:	276	Resp:	41495
Ion Ratio	Lower	Upper	
276	100		
277	24.4	19.2	28.8
138	22.8	16.8	25.2



Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Manual Integrations
APPROVED

mohammad
8/2/2021 3:47:23 PM