

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN073021\
 Data File : BN015726.D
 Acq On : 30 Jul 2021 19:56
 Operator : CG/JU
 Sample : SSTDICC3.2
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC3.2

Quant Time: Jul 31 00:32:41 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 30 18:55:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.790	152	39089	0.400	ng	0.00
7) Naphthalene-d8	10.572	136	162589	0.400	ng	0.00
13) Acenaphthene-d10	14.421	164	87294	0.400	ng	0.00
19) Phenanthrene-d10	17.164	188	164351	0.400	ng	0.00
29) Chrysene-d12	21.376	240	156002	0.400	ng	0.00
35) Perylene-d12	23.717	264	162887	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.383	112	320892	3.053	ng	0.00
5) Phenol-d6	6.951	99	389516	2.929	ng	0.00
8) Nitrobenzene-d5	8.937	82	357912	3.288	ng	0.00
11) 2-Methylnaphthalene-d10	12.167	152	785995	3.073	ng	0.00
14) 2,4,6-Tribromophenol	15.918	330	140361	3.964	ng	0.00
15) 2-Fluorobiphenyl	13.051	172	1024166	3.080	ng	0.00
27) Fluoranthene-d10	19.207	212	1450896	3.121	ng	0.00
31) Terphenyl-d14	19.818	244	1189043	2.997	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.346	88	45069	3.297	ng	# 94
3) n-Nitrosodimethylamine	3.696	42	94807	3.535	ng	# 96
6) bis(2-Chloroethyl)ether	7.218	93	320486	3.069	ng	99
9) Naphthalene	10.622	128	1325979	3.142	ng	100
10) Hexachlorobutadiene	10.914	225	234747	2.875	ng	# 100
12) 2-Methylnaphthalene	12.239	142	854600	2.990	ng	99
16) Acenaphthylene	14.142	152	1297483	3.271	ng	100
17) Acenaphthene	14.485	154	821603	3.327	ng	96
18) Fluorene	15.474	166	993147	3.199	ng	99
20) 4,6-Dinitro-2-methylph...	15.550	198	72636	5.936	ng	# 84
21) 4-Bromophenyl-phenylether	16.360	248	319371	3.269	ng	# 65
22) Hexachlorobenzene	16.482	284	340019	3.568	ng	99
23) Atrazine	16.640	200	271193	3.083	ng	99
24) Pentachlorophenol	16.823	266	157560	4.513	ng	100
25) Phenanthrene	17.212	178	1501675	3.313	ng	100
26) Anthracene	17.298	178	1421729	3.242	ng	100
28) Fluoranthene	19.242	202	1598376	3.084	ng	99
30) Pyrene	19.604	202	1686925	3.163	ng	100
32) Benzo(a)anthracene	21.355	228	1665571	3.145	ng	99
33) Chrysene	21.408	228	1614665	3.210	ng	99
34) Bis(2-ethylhexyl)phtha...	21.302	149	1042479	3.435	ng	99
36) Indeno(1,2,3-cd)pyrene	26.124	276	1956122	3.292	ng	99
37) Benzo(b)fluoranthene	23.009	252	1711855	3.102	ng	99
38) Benzo(k)fluoranthene	23.057	252	1756318	3.229	ng	# 97
39) Benzo(a)pyrene	23.615	252	1584286	3.178	ng	99
40) Dibenzo(a,h)anthracene	26.148	278	1654348	3.305	ng	96
41) Benzo(g,h,i)perylene	26.856	276	1667374	3.302	ng	100


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN073021\
Data File : BN015726.D
Acq On : 30 Jul 2021 19:56
Operator : CG/JU
Sample : SSTDIC3.2
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDIC3.2

Quant Time: Jul 31 00:32:41 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 30 18:55:25 2021
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 0.400 ng
RT: 7.790 min Scan# 511
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 39089
Ion Ratio Lower Upper
152 100
150 156.4 123.9 185.9
115 59.4 47.8 71.6



#2
1,4-Dioxane
Concen: 3.297 ng
RT: 3.346 min Scan# 29
Delta R.T. -0.009 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion: 88 Resp: 45069
Ion Ratio Lower Upper
88 100
43 27.0 17.2 25.8#
58 74.3 62.2 93.2









#3
n-Nitrosodimethylamine
Concen: 3.535 ng
RT: 3.696 min Scan# 67
Delta R.T. -0.018 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 42 Resp: 94807
Ion Ratio Lower Upper
42 100
74 150.7 116.6 174.8
44 6.6 6.6 10.0#



#4
2-Fluorophenol
Concen: 3.053 ng
RT: 5.383 min Scan# 250
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion: 112 Resp: 320892
Ion Ratio Lower Upper
112 100
64 52.0 42.1 63.1
63 26.9 21.9 32.9









#5
Phenol-d6
Concen: 2.929 ng
RT: 6.951 min Scan# 420
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion: 99 Resp: 389516
Ion Ratio Lower Upper
99 100
42 15.5 12.9 19.3
71 29.1 23.8 35.6



#6
bis(2-Chloroethyl)ether
Concen: 3.069 ng
RT: 7.218 min Scan# 449
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion: 93 Resp: 320486
Ion Ratio Lower Upper
93 100
63 71.8 58.6 87.8
95 33.0 26.9 40.3









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.572 min Scan# 746
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:136 Resp: 162589




Ion	Ratio	Lower	Upper
136	100		
137	11.0	8.7	13.1
54	6.6	5.5	8.3
68	5.3	4.5	6.7




#8
Nitrobenzene-d5
Concen: 3.288 ng
RT: 8.937 min Scan# 617
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion: 82 Resp: 357912

Ion	Ratio	Lower	Upper
82	100		
128	45.9	33.2	49.8
54	45.3	38.4	57.6









#9
Naphthalene
Concen: 3.142 ng
RT: 10.622 min Scan# 750
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:128 Resp: 1325979
Ion Ratio Lower Upper
128 100
129 11.1 8.9 13.3
127 12.8 10.4 15.6



#10
Hexachlorobutadiene
Concen: 2.875 ng
RT: 10.914 min Scan# 773
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion:225 Resp: 234747
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.0 51.0 76.4









#11
2-Methylnaphthalene-d10
Concen: 3.073 ng
RT: 12.167 min Scan# 968
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:152 Resp: 785995
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.4



#12
2-Methylnaphthalene
Concen: 2.990 ng
RT: 12.239 min Scan# 984
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion:142 Resp: 854600
Ion Ratio Lower Upper
142 100
141 87.8 70.5 105.7
115 33.4 27.3 40.9









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.421 min Scan# 1195
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:164 Resp: 87294
Ion Ratio Lower Upper
164 100
162 100.6 79.8 119.6
160 44.6 35.8 53.6



#14
2,4,6-Tribromophenol
Concen: 3.964 ng
RT: 15.918 min Scan# 1313
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion:330 Resp: 140361
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 34.8 28.8 43.2









#15
2-Fluorobiphenyl
Concen: 3.080 ng
RT: 13.051 min Scan# 1087
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:172 Resp: 1024166
Ion Ratio Lower Upper
172 100
171 33.0 26.9 40.3
170 21.1 17.4 26.0



#16
Acenaphthylene
Concen: 3.271 ng
RT: 14.142 min Scan# 1173
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion:152 Resp: 1297483
Ion Ratio Lower Upper
152 100
151 19.4 15.7 23.5
153 13.1 10.4 15.6









#17
Acenaphthene
Concen: 3.327 ng
RT: 14.485 min Scan# 1200
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:154 Resp: 821603
Ion Ratio Lower Upper
154 100
153 108.3 89.7 134.5
152 51.5 43.8 65.6



#18
Fluorene
Concen: 3.199 ng
RT: 15.474 min Scan# 1278
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion:166 Resp: 993147
Ion Ratio Lower Upper
166 100
165 96.5 78.1 117.1
167 13.6 10.8 16.2









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.164 min Scan# 1415
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:188 Resp: 164351
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.2 10.2 15.4



#20
4,6-Dinitro-2-methylphenol
Concen: 5.936 ng
RT: 15.550 min Scan# 1284
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion:198 Resp: 72636
Ion Ratio Lower Upper
198 100
182 5.5 9.4 14.0#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 3.269 ng
RT: 16.360 min Scan# 1349
Delta R.T. -0.012 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:248 Resp: 319371
Ion Ratio Lower Upper
248 100
250 89.8 83.9 125.9
141 89.6 26.9 40.3#



#22
Hexachlorobenzene
Concen: 3.568 ng
RT: 16.482 min Scan# 1359
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion:284 Resp: 340019
Ion Ratio Lower Upper
284 100
142 36.0 29.4 44.0
249 29.7 23.8 35.6









#23
Atrazine
Concen: 3.083 ng
RT: 16.640 min Scan# 1372
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:200 Resp: 271193
Ion Ratio Lower Upper
200 100
173 24.6 18.9 28.3
215 48.0 38.1 57.1



#24
Pentachlorophenol
Concen: 4.513 ng
RT: 16.823 min Scan# 1387
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion:266 Resp: 157560
Ion Ratio Lower Upper
266 100
264 62.5 50.2 75.2
268 63.7 51.1 76.7









#25
Phenanthrene
Concen: 3.313 ng
RT: 17.212 min Scan# 1419
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:178 Resp: 1501675
Ion Ratio Lower Upper
178 100
176 18.8 15.3 22.9
179 15.4 12.4 18.6



#26
Anthracene
Concen: 3.242 ng
RT: 17.298 min Scan# 1426
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion:178 Resp: 1421729
Ion Ratio Lower Upper
178 100
176 18.4 14.6 22.0
179 15.5 12.2 18.4









#27
Fluoranthene-d10
Concen: 3.121 ng
RT: 19.207 min Scan# 1725
Delta R.T. -0.005 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:212 Resp: 1450896
Ion Ratio Lower Upper
212 100
106 13.4 10.6 16.0
104 7.4 5.9 8.9



#28
Fluoranthene
Concen: 3.084 ng
RT: 19.242 min Scan# 1732
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion:202 Resp: 1598376
Ion Ratio Lower Upper
202 100
101 11.5 9.0 13.6
203 17.6 13.8 20.8









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.376 min Scan# 2015
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:240 Resp: 156002
Ion Ratio Lower Upper
240 100
120 8.2 7.4 11.0
236 25.7 20.9 31.3



#30
Pyrene
Concen: 3.163 ng
RT: 19.604 min Scan# 1805
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion:202 Resp: 1686925
Ion Ratio Lower Upper
202 100
200 20.7 16.5 24.7
203 17.8 14.0 21.0









#31
Terphenyl-d14
Concen: 2.997 ng
RT: 19.818 min Scan# 1848
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:244 Resp: 1189043
Ion Ratio Lower Upper
244 100
212 7.3 5.9 8.9
122 11.6 9.6 14.4



#32
Benzo(a)anthracene
Concen: 3.145 ng
RT: 21.355 min Scan# 2013
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion:228 Resp: 1665571
Ion Ratio Lower Upper
228 100
226 27.2 21.9 32.9
229 19.8 15.4 23.0









#33
Chrysene
Concen: 3.210 ng
RT: 21.408 min Scan# 2018
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:228 Resp: 1614665
Ion Ratio Lower Upper
228 100
226 30.2 24.3 36.5
229 19.6 15.3 22.9



#34
Bis(2-ethylhexyl)phthalate
Concen: 3.435 ng
RT: 21.302 min Scan# 2008
Delta R.T. -0.010 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion:149 Resp: 1042479
Ion Ratio Lower Upper
149 100
167 27.3 21.5 32.3
279 4.0 3.2 4.8









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.717 min Scan# 2495
Delta R.T. -0.003 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:264 Resp: 162887




Ion	Ratio	Lower	Upper
264	100		
260	24.6	19.4	29.2
265	24.8	19.9	29.9




#36
Indeno(1,2,3-cd)pyrene
Concen: 3.292 ng
RT: 26.124 min Scan# 3198
Delta R.T. -0.003 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion:276 Resp: 1956122

Ion	Ratio	Lower	Upper
276	100		
138	24.9	20.6	30.8
277	25.2	20.1	30.1









#37
Benzo(b)fluoranthene
Concen: 3.102 ng
RT: 23.009 min Scan# 2288
Delta R.T. -0.003 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 1711855
Ion Ratio Lower Upper
252 100
253 22.3 18.1 27.1
125 10.1 8.3 12.5



#38
Benzo(k)fluoranthene
Concen: 3.229 ng
RT: 23.057 min Scan# 2302
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion:252 Resp: 1756318
Ion Ratio Lower Upper
252 100
253 22.2 18.1 27.1
125 10.4 10.4 15.6#









#39
Benzo(a)pyrene
Concen: 3.178 ng
RT: 23.615 min Scan# 2465
Delta R.T. -0.003 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56




Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2


Tgt Ion:252 Resp: 1584286
Ion Ratio Lower Upper
252 100
253 22.3 18.1 27.1
125 11.0 9.2 13.8



#40
Dibenzo(a,h)anthracene
Concen: 3.305 ng
RT: 26.148 min Scan# 3205
Delta R.T. 0.000 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Tgt Ion:278 Resp: 1654348
Ion Ratio Lower Upper
278 100
139 16.2 16.0 24.0
279 23.7 19.2 28.8





#41
Benzo(g,h,i)perylene
Concen: 3.302 ng
RT: 26.856 min Scan# 3412
Delta R.T. -0.003 min
Lab File: BN015726.D
Acq: 30 Jul 2021 19:56

Instrument :
BNA_N
ClientSampleId :
SSTDICC3.2

Tgt Ion:276 Resp: 1667374
Ion Ratio Lower Upper
276 100
277 23.9 19.2 28.8
138 21.0 16.8 25.2

