

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN073021\
 Data File : BN015727.D
 Acq On : 30 Jul 2021 20:33
 Operator : CG/JU
 Sample : SSTDICC5.0
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Manual Integrations
 APPROVED

mohammad
 8/2/2021 3:47:43 PM

Quant Time: Jul 31 00:32:50 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073021.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Jul 30 18:55:25 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.790	152	43386	0.400	ng	0.00
7) Naphthalene-d8	10.571	136	170575	0.400	ng	0.00
13) Acenaphthene-d10	14.421	164	90883	0.400	ng	0.00
19) Phenanthrene-d10	17.163	188	168292	0.400	ng	0.00
29) Chrysene-d12	21.376	240	158102	0.400	ng	0.00
35) Perylene-d12	23.720	264	168808	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.374	112	516489	4.427	ng	0.00
5) Phenol-d6	6.960	99	620760	4.206	ng	0.00
8) Nitrobenzene-d5	8.936	82	590672	5.172	ng	0.00
11) 2-Methylnaphthalene-d10	12.167	152	1210151	4.509	ng	0.00
14) 2,4,6-Tribromophenol	15.918	330	228596	6.200	ng	0.00
15) 2-Fluorobiphenyl	13.050	172	1573363	4.544	ng	0.00
27) Fluoranthene-d10	19.211	212	2191824	4.604	ng	0.00
31) Terphenyl-d14	19.817	244	1788204	4.447	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.354	88	80969m	5.337	ng	
3) n-Nitrosodimethylamine	3.677	42	155976	5.240	ng	# 94
6) bis(2-Chloroethyl)ether	7.218	93	504207	4.350	ng	98
9) Naphthalene	10.622	128	2075338	4.688	ng	100
10) Hexachlorobutadiene	10.913	225	369468	4.313	ng	# 100
12) 2-Methylnaphthalene	12.238	142	1321498	4.408	ng	99
16) Acenaphthylene	14.142	152	1999392	4.841	ng	100
17) Acenaphthene	14.484	154	1273903	4.955	ng	96
18) Fluorene	15.474	166	1527585	4.726	ng	98
21) 4-Bromophenyl-phenylether	16.360	248	494279	4.940	ng	# 65
22) Hexachlorobenzene	16.482	284	525662	5.387	ng	99
23) Atrazine	16.640	200	422667	4.693	ng	99
24) Pentachlorophenol	16.822	266	259775	7.266	ng	100
25) Phenanthrene	17.212	178	2278646	4.909	ng	100
26) Anthracene	17.297	178	2198628	4.897	ng	100
28) Fluoranthene	19.241	202	2432472	4.584	ng	99
30) Pyrene	19.604	202	2557346	4.731	ng	99
32) Benzo(a)anthracene	21.355	228	2504299	4.666	ng	99
33) Chrysene	21.408	228	2447176	4.800	ng	100
34) Bis(2-ethylhexyl)phtha...	21.301	149	1669648	5.429	ng	99
36) Indeno(1,2,3-cd)pyrene	26.130	276	3039506	4.935	ng	99
37) Benzo(b)fluoranthene	23.012	252	2628775	4.597	ng	99
38) Benzo(k)fluoranthene	23.060	252	2656223	4.712	ng	# 97
39) Benzo(a)pyrene	23.618	252	2443370	4.729	ng	99
40) Dibenzo(a,h)anthracene	26.151	278	2569611	4.953	ng	96
41) Benzo(g,h,i)perylene	26.862	276	2582627	4.936	ng	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN073021\
Data File : BN015727.D
Acq On : 30 Jul 2021 20:33
Operator : CG/JU
Sample : SSTDICC5.0
Misc :
ALS Vial : 8 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Manual Integrations
APPROVED

mohammad
8/2/2021 3:47:43 PM

Quant Time: Jul 31 00:32:50 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN073021.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Jul 30 18:55:25 2021
Response via : Initial Calibration





#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.790 min Scan# 511
 Delta R.T. -0.000 min
 Lab File: BN015727.D
 Acq: 30 Jul 2021 20:33





Tgt Ion: 152 Resp: 43386

Ion	Ratio	Lower	Upper
152	100		
150	158.1	123.9	185.9
115	59.9	47.8	71.6

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0

Manual Integrations
 APPROVED




mohammad
 8/2/2021 3:47:43 PM




#2
 1,4-Dioxane
 Concen: 5.337 ng m
 RT: 3.354 min Scan# 30
 Delta R.T. -0.000 min
 Lab File: BN015727.D
 Acq: 30 Jul 2021 20:33

Tgt Ion: 88 Resp: 80969

Ion	Ratio	Lower	Upper
88	100		
43	29.6	17.2	25.8
58	81.7	62.2	93.2









#3
n-Nitrosodimethylamine
Concen: 5.240 ng
RT: 3.677 min Scan# 65
Delta R.T. -0.037 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33

Tgt Ion: 42 Resp: 155976
Ion Ratio Lower Upper
42 100
74 153.3 116.6 174.8
44 6.5 6.6 10.0#

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0




Manual Integrations
APPROVED


mohammad
8/2/2021 3:47:43 PM



#4
2-Fluorophenol
Concen: 4.427 ng
RT: 5.374 min Scan# 249
Delta R.T. -0.009 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33

Tgt Ion: 112 Resp: 516489
Ion Ratio Lower Upper
112 100
64 51.7 42.1 63.1
63 26.7 21.9 32.9









#5
Phenol-d6
Concen: 4.206 ng
RT: 6.960 min Scan# 421
Delta R.T. 0.009 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Tgt Ion: 99 Resp: 620760
Ion Ratio Lower Upper
99 100
42 15.3 12.9 19.3
71 29.1 23.8 35.6




Manual Integrations
APPROVED


mohammad
8/2/2021 3:47:43 PM



#6
bis(2-Chloroethyl)ether
Concen: 4.350 ng
RT: 7.218 min Scan# 449
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33


Tgt Ion: 93 Resp: 504207
Ion Ratio Lower Upper
93 100
63 71.0 58.6 87.8
95 32.8 26.9 40.3





#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.571 min Scan# 746
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33



Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0




Tgt Ion:136 Resp: 170575
Ion Ratio Lower Upper
136 100
137 10.9 8.7 13.1
54 6.5 5.5 8.3
68 5.4 4.5 6.7

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:43 PM




#8
Nitrobenzene-d5
Concen: 5.172 ng
RT: 8.936 min Scan# 617
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33



Tgt Ion: 82 Resp: 590672
Ion Ratio Lower Upper
82 100
128 45.7 33.2 49.8
54 45.6 38.4 57.6





#9
Naphthalene
Concen: 4.688 ng
RT: 10.622 min Scan# 750
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33





Tgt Ion:128 Resp: 2075338

Ion	Ratio	Lower	Upper
128	100		
129	11.2	8.9	13.3
127	12.7	10.4	15.6

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Manual Integrations
APPROVED




mohammad
8/2/2021 3:47:43 PM




#10
Hexachlorobutadiene
Concen: 4.313 ng
RT: 10.913 min Scan# 773
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33

Tgt Ion:225 Resp: 369468





Ion	Ratio	Lower	Upper
225	100		
223	0.0	0.0	0.0
227	64.0	51.0	76.4








#11
2-Methylnaphthalene-d10
Concen: 4.509 ng
RT: 12.167 min Scan# 968
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33

Tgt Ion:152 Resp: 1210151
Ion Ratio Lower Upper
152 100
151 20.1 16.2 24.4



#12
2-Methylnaphthalene
Concen: 4.408 ng
RT: 12.238 min Scan# 984
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33


Tgt Ion:142 Resp: 1321498
Ion Ratio Lower Upper
142 100
141 87.4 70.5 105.7
115 32.9 27.3 40.9



Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Manual Integrations
APPROVED

mohammad
8/2/2021 3:47:43 PM







#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.421 min Scan# 1195
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33

Tgt Ion:164 Resp: 90883
Ion Ratio Lower Upper
164 100
162 99.6 79.8 119.6
160 44.3 35.8 53.6

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0




Manual Integrations
APPROVED


mohammad
8/2/2021 3:47:43 PM



#14
2,4,6-Tribromophenol
Concen: 6.200 ng
RT: 15.918 min Scan# 1313
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33


Tgt Ion:330 Resp: 228596
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 34.1 28.8 43.2





#15
2-Fluorobiphenyl
Concen: 4.544 ng
RT: 13.050 min Scan# 1087
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0




Tgt Ion:172 Resp: 1573363
Ion Ratio Lower Upper
172 100
171 33.2 26.9 40.3
170 21.3 17.4 26.0

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:43 PM




#16
Acenaphthylene
Concen: 4.841 ng
RT: 14.142 min Scan# 1173
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33




Tgt Ion:152 Resp: 1999392
Ion Ratio Lower Upper
152 100
151 19.4 15.7 23.5
153 13.2 10.4 15.6





#17
Acenaphthene
Concen: 4.955 ng
RT: 14.484 min Scan# 1200
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0



Tgt Ion:154 Resp: 1273903
Ion Ratio Lower Upper
154 100
153 108.0 89.7 134.5
152 51.3 43.8 65.6




Manual Integrations
APPROVED


mohammad
8/2/2021 3:47:43 PM



#18
Fluorene
Concen: 4.726 ng
RT: 15.474 min Scan# 1278
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33


Tgt Ion:166 Resp: 1527585
Ion Ratio Lower Upper
166 100
165 95.9 78.1 117.1
167 13.6 10.8 16.2





#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.163 min Scan# 1415
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0




Tgt Ion:188 Resp: 168292
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 12.4 10.2 15.4

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:43 PM




#21
4-Bromophenyl-phenylether
Concen: 4.940 ng
RT: 16.360 min Scan# 1349
Delta R.T. -0.012 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33




Tgt Ion:248 Resp: 494279
Ion Ratio Lower Upper
248 100
250 89.4 83.9 125.9
141 88.2 26.9 40.3#





#22
Hexachlorobenzene
Concen: 5.387 ng
RT: 16.482 min Scan# 1359
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0




Tgt Ion:284 Resp: 525662
Ion Ratio Lower Upper
284 100
142 35.5 29.4 44.0
249 29.5 23.8 35.6

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:43 PM




#23
Atrazine
Concen: 4.693 ng
RT: 16.640 min Scan# 1372
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33



Tgt Ion:200 Resp: 422667
Ion Ratio Lower Upper
200 100
173 25.6 18.9 28.3
215 47.8 38.1 57.1









#24
Pentachlorophenol
Concen: 7.266 ng
RT: 16.822 min Scan# 1387
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33

Tgt Ion:266 Resp: 259775
Ion Ratio Lower Upper
266 100
264 62.4 50.2 75.2
268 64.0 51.1 76.7

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0




Manual Integrations
APPROVED


mohammad
8/2/2021 3:47:43 PM



#25
Phenanthrene
Concen: 4.909 ng
RT: 17.212 min Scan# 1419
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33


Tgt Ion:178 Resp: 2278646
Ion Ratio Lower Upper
178 100
176 18.9 15.3 22.9
179 15.6 12.4 18.6





#26
 Anthracene
 Concen: 4.897 ng
 RT: 17.297 min Scan# 1426
 Delta R.T. -0.000 min
 Lab File: BN015727.D
 Acq: 30 Jul 2021 20:33

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC5.0





Tgt Ion:178 Resp: 2198628


Ion	Ratio	Lower	Upper
178	100		
176	18.3	14.6	22.0
179	15.6	12.2	18.4

Manual Integrations
 APPROVED

mohammad
 8/2/2021 3:47:43 PM






#27
 Fluoranthene-d10
 Concen: 4.604 ng
 RT: 19.211 min Scan# 1726
 Delta R.T. -0.000 min
 Lab File: BN015727.D
 Acq: 30 Jul 2021 20:33



Tgt Ion:212 Resp: 2191824


Ion	Ratio	Lower	Upper
212	100		
106	13.5	10.6	16.0
104	7.5	5.9	8.9





#28
Fluoranthene
Concen: 4.584 ng
RT: 19.241 min Scan# 1732
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0



Tgt Ion:202 Resp: 2432472
Ion Ratio Lower Upper
202 100
101 11.5 9.0 13.6
203 17.7 13.8 20.8




Manual Integrations
APPROVED


mohammad
8/2/2021 3:47:43 PM



#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.376 min Scan# 2015
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33


Tgt Ion:240 Resp: 158102
Ion Ratio Lower Upper
240 100
120 9.2 7.4 11.0
236 25.8 20.9 31.3





#30
Pyrene
Concen: 4.731 ng
RT: 19.604 min Scan# 1805
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0




Tgt Ion:202 Resp: 2557346
Ion Ratio Lower Upper
202 100
200 20.7 16.5 24.7
203 17.9 14.0 21.0

Manual Integrations
APPROVED



mohammad
8/2/2021 3:47:43 PM




#31
Terphenyl-d14
Concen: 4.447 ng
RT: 19.817 min Scan# 1848
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33



Tgt Ion:244 Resp: 1788204
Ion Ratio Lower Upper
244 100
212 7.5 5.9 8.9
122 12.4 9.6 14.4









#32
Benzo(a)anthracene
Concen: 4.666 ng
RT: 21.355 min Scan# 2013
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33

Instrument :
BNA_N
ClientSampled :
SSTDICC5.0

Tgt Ion:228 Resp: 2504299
Ion Ratio Lower Upper
228 100
226 27.4 21.9 32.9
229 19.9 15.4 23.0




Manual Integrations
APPROVED


mohammad
8/2/2021 3:47:43 PM




#33
Chrysene
Concen: 4.800 ng
RT: 21.408 min Scan# 2018
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33

Tgt Ion:228 Resp: 2447176
Ion Ratio Lower Upper
228 100
226 30.4 24.3 36.5
229 19.6 15.3 22.9








#34
Bis(2-ethylhexyl)phthalate
Concen: 5.429 ng
RT: 21.301 min Scan# 2008
Delta R.T. -0.011 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33






Tgt Ion:149 Resp: 1669648
Ion Ratio Lower Upper
149 100
167 27.5 21.5 32.3
279 4.0 3.2 4.8



#35
Perylene-d12
Concen: 0.400 ng
RT: 23.720 min Scan# 2496
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33


Tgt Ion:264 Resp: 168808
Ion Ratio Lower Upper
264 100
260 24.4 19.4 29.2
265 24.7 19.9 29.9



Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Manual Integrations
APPROVED

mohammad
8/2/2021 3:47:43 PM







#36
Indeno(1,2,3-cd)pyrene
Concen: 4.935 ng
RT: 26.130 min Scan# 3200
Delta R.T. 0.003 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33

Tgt Ion:276 Resp: 3039506
Ion Ratio Lower Upper
276 100
138 25.0 20.6 30.8
277 25.2 20.1 30.1

Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0




Manual Integrations
APPROVED


mohammad
8/2/2021 3:47:43 PM




#37
Benzo(b)fluoranthene
Concen: 4.597 ng
RT: 23.012 min Scan# 2289
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33

Tgt Ion:252 Resp: 2628775
Ion Ratio Lower Upper
252 100
253 22.2 18.1 27.1
125 10.3 8.3 12.5









#38
Benzo(k)fluoranthene
Concen: 4.712 ng
RT: 23.060 min Scan# 2303
Delta R.T. 0.003 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33





Tgt Ion:252 Resp: 2656223
Ion Ratio Lower Upper
252 100
253 22.3 18.1 27.1
125 10.2 10.4 15.6#



#39
Benzo(a)pyrene
Concen: 4.729 ng
RT: 23.618 min Scan# 2466
Delta R.T. -0.000 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33




Tgt Ion:252 Resp: 2443370
Ion Ratio Lower Upper
252 100
253 22.2 18.1 27.1
125 11.1 9.2 13.8




Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Manual Integrations
APPROVED




mohammad
8/2/2021 3:47:43 PM




#40
Dibenzo(a,h)anthracene
Concen: 4.953 ng
RT: 26.151 min Scan# 3206
Delta R.T. 0.003 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33





Tgt Ion:278 Resp: 2569611
Ion Ratio Lower Upper
278 100
139 16.5 16.0 24.0
279 23.8 19.2 28.8



#41
Benzo(g,h,i)perylene
Concen: 4.936 ng
RT: 26.862 min Scan# 3414
Delta R.T. 0.003 min
Lab File: BN015727.D
Acq: 30 Jul 2021 20:33



Tgt Ion:276 Resp: 2582627
Ion Ratio Lower Upper
276 100
277 24.0 19.2 28.8
138 21.1 16.8 25.2



Instrument :
BNA_N
ClientSampleId :
SSTDICC5.0

Manual Integrations
APPROVED

mohammad
8/2/2021 3:47:43 PM