

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN073118\
 Data File : BN002213.D
 Acq On : 31 Jul 2018 16:23
 Operator : JU/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Jul 31 16:56:02 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN073018.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Jul 30 16:17:29 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	93	0.00
2	1,4-Dioxane	0.400	0.352	12.0	90	0.00
3	n-Nitrosodimethylamine	0.400	0.353	11.8	84	0.00
4 S	2-Fluorophenol	0.400	0.379	5.3	92	0.00
5 S	Phenol-d6	0.400	0.406	-1.5	101	0.00
6	bis(2-Chloroethyl)ether	0.400	0.370	7.5	86	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	96	0.00
8 S	Nitrobenzene-d5	0.400	0.380	5.0	93	0.00
9	Naphthalene	0.400	0.395	1.3	96	-0.01
10	Hexachlorobutadiene	0.400	0.388	3.0	95	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.382	4.5	94	0.00
12	2-Methylnaphthalene	0.400	0.386	3.5	93	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	86	0.00
14 S	2,4,6-Tribromophenol	0.400	0.327	18.3	72	0.00
15 S	2-Fluorobiphenyl	0.400	0.399	0.3	91	0.00
16	Acenaphthylene	0.400	0.381	4.8	86	0.00
17	Acenaphthene	0.400	0.389	2.8	87	0.00
18	Fluorene	0.400	0.374	6.5	81	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	69	0.00
20	4-Bromophenyl-phenylether	0.400	0.399	0.3	75	0.00
21	Hexachlorobenzene	0.400	0.408	-2.0	75	0.00
22	Pentachlorophenol	0.400	0.372	7.0	69	0.00
23	Phenanthrene	0.400	0.380	5.0	68	0.00
24	Anthracene	0.400	0.369	7.8	67	0.00
25 SURR	Fluoranthene-d10	0.400	0.347	13.3	59	0.00
26	Fluoranthene	0.400	0.342	14.5	58	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	53	0.00
28	Pvrene	0.400	0.402	-0.5	57	0.00
29 S	Terphenyl-d14	0.400	0.392	2.0	56	0.00
30	Benzo(a)anthracene	0.400	0.379	5.3	52	0.00
31	Chrysene	0.400	0.389	2.8	54	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.372	7.0	50	-0.01
33	Indeno(1,2,3-cd)pyrene	0.400	0.453	-13.2	71	0.00
34 I	Perylene-d12	0.400	0.400	0.0	62	0.00
35	Benzo(b)fluoranthene	0.400	0.369	7.8	58	0.00
36	Benzo(k)fluoranthene	0.400	0.370	7.5	60	0.00
37 C	Benzo(a)pyrene	0.400	0.394	1.5	64	0.00
38	Dibenzo(a,h)anthracene	0.400	0.406	-1.5	72	0.00
39	Benzo(a,h,i)perylene	0.400	0.406	-1.5	71	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA N\Data\BN073118\
Data File : BN002213.D
Acq On : 31 Jul 2018 16:23
Operator : JU/SJ
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Jul 31 16:56:02 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN073018.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Jul 30 16:17:29 2018
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				