




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080123\
 Data File : BN026772.D
 Acq On : 01 Aug 2023 07:04
 Operator : MA/JU
 Sample : DFTPP
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 02 01:19:41 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 02 01:19:36 2023
 Response via : Initial Calibration



TIC: BN026772.D\data.ms

(70) Pentachlorophenol (C)




17.116min (0.000) 25386.37 ng

response 701275

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	69.21
264.00	61.60	65.58
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080123\
Data File : BN026772.D
Acq On : 01 Aug 2023 07:04
Operator : MA/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Aug 02 01:19:41 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 02 01:19:36 2023
Response via : Initial Calibration



TIC: BN026772.D\data.ms

(77) Benzidine

19.745min (0.000) 0.00 ng

response 2997095

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	13.97
183.00	13.20	11.15
0.00	0.00	0.00