

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN080123\
 Data File : BN026780.D
 Acq On : 01 Aug 2023 12:32
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN080123

Quant Time: Aug 02 00:36:38 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN080123.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 02 00:35:47 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
2	1,4-Dioxane	0.467	0.490	-4.9	88	0.00
3	n-Nitrosodimethylamine	0.521	0.551	-5.8	85	0.00
4 S	2-Fluorophenol	1.013	0.979	3.4	83	0.00
5 S	Phenol-d6	1.312	1.276	2.7	84	0.00
6	bis(2-Chloroethyl)ether	1.183	1.258	-6.3	87	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	86	0.00
8 S	Nitrobenzene-d5	0.237	0.236	0.4	90	0.00
9	Naphthalene	1.100	1.136	-3.3	87	0.00
10	Hexachlorobutadiene	0.199	0.212	-6.5	88	0.00
11 SURR	2-Methylnaphthalene-d10	0.574	0.585	-1.9	86	0.00
12	2-Methylnaphthalene	0.751	0.767	-2.1	86	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	86	0.00
14 S	2,4,6-Tribromophenol	0.136	0.114	16.2	80	0.00
15 S	2-Fluorobiphenyl	1.633	1.663	-1.8	86	0.00
16	Acenaphthylene	1.977	1.959	0.9	85	0.00
17	Acenaphthene	1.240	1.270	-2.4	86	0.00
18	Fluorene	1.615	1.659	-2.7	87	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	86	0.00
20	4,6-Dinitro-2-methylphenol	0.006	0.007	-16.7	87	0.00
21	4-Bromophenyl-phenylether	0.242	0.248	-2.5	86	0.00
22	Hexachlorobenzene	0.276	0.299	-8.3	91	0.00
23	Atrazine	0.160	0.143	10.6	82	0.00
24	Pentachlorophenol	0.071	0.065	8.5	82	0.00
25	Phenanthrene	1.215	1.266	-4.2	88	0.00
26	Anthracene	1.085	1.059	2.4	85	0.00
27 SURR	Fluoranthene-d10	0.960	0.948	1.3	85	0.00
28	Fluoranthene	1.353	1.364	-0.8	86	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	87	0.00
30	Pyrene	1.728	1.748	-1.2	86	0.00
31 S	Terphenyl-d14	0.834	0.799	4.2	85	0.00
32	Benzo(a)anthracene	1.374	1.358	1.2	85	0.00
33	Chrysene	1.518	1.562	-2.9	87	0.00
34	Bis(2-ethylhexyl)phthalate	0.455	0.440	3.3	78	0.00
35 I	Perylene-d12	1.000	1.000	0.0	86	0.00
36	Indeno(1,2,3-cd)pyrene	1.632	1.623	0.6	89	0.00
37	Benzo(b)fluoranthene	1.574	1.571	0.2	88	0.00
38	Benzo(k)fluoranthene	1.617	1.635	-1.1	88	0.00
39 C	Benzo(a)pyrene	1.435	1.498	-4.4	91	0.00
40	Dibenzo(a,h)anthracene	1.268	1.308	-3.2	89	0.00
41	Benzo(g,h,i)perylene	1.501	1.464	2.5	87	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN080123