

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN080318\
 Data File : BN002229.D
 Acq On : 03 Aug 2018 19:17
 Operator : JU/SJ
 Sample : J4265-10
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C0AB4

Quant Time: Aug 04 03:47:14 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN071318MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Sat Aug 04 03:46:32 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.69	152	259499	20.00	ng/ul	0.00
18) Naphthalene-d8	10.46	136	1186591	20.00	ng/ul	0.00
35) Acenaphthene-d10	14.32	164	681833	20.00	ng/ul	0.00
61) Phenanthrene-d10	17.06	188	1385107	20.00	ng/ul	0.00
77) Chrysene-d12	21.26	240	1039894	20.00	ng/ul	0.00
85) Perylene-d12	23.49	264	976972	20.00	ng/ul	0.00

System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.25	96	39350	6.84	ng/uL	0.00
5) Phenol-d5	6.86	99	205398	8.40	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	7.03	67	486289	32.66	ng/ul	0.00
9) 2-Chlorophenol-d4	7.22	132	514866	26.67	ng/ul	0.00
13) 4-Methylphenol-d8	8.39	113	368407	18.72	ng/ul	0.00
19) Nitrobenzene-d5	8.83	128	321488	32.70	ng/ul	0.00
22) 2-Nitrophenol-d4	9.55	143	347480	31.27	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	10.09	165	575350	28.59	ng/ul	0.00
29) 4-Chloroaniline-d4	10.60	131	48351	2.09	ng/ul	0.00
43) Dimethylphthalate-d6	13.73	166	1911237	33.07	ng/ul	0.00
46) Acenaphthylene-d8	14.01	160	2449378	34.07	ng/ul	0.00
51) 4-Nitrophenol-d4	14.54	143	65408	5.92	ng/ul	0.00
57) Fluorene-d10	15.32	176	1613813	32.47	ng/ul	0.00
62) 4,6-Dinitro-2-methylphenol	15.44	200	253341	28.08	ng/ul	0.00
70) Anthracene-d10	17.16	188	2304117	34.12	ng/ul	0.00
78) Pyrene-d10	19.46	212	2149345	39.76	ng/ul	0.00
89) Benzo(a)pyrene-d12	23.36	264	1743738	33.05	ng/ul	0.01

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN080318\
Data File : BN002229.D
Acq On : 03 Aug 2018 19:17
Operator : JU/SJ
Sample : J4265-10
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C0AB4

Quant Time: Aug 04 03:47:14 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN071318MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Sat Aug 04 03:46:32 2018
Response via : Initial Calibration

